

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111924\
 Data File : BN035194.D
 Acq On : 20 Nov 2024 12:07
 Operator : RC/JU
 Sample : P4875-13
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E2A01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/21/2024
 Supervised By :mohammad ahmed 11/22/2024

Quant Time: Nov 20 12:34:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN111624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 18 00:44:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.541	152	3129	0.400	ng/ul	0.00
4) Naphthalene-d8	10.300	136	7440	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.180	164	4911	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.937	188	10916m	0.400	ng/ul	0.00
17) Chrysene-d12	21.148	240	10064	0.400	ng/ul	0.00
23) Perylene-d12	23.320	264	10466m	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.136	96	13516	4.726	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.901	152	2635	0.242	ng/ul	-0.01
18) Fluoranthene-d10	18.977	212	9906	0.357	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111924\
Data File : BN035194.D
Acq On : 20 Nov 2024 12:07
Operator : RC/JU
Sample : P4875-13
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E2A01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/21/2024
Supervised By :mohammad ahmed 11/22/2024

Quant Time: Nov 20 12:34:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN111624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 18 00:44:37 2024
Response via : Initial Calibration

