


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN112020\
 Data File : BN012654.D
 Acq On : 20 Nov 2020 11:14
 Operator : CG/JU
 Sample : DFTPP18
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP18

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN112020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Nov 20 18:13:27 2020



AutoFind: Scans 2520, 2521, 2522; Background Corrected with Scan 2513

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.5	312341	PASS
68	69	0.00	2	1.9	5985	PASS
69	198	0.00	100	45.7	307200	PASS
70	69	0.00	2	0.6	1748	PASS
127	198	10	80	48.5	325440	PASS
197	198	0.00	2	0.3	1705	PASS
198	198	100	100	100.0	671488	PASS
199	198	5	9	6.7	44866	PASS
275	198	10	60	29.0	194410	PASS
365	198	1	100	4.4	29253	PASS
441	198	0.01	100	15.6	104501	PASS
442	198	50	100	97.6	655210	PASS
443	442	15	24	17.4	113773	PASS

Scan 2513 (17.269 min): BN012654.D
DFTPP18

m/z	Abundance
40.10	849.0
44.00	1490.0
73.10	1019.0
83.80	471.0
193.10	307.0
207.00	1781.0
208.00	541.0
280.90	704.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

m/z	Abundance
37.10	1806.0
38.10	4803.0
39.10	24888.0
40.00	1653.0
41.10	597.0
43.00	398.0
43.90	2039.0
45.00	326.0
49.10	2776.0
50.10	93536.0
51.10	330688.0
52.10	15860.0
53.00	665.0
55.00	1667.0
56.00	10723.0
57.00	19448.0
58.00	692.0
59.00	530.0
60.00	317.0
61.00	4688.0
62.00	4882.0
63.00	13509.0
64.00	2065.0
65.00	5780.0
66.00	430.0
67.00	599.0
68.20	6165.0
69.00	320128.0
70.00	2188.0
73.10	4106.0
74.00	34200.0
75.00	51840.0
76.10	17856.0
77.10	345088.0
78.10	22736.0
79.00	28624.0
80.00	21152.0
81.00	27288.0
82.00	6668.0
83.00	5337.0
83.90	876.0
85.00	4033.0
86.00	8223.0
87.10	3195.0
88.00	1350.0
88.80	352.0
91.00	5963.0
92.00	7370.0
93.00	48640.0
94.00	2873.0
95.00	968.0
96.00	2256.0
98.00	39032.0
99.00	26408.0
99.90	2362.0
101.00	14436.0
102.10	722.0
103.00	4619.0
104.00	9235.0
105.00	8828.0
106.10	2703.0
107.00	104960.0
108.00	14210.0
109.00	3545.0
110.00	178688.0
111.00	29184.0
112.10	2966.0
113.10	1183.0
114.30	424.0
115.00	479.0
116.00	7070.0
117.00	114112.0
118.00	8595.0
119.00	1209.0
120.00	1405.0
121.00	641.0
122.00	7536.0
123.00	9509.0
124.00	5208.0
125.00	5203.0
127.00	303040.0
128.00	25536.0
129.00	135936.0
130.00	12401.0
131.00	1928.0
132.00	1086.0
133.10	520.0
134.00	4481.0
135.00	10830.0
136.00	4057.0
137.00	6159.0
138.00	1534.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

138.90	694.0
140.00	1898.0
141.00	19912.0
142.10	5190.0
143.00	4441.0
144.00	1154.0
144.90	1153.0
146.00	3239.0
147.00	9539.0
148.00	24304.0
149.00	5648.0
150.10	1342.0
151.20	2467.0
152.00	1583.0
153.00	5343.0
154.00	3855.0
155.00	10207.0
156.00	14300.0
157.00	2753.0
158.00	2876.0
159.00	2739.0
160.00	5572.0
161.00	7785.0
162.00	2189.0
162.90	650.0
164.00	922.0
165.00	7068.0
166.00	5502.0
167.00	38704.0
168.00	19568.0
169.00	3281.0
169.90	1075.0
171.00	1619.0
172.00	3428.0
173.00	4048.0
174.10	7611.0
175.00	12428.0
176.00	4243.0
176.90	6908.0
178.00	2825.0
179.00	27056.0
180.00	18616.0
181.00	8285.0
182.00	1444.0
183.20	554.0
184.00	2285.0
185.00	11356.0
186.00	95800.0
187.00	28360.0
188.10	3099.0
189.00	6900.0
189.90	1063.0
191.00	3254.0
192.00	7998.0
193.00	9538.0
194.10	1629.0
195.10	968.0
196.00	15915.0
198.00	580160.0
199.00	42272.0
200.00	3116.0
201.40	1793.0
202.00	1333.0
203.00	4399.0
204.00	23568.0
205.00	40552.0
206.10	166656.0
207.00	24168.0
207.90	6379.0
209.00	1984.0
210.10	2612.0
211.00	6625.0
211.90	933.0
213.00	663.0
215.00	1793.0
216.00	4072.0
217.00	45248.0
218.00	6046.0
219.00	891.0
220.10	721.0
221.00	22000.0
223.00	9351.0
224.10	88896.0
225.10	20488.0
226.00	3262.0
227.00	41848.0
228.00	6621.0
229.00	8656.0
229.90	1274.0
231.00	3780.0
231.90	459.0
233.00	622.0
234.00	2236.0
235.00	3000.0
236.00	1883.0
237.10	3956.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

238.10	737.0
239.00	2408.0
240.00	1588.0
241.00	1989.0
241.90	5621.0
243.00	5210.0
244.10	63456.0
245.00	9811.0
246.00	15739.0
247.00	3276.0
248.00	1021.0
249.00	2382.0
249.90	381.0
250.90	644.0
252.10	504.0
253.10	1445.0
255.00	333888.0
256.00	52704.0
257.10	3934.0
258.00	23184.0
259.00	3452.0
259.90	684.0
261.10	771.0
262.90	445.0
264.10	558.0
265.00	10144.0
266.00	1130.0
270.00	355.0
271.00	751.0
272.10	1051.0
273.00	11384.0
274.00	26712.0
275.00	157376.0
276.00	21608.0
277.00	15088.0
277.90	2609.0
279.00	457.0
281.10	517.0
282.00	588.0
283.00	1929.0
284.00	1023.0
285.00	1878.0
286.00	499.0
289.00	417.0
290.00	353.0
291.00	461.0
292.00	753.0
293.00	3124.0
294.00	844.0
295.10	1319.0
296.00	49872.0
297.00	7251.0
298.00	445.0
300.90	541.0
302.00	513.0
303.00	5612.0
303.90	1670.0
304.90	607.0
307.90	991.0
308.80	533.0
310.20	534.0
313.00	339.0
313.90	1597.0
314.90	4954.0
316.00	2760.0
316.90	530.0
319.90	420.0
321.00	1782.0
322.00	647.0
323.10	13931.0
324.00	2382.0
327.00	3394.0
328.00	1614.0
329.10	325.0
332.00	813.0
333.10	1242.0
334.00	8207.0
335.00	2278.0
335.90	320.0
338.90	312.0
339.90	382.0
341.00	2005.0
346.00	2921.0
346.90	544.0
350.90	393.0
352.00	4616.0
352.90	2495.0
354.10	4056.0
355.00	1007.0
365.00	24072.0
365.90	2975.0
370.00	373.0
371.00	1032.0
372.00	5966.0
373.00	1611.0
377.10	328.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

383.00	1844.0
384.00	362.0
390.10	1121.0
391.10	689.0
391.90	656.0
401.10	453.0
402.00	2137.0
403.00	2996.0
404.00	752.0
421.00	2000.0
422.00	2497.0
423.00	21432.0
424.00	7333.0
424.90	910.0
441.00	65480.0
442.00	429568.0
443.00	68464.0
444.10	7082.0
445.00	574.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

m/z	Abundance
36.10	360.0
37.00	1541.0
38.00	5505.0
39.10	29904.0
40.00	1891.0
41.00	1020.0
43.00	325.0
44.00	2105.0
45.10	978.0
49.10	2974.0
50.00	105120.0
51.00	382080.0
52.00	18936.0
53.00	821.0
55.00	2438.0
56.00	11292.0
57.00	22888.0
58.00	1352.0
59.20	431.0
61.00	6375.0
62.00	5994.0
63.10	12971.0
64.00	2202.0
65.10	5947.0
65.90	499.0
67.10	772.0
68.10	7624.0
69.00	372736.0
70.00	1827.0
71.90	416.0
73.00	3835.0
74.00	45216.0
75.00	63512.0
76.10	22176.0
77.10	405248.0
78.10	27376.0
79.00	31912.0
80.00	24888.0
81.00	31880.0
82.00	7213.0
83.00	6137.0
84.00	587.0
85.00	5543.0
86.00	9406.0
87.00	4581.0
88.10	1318.0
89.10	975.0
91.00	8155.0
92.00	7467.0
93.00	59176.0
94.10	3241.0
95.10	900.0
96.10	2499.0
96.90	950.0
98.00	49120.0
99.00	31928.0
100.00	3093.0
101.00	16247.0
101.90	694.0
103.00	7141.0
104.00	10162.0
105.00	10958.0
106.10	4026.0
107.00	132864.0
108.00	19784.0
109.20	4173.0
110.00	224512.0
111.00	35536.0
112.00	3613.0
113.00	1986.0
115.00	771.0
116.00	8520.0
117.00	142464.0
118.00	9189.0
119.00	1663.0
119.90	2054.0
120.90	666.0
122.00	10218.0
123.00	13142.0
124.00	6528.0
125.00	5461.0
126.20	2770.0
127.00	417152.0
128.00	33560.0
129.00	177152.0
130.00	14738.0
131.00	3034.0
132.00	1228.0
132.80	625.0
134.00	5328.0
135.00	15450.0
136.00	6514.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

137.00	6716.0
138.10	1578.0
139.00	567.0
140.00	2312.0
141.00	26832.0
142.00	8055.0
143.00	4865.0
144.00	1691.0
144.90	1016.0
146.00	4563.0
147.00	12848.0
148.00	32800.0
149.00	6977.0
150.10	1382.0
151.00	3225.0
151.80	1813.0
153.00	6647.0
154.00	4496.0
155.00	12988.0
156.00	20712.0
157.00	3655.0
158.00	4657.0
158.90	3403.0
160.00	7835.0
161.00	11594.0
162.00	3154.0
163.00	1120.0
164.00	1364.0
165.00	9608.0
166.00	7074.0
167.00	50512.0
168.00	26296.0
169.00	4345.0
170.10	1537.0
171.00	2428.0
172.00	3710.0
173.00	5164.0
174.00	8945.0
175.00	20008.0
176.10	4842.0
177.00	10772.0
178.00	3195.0
179.00	33392.0
180.00	23592.0
181.00	10420.0
182.00	2031.0
183.00	837.0
184.00	2992.0
185.00	16416.0
186.00	126752.0
187.10	38352.0
188.00	4497.0
189.00	9516.0
189.90	1635.0
191.10	4557.0
192.00	10766.0
193.00	13707.0
194.00	2550.0
194.90	1264.0
196.00	23544.0
198.00	834816.0
199.00	56176.0
200.00	4064.0
201.40	3787.0
203.00	6665.0
204.00	32344.0
205.10	54936.0
206.00	226688.0
207.10	32688.0
208.00	9472.0
209.00	3439.0
210.20	3926.0
211.00	8639.0
212.90	470.0
214.90	2993.0
216.00	5346.0
217.00	65432.0
218.00	7290.0
219.00	810.0
221.00	32472.0
221.90	6618.0
223.00	13967.0
224.00	124304.0
225.00	34208.0
226.10	3905.0
227.00	62280.0
228.00	8350.0
229.00	11287.0
230.00	1449.0
231.00	5017.0
232.00	860.0
233.00	1189.0
234.00	4420.0
235.00	3799.0
236.00	4092.0
237.00	4609.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

238.10	907.0
239.00	2298.0
240.00	1590.0
241.00	3463.0
242.10	7693.0
243.00	7011.0
244.00	99344.0
245.00	12212.0
246.00	22704.0
247.00	4553.0
248.00	1091.0
249.00	4224.0
249.90	596.0
251.00	1000.0
251.90	1072.0
253.00	2250.0
255.00	515712.0
256.00	77584.0
257.00	5077.0
258.00	34280.0
259.00	5638.0
260.00	793.0
260.90	1008.0
263.00	379.0
264.20	507.0
265.00	13373.0
266.00	2409.0
266.90	352.0
268.00	383.0
269.90	637.0
270.90	920.0
272.00	1269.0
273.00	16263.0
274.00	45504.0
275.00	232768.0
276.10	29128.0
277.00	21088.0
278.00	3880.0
279.10	711.0
280.90	542.0
282.10	760.0
283.00	2784.0
284.10	1303.0
285.10	3991.0
286.00	697.0
287.00	418.0
288.00	497.0
289.00	1128.0
290.00	1057.0
290.90	718.0
292.00	1330.0
293.00	4659.0
294.00	1364.0
296.00	75784.0
297.00	9837.0
298.00	944.0
301.00	1194.0
302.00	1319.0
303.00	7936.0
304.00	2680.0
305.10	440.0
308.00	1241.0
308.90	717.0
310.00	800.0
312.10	369.0
313.00	945.0
314.00	3107.0
315.00	7911.0
316.10	3470.0
317.10	692.0
321.10	2186.0
322.00	996.0
323.00	21832.0
324.10	3556.0
324.90	549.0
325.90	475.0
326.90	3997.0
328.00	1990.0
329.00	540.0
332.00	1379.0
333.00	2062.0
334.00	14105.0
335.10	3817.0
335.90	417.0
338.90	332.0
341.00	2640.0
341.90	671.0
346.00	4321.0
347.10	664.0
351.00	628.0
352.00	6471.0
353.00	4917.0
354.10	7043.0
355.10	970.0
359.00	332.0
359.90	447.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

365.00	38176.0
365.90	4589.0
367.00	363.0
369.70	496.0
370.00	427.0
371.00	1906.0
372.00	11489.0
373.00	3534.0
382.90	3184.0
384.10	849.0
390.10	1441.0
391.00	817.0
392.10	941.0
393.00	445.0
401.00	538.0
402.00	4396.0
403.00	6589.0
404.00	2083.0
405.00	513.0
414.90	441.0
421.00	5372.0
422.00	4935.0
423.00	37680.0
424.00	8227.0
424.80	1191.0
441.00	123720.0
442.00	716160.0
443.00	146304.0
444.10	13424.0
445.00	466.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

m/z	Abundance
37.10	1292.0
38.00	3383.0
39.10	15743.0
39.90	1244.0
41.00	512.0
43.00	479.0
44.00	2220.0
45.10	484.0
49.10	2594.0
50.10	67088.0
51.10	224256.0
52.00	11404.0
52.90	490.0
55.00	1395.0
56.00	6543.0
57.00	14577.0
57.90	853.0
61.00	2647.0
62.00	3739.0
63.00	9668.0
64.10	981.0
65.00	3869.0
66.90	497.0
68.00	4167.0
69.00	228736.0
70.00	1231.0
73.00	2864.0
74.00	28360.0
75.00	38496.0
76.00	12856.0
77.10	249600.0
78.00	15818.0
79.00	21200.0
80.00	13854.0
81.00	20024.0
82.00	4464.0
83.00	4066.0
83.90	516.0
85.00	3266.0
86.00	6394.0
87.00	2836.0
87.90	845.0
89.00	635.0
91.00	5385.0
92.00	5353.0
93.00	37616.0
94.10	2279.0
95.00	321.0
96.00	1772.0
97.10	637.0
98.00	32400.0
99.00	19840.0
99.90	2207.0
101.00	10958.0
102.00	570.0
102.90	2993.0
104.00	7129.0
105.00	7638.0
106.00	1871.0
107.00	82008.0
108.00	12324.0
110.00	141120.0
111.00	21696.0
112.00	2627.0
113.00	990.0
115.00	370.0
116.10	4315.0
117.00	94688.0
118.00	5761.0
119.10	921.0
120.00	633.0
121.00	760.0
122.00	6922.0
123.00	8956.0
124.00	3339.0
125.00	4249.0
127.00	256128.0
128.00	21104.0
129.00	109920.0
130.00	8595.0
131.00	1823.0
132.00	1078.0
133.20	585.0
134.00	4097.0
135.00	10578.0
136.00	3663.0
137.00	4652.0
137.90	1332.0
139.00	1301.0
140.00	1933.0
141.00	16382.0
142.00	5613.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

143.00	4017.0
144.00	902.0
145.00	664.0
146.00	2797.0
147.00	8219.0
148.00	20280.0
149.00	4152.0
150.00	1332.0
151.10	1954.0
152.10	1117.0
153.00	5463.0
154.00	3490.0
155.00	9809.0
156.00	13128.0
157.10	1870.0
157.90	3332.0
159.00	2651.0
160.00	5899.0
161.00	6742.0
162.00	1861.0
162.90	560.0
163.90	888.0
165.00	6923.0
166.10	4277.0
167.00	36360.0
168.00	16159.0
169.10	2776.0
169.90	1090.0
171.00	1768.0
172.10	3677.0
173.00	3911.0
174.00	7372.0
175.00	14552.0
176.00	3418.0
177.00	6224.0
178.00	2504.0
179.00	25400.0
180.00	16600.0
181.00	7164.0
182.00	1471.0
182.90	643.0
184.00	2364.0
185.10	12341.0
186.00	88224.0
187.00	26176.0
188.00	2447.0
189.00	6453.0
189.90	877.0
191.00	2829.0
192.00	9768.0
193.00	9228.0
194.10	1984.0
195.00	901.0
196.00	17456.0
197.10	5115.0
197.90	599488.0
199.00	36152.0
199.90	2271.0
201.50	1773.0
203.00	4637.0
204.00	22936.0
205.00	40496.0
206.00	164544.0
207.00	21840.0
208.00	6498.0
208.90	1683.0
210.20	2629.0
211.00	6074.0
212.00	978.0
213.00	624.0
214.90	2211.0
216.10	3629.0
217.00	42728.0
218.00	5246.0
218.90	484.0
221.00	21800.0
222.00	5668.0
223.00	9938.0
224.00	90496.0
225.00	23608.0
227.00	45776.0
228.00	6919.0
229.00	8534.0
230.00	1211.0
231.00	3133.0
232.00	605.0
233.00	698.0
234.00	2852.0
235.00	3632.0
236.00	1928.0
237.00	3781.0
238.00	687.0
239.00	1962.0
239.90	1163.0
241.00	2332.0
242.00	5460.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

243.10	5945.0
244.00	72176.0
245.00	9635.0
246.00	16384.0
247.00	3203.0
248.00	805.0
248.90	3476.0
250.10	699.0
251.10	744.0
252.00	943.0
253.10	1824.0
255.00	366784.0
256.00	59128.0
257.00	4295.0
258.00	25008.0
259.00	4469.0
259.90	424.0
261.00	800.0
264.00	988.0
265.00	10582.0
266.00	1270.0
270.10	729.0
271.10	1059.0
272.00	912.0
273.00	11587.0
274.00	31584.0
275.00	193088.0
276.00	22672.0
277.00	19416.0
278.00	3044.0
279.10	739.0
280.90	933.0
282.10	719.0
283.10	1718.0
283.90	1134.0
285.00	2868.0
286.00	655.0
289.00	735.0
290.90	330.0
292.00	780.0
293.00	4049.0
293.80	810.0
295.10	1735.0
296.00	61176.0
297.00	7620.0
298.00	514.0
301.00	549.0
301.80	755.0
302.30	576.0
303.00	6620.0
304.00	1391.0
305.10	314.0
308.10	697.0
308.90	727.0
309.90	952.0
311.00	686.0
312.90	686.0
314.00	2664.0
315.00	6953.0
316.00	3839.0
317.10	639.0
321.10	1800.0
322.00	1319.0
323.10	17304.0
324.00	3179.0
324.90	578.0
325.80	324.0
327.00	4365.0
327.90	1735.0
331.90	1361.0
332.90	1888.0
334.00	12200.0
335.00	4411.0
336.00	505.0
338.90	309.0
341.10	1273.0
342.00	819.0
346.00	3350.0
346.90	753.0
351.20	492.0
352.00	6164.0
353.10	4544.0
354.00	6431.0
355.00	1099.0
358.90	322.0
364.90	25512.0
366.00	3990.0
369.90	539.0
371.00	1524.0
372.00	10790.0
373.00	3068.0
377.00	390.0
383.00	2675.0
384.00	929.0
389.90	1363.0
391.10	874.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

391.80	471.0
400.90	670.0
401.90	3763.0
403.00	5431.0
404.00	1689.0
414.80	334.0
421.00	3902.0
422.00	4652.0
423.00	31568.0
424.00	9250.0
424.90	1139.0
441.00	124304.0
442.10	819904.0
443.00	126552.0
444.00	11728.0
445.00	625.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18