

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112122\
 Data File : BN022805.D
 Acq On : 22 Nov 2022 02:11
 Operator : CG/JU
 Sample : N5690-08
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BGCC9

Quant Time: Nov 22 03:52:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Nov 19 05:51:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.846	152	56867	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	272679	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.516	164	180019	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.269	188	371100	20.000	ng/u1	0.00
79) Chrysene-d12	21.474	240	362075	20.000	ng/u1	0.00
88) Perylene-d12	23.862	264	358763	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.175	96	6704	4.025	ng/uL	0.00
4) Pyridine-d5	3.622	84	28602	6.072	ng/u1	0.01
7) Phenol-d5	7.016	99	33768	6.287	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.181	67	120578	37.443	ng/u1	0.00
11) 2-Chlorophenol-d4	7.375	132	100809	25.629	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	67349	15.884	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	74682	35.383	ng/u1	0.00
24) 2-Nitrophenol-d4	9.751	143	75351	31.995	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.293	165	117706	29.317	ng/u1	0.00
31) 4-Chloroaniline-d4	10.816	131	174817	28.586	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	517303	39.507	ng/u1	0.00
49) Acenaphthylene-d8	14.210	160	544570	37.746	ng/u1	0.00
54) 4-Nitrophenol-d4	14.745	143	12498	4.788	ng/u1	0.00
60) Fluorene-d10	15.516	176	433166	40.336	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	75703	32.094	ng/u1	0.00
73) Anthracene-d10	17.375	188	723966	44.553	ng/u1	0.00
81) Pyrene-d10	19.674	212	811243	43.234	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.710	264	797239	44.206	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112122\
Data File : BN022805.D
Acq On : 22 Nov 2022 02:11
Operator : CG/JU
Sample : N5690-08
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BGCC9

Quant Time: Nov 22 03:52:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN11522.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Nov 19 05:51:00 2022
Response via : Initial Calibration

