

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112122\
 Data File : BN022809.D
 Acq On : 22 Nov 2022 05:14
 Operator : CG/JU
 Sample : PB149186BL
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK186

Quant Time: Nov 22 05:48:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 22 05:15:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.846	152	59029	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	283862	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.516	164	191724	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.269	188	392500	20.000	ng/u1	0.00
79) Chrysene-d12	21.468	240	363575	20.000	ng/u1	0.00
88) Perylene-d12	23.863	264	366827	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.175	96	10082	5.831	ng/uL	0.00
4) Pyridine-d5	3.611	84	135009	27.612	ng/u1	0.00
7) Phenol-d5	7.016	99	165971	29.770	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.181	67	111504	33.357	ng/u1	0.00
11) 2-Chlorophenol-d4	7.375	132	126679	31.027	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	136302	30.968	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	67797	30.856	ng/u1	0.00
24) 2-Nitrophenol-d4	9.751	143	74414	30.353	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	122365	29.277	ng/u1	0.00
31) 4-Chloroaniline-d4	10.816	131	196201	30.818	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	447041	32.057	ng/u1	0.00
49) Acenaphthylene-d8	14.210	160	481287	31.323	ng/u1	0.00
54) 4-Nitrophenol-d4	14.734	143	73617	26.481	ng/u1	0.00
60) Fluorene-d10	15.510	176	368342	32.205	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	62101	24.892	ng/u1	0.00
73) Anthracene-d10	17.369	188	578336	33.650	ng/u1	0.00
81) Pyrene-d10	19.674	212	628821	33.374	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.704	264	611924	33.185	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112122\
Data File : BN022809.D
Acq On : 22 Nov 2022 05:14
Operator : CG/JU
Sample : PB149186BL
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK186

Quant Time: Nov 22 05:48:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN11522.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 22 05:15:51 2022
Response via : Initial Calibration

