

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112221\
 Data File : BN017539.D
 Acq On : 22 Nov 2021 10:26
 Operator : CG/JU
 Sample : SST00545
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST005245

Quant Time: Nov 22 14:50:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN112221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 22 14:49:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	253203	20.000	ng/ul	0.00
20) Naphthalene-d8	10.634	136	1106859	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.475	164	705568	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.222	188	1470892	20.000	ng/ul	0.00
79) Chrysene-d12	21.410	240	1415150	20.000	ng/ul	0.00
88) Perylene-d12	23.774	264	1240114	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	14307	2.047	ng/uL	-0.01
4) Pyridine-d5	3.711	84	77316	4.129	ng/ul	0.00
7) Phenol-d5	6.999	99	93937	3.993	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.169	67	66973	4.796	ng/ul	0.00
11) 2-Chlorophenol-d4	7.369	132	75486	4.080	ng/ul	0.00
15) 4-Methylphenol-d8	8.540	113	73930	3.849	ng/ul	-0.01
21) Nitrobenzene-d5	8.987	128	31778	3.707	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.716	143	28145	2.966	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.252	165	66825	3.844	ng/ul	0.00
31) 4-Chloroaniline-d4	10.769	131	107285	4.143	ng/ul	0.00
46) Dimethylphthalate-d6	13.887	166	229649	4.348	ng/ul	0.00
49) Acenaphthylene-d8	14.169	160	289480	4.347	ng/ul	0.00
54) 4-Nitrophenol-d4	14.657	143	28188	2.705	ng/ul	0.00
60) Fluorene-d10	15.469	176	212698	4.715	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.581	200	21464	2.254	ng/ul	0.00
73) Anthracene-d10	17.316	188	327911	4.617	ng/ul	0.00
81) Pyrene-d10	19.616	212	377311	4.546	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.621	264	283130	4.296	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.334	88	12787	1.850	ng/uL	97
5) Pyridine	3.728	79	80107	4.228	ng/ul	97
6) Benzaldehyde	6.975	77	63419	4.524	ng/ul	90
8) Phenol	7.022	94	103408	4.345	ng/ul	95
10) Bis(2-Chloroethyl)ether	7.258	93	89837	4.803	ng/ul	99
12) 2-Chlorophenol	7.405	128	79790	4.170	ng/ul	98
13) 2-Methylphenol	8.275	108	72525	4.005	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.375	45	123315	4.494	ng/ul	96
16) Acetophenone	8.658	105	128041	4.498	ng/ul	95
17) N-Nitroso-di-n-propyla...	8.640	70	59398	4.157	ng/ul#	95
18) 4-Methylphenol	8.605	108	82988	4.148	ng/ul	98
19) Hexachloroethane	8.922	117	33787	4.585	ng/ul	93
22) Nitrobenzene	9.034	77	87108	4.519	ng/ul	95
23) Isophorone	9.557	82	156874	4.008	ng/ul	96
25) 2-Nitrophenol	9.746	139	34190	3.309	ng/ul#	89
26) 2,4-Dimethylphenol	9.810	107	93221	4.502	ng/ul	92
27) Bis(2-Chloroethoxy)met...	10.046	93	117070	4.677	ng/ul	98
29) 2,4-Dichlorophenol	10.281	162	73235	4.241	ng/ul	97
30) Naphthalene	10.687	128	286123	4.751	ng/ul	99
32) 4-Chloroaniline	10.793	127	111860	4.367	ng/ul	99
33) Hexachlorobutadiene	10.987	225	48796	4.641	ng/ul	93
34) Caprolactam	11.516	113	15461	2.488	ng/ul	89
35) 4-Chloro-3-methylphenol	11.916	107	76853	4.000	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112221\
 Data File : BN017539.D
 Acq On : 22 Nov 2021 10:26
 Operator : CG/JU
 Sample : SST00545
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST005245

Quant Time: Nov 22 14:50:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN112221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 22 14:49:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 2-Methylnaphthalene	12.298	142	193838	4.662	ng/ul	100
37) 1-Methylnaphthalene	12.516	142	192681	4.499	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.669	216	98268	4.778	ng/ul	93
40) Hexachlorocyclopentadiene	12.657	237	47439	3.630	ng/ul	96
41) 2,4,6-Trichlorophenol	12.910	196	49818	3.736	ng/ul	97
42) 2,4,5-Trichlorophenol	12.975	196	57252	3.856	ng/ul	96
43) 1,1'-Biphenyl	13.310	154	263872	4.830	ng/ul	99
44) 2-Chloronaphthalene	13.351	162	199467	4.794	ng/ul	98
45) 2-Nitroaniline	13.545	65	39348	3.474	ng/ul	91
47) Dimethylphthalate	13.934	163	233926	4.410	ng/ul	100
48) 2,6-Dinitrotoluene	14.045	165	33171	3.184	ng/ul#	80
50) Acenaphthylene	14.193	152	316110	4.635	ng/ul	100
51) 3-Nitroaniline	14.369	138	37444	3.106	ng/ul	91
52) Acenaphthene	14.540	153	210548	4.774	ng/ul	99
53) 2,4-Dinitrophenol	14.575	184	10896	1.635	ng/ul	97
55) 4-Nitrophenol	14.669	109	26339	3.581	ng/ul	91
56) Dibenzofuran	14.875	168	308513	4.900	ng/ul	100
57) 2,4-Dinitrotoluene	14.828	165	50612	3.369	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	15.098	232	41882	3.417	ng/ul	98
59) Diethylphthalate	15.298	149	231069	4.368	ng/ul	98
61) Fluorene	15.522	166	247642	4.990	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.522	204	120479	4.937	ng/ul	93
63) 4-Nitroaniline	15.528	138	37091	3.113	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.592	198	23041	2.425	ng/ul#	85
67) N-Nitrosodiphenylamine	15.728	169	200439	4.477	ng/ul	98
68) 4-Bromophenyl-phenylether	16.416	248	64174	4.094	ng/ul#	91
69) Hexachlorobenzene	16.534	284	72302	3.964	ng/ul#	90
70) Atrazine	16.681	200	61605	3.750	ng/ul	98
71) Pentachlorophenol	16.869	266	27838	2.542	ng/ul	96
72) Phenanthrene	17.263	178	392480	4.843	ng/ul	99
74) Anthracene	17.351	178	388908	4.739	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.275	216	98011	4.558	ng/uL	92
76) Pentachlorobenzene	14.798	250	96322	4.424	ng/uL	96
77) Carbazole	17.622	167	338412	4.670	ng/ul	98
78) Di-n-butylphthalate	18.198	149	334913	3.910	ng/ul#	99
80) Fluoranthene	19.280	202	444494	4.485	ng/ul	100
82) Pyrene	19.639	202	469605	4.629	ng/ul	99
83) Butylbenzylphthalate	20.551	149	126997	3.362	ng/ul	90
84) 3,3'-Dichlorobenzidine	21.327	252	99167	3.148	ng/ul	96
85) Benzo(a)anthracene	21.392	228	438627	4.708	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.339	149	210017	3.684	ng/ul#	99
87) Chrysene	21.445	228	444804	4.844	ng/ul	99
89) Di-n-octyl phthalate	22.251	149	307687	3.395	ng/ul	100
90) Benzo(b)fluoranthene	23.057	252	396334	4.734	ng/ul	99
91) Benzo(k)fluoranthene	23.104	252	386187	4.742	ng/ul	98
93) Benzo(a)pyrene	23.674	252	375985	4.667	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.215	276	352917	4.294	ng/ul	100
95) Dibenzo(a,h)anthracene	26.233	278	297968	4.289	ng/ul	98
96) Benzo(g,h,i)perylene	26.968	276	291617	4.238	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :

BNA_N

ClientSampleId :

SSTD005245



1,4-Dioxane

Concen: 1.850 ng/uL

RT: 3.334 min Scan# 5

Delta R.T. -0.006 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion: 88 Resp: 12787

Ion Ratio Lower Upper

88 100

43 31.4 24.4 36.6

58 80.1 61.9 92.9



Pyridine

Concen: 4.228 ng/uL

RT: 3.728 min Scan# 126

Delta R.T. 0.000 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26




Tgt Ion: 79 Resp: 80107

Ion Ratio Lower Upper


79 100

52 79.3 66.7 100.1

51 38.1 30.6 46.0



Abundance Scan 679 (6.981 min): BN017541.D\data.ms (-672#6)



Benzaldehyde

Concen: 4.524 ng/u1

RT: 6.975 min Scan# 6

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Instrument :

BNA_N

ClientSampleId :

SSTD005245

Tgt Ion: 77 Resp: 63419



Ion Ratio Lower Upper

77 100


105 95.0 83.8 125.6

106 88.0 79.6 119.4


Abundance Scan 678 (6.975 min): BN017539.D\data.ms



Abundance Scan 678 (6.975 min): BN017539.D\data.ms (-645) (-)



Abundance Scan 688 (7.034 min): BN017541.D\data.ms (-680#8)



Phenol

Concen: 4.345 ng/u1

RT: 7.022 min Scan# 686

Delta R.T. -0.011 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Tgt Ion: 94 Resp: 103408



Ion Ratio Lower Upper

94 100


65 27.4 21.7 32.5


66 39.8 28.3 42.5

Abundance Scan 686 (7.022 min): BN017539.D\data.ms



Abundance Scan 686 (7.022 min): BN017539.D\data.ms (-654) (-)









Bis(2-Chloroethyl)ether
 Concen: 4.803 ng/ul
 RT: 7.258 min Scan# 726
 Delta R.T. -0.011 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument : BNA_N
 ClientSampleId : SST005245


Tgt Ion:	93	Resp:	89837
Ion	Ratio	Lower	Upper
93	100		
63	78.7	63.6	95.4
95	31.9	26.2	39.2



2-Chlorophenol
 Concen: 4.170 ng/ul
 RT: 7.405 min Scan# 751
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:	128	Resp:	79790
Ion	Ratio	Lower	Upper
128	100		
64	45.3	36.2	54.2
130	29.9	26.4	39.6





2-Methylphenol

Concen: 4.005 ng/ul

RT: 8.275 min Scan# 899

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Instrument :

BNA_N

ClientSampleId :





SSTD005245

Tgt Ion:108 Resp: 72525

Ion Ratio Lower Upper

108 100

107 88.0 71.1 106.7



2,2'-oxybis(1-Chloropropane)

Concen: 4.494 ng/ul

RT: 8.375 min Scan# 916

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26




Tgt Ion: 45 Resp: 123315


Ion Ratio Lower Upper

45 100

77 15.0 10.7 16.1

79 11.4 8.1 12.1





Acetophenone

Concen: 4.498 ng/ul

RT: 8.658 min Scan# 965

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion:105 Resp: 128041

Ion Ratio Lower Upper

105 100

77 78.2 58.6 87.8

51 27.0 21.3 31.9



N-Nitroso-di-n-propylamine

Concen: 4.157 ng/ul

RT: 8.640 min Scan# 961

Delta R.T. -0.011 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Tgt Ion: 70 Resp: 59398




Ion Ratio Lower Upper


70 100

42 58.5 49.8 74.8

101 7.9 8.2 12.2#

130 19.4 18.0 27.0





4-Methylphenol

Concen: 4.148 ng/ul

RT: 8.605 min Scan# 918

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Instrument :

BNA_N

ClientSampleId :





SSTD005245

Tgt Ion:108 Resp: 82988

Ion Ratio Lower Upper

108 100

107 108.6 88.7 133.1



Hexachloroethane

Concen: 4.585 ng/ul

RT: 8.922 min Scan# 1009

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26




Tgt Ion:117 Resp: 33787


Ion Ratio Lower Upper

117 100

201 87.2 78.1 117.1

199 56.8 46.6 70.0





Nitrobenzene

Concen: 4.519 ng/u1

RT: 9.034 min Scan# 1

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion: 77 Resp: 87108

Ion Ratio Lower Upper

77 100

123 43.8 37.9 56.9

65 14.4 10.6 16.0



Isophorone

Concen: 4.008 ng/u1

RT: 9.557 min Scan# 1117

Delta R.T. -0.006 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26




Tgt Ion: 82 Resp: 156874

Ion Ratio Lower Upper


82 100

95 6.9 5.6 8.4

138 16.0 15.0 22.4



Abundance Scan 1150 (9.751 min): BN017541.D\data.ms (-11425) (-)



2-Nitrophenol

Concen: 3.309 ng/ul

RT: 9.746 min Scan# 11425

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Instrument :

BNA_N

ClientSampleId :

SSTD005245

Tgt Ion:139 Resp: 34190



Ion Ratio Lower Upper

139 100


65 53.7 38.4 57.6

109 35.5 21.9 32.9#


Abundance Scan 1149 (9.746 min): BN017539.D\data.ms



Abundance Scan 1149 (9.746 min): BN017539.D\data.ms (-1116) (-)



Abundance Scan 1161 (9.816 min): BN017541.D\data.ms (-11526) (-)



2,4-Dimethylphenol

Concen: 4.502 ng/ul

RT: 9.810 min Scan# 1160

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Tgt Ion:107 Resp: 93221



Ion Ratio Lower Upper

107 100


121 46.3 41.2 61.8


122 81.5 71.4 107.0

Abundance Scan 1160 (9.810 min): BN017539.D\data.ms



Abundance Scan 1160 (9.810 min): BN017539.D\data.ms (-1127) (-)









Bis(2-Chloroethoxy)methane
 Concen: 4.677 ng/ul
 RT: 10.046 min Scan# 1167
 Delta R.T. -0.005 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument :
 BNA_N
 ClientSampleId :
 SST005245


Tgt Ion:	93	Resp:	117070
Ion	Ratio	Lower	Upper
93	100		
95	30.9	25.4	38.2
123	12.0	9.0	13.6



2,4-Dichlorophenol
 Concen: 4.241 ng/ul
 RT: 10.281 min Scan# 1240
 Delta R.T. -0.005 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:	162	Resp:	73235
Ion	Ratio	Lower	Upper
162	100		
164	65.6	53.6	80.4
98	40.6	29.9	44.9









Naphthalene
 Concen: 4.751 ng/u1
 RT: 10.687 min Scan# 1309
 Delta R.T. -0.005 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument :
 BNA_N
 ClientSampleId :
 SST005245


Tgt Ion:128 Resp: 286123
 Ion Ratio Lower Upper
 128 100
 129 10.7 8.4 12.6
 127 13.1 10.3 15.5



4-Chloroaniline
 Concen: 4.367 ng/u1
 RT: 10.793 min Scan# 1327
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:127 Resp: 111860
 Ion Ratio Lower Upper
 127 100
 129 31.9 25.8 38.6









Hexachlorobutadiene
Concen: 4.641 ng/u1
RT: 10.987 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:225 Resp: 48796
Ion Ratio Lower Upper
225 100
223 64.2 49.4 74.2
227 69.9 49.0 73.6



Caprolactam
Concen: 2.488 ng/u1
RT: 11.516 min Scan# 1450
Delta R.T. -0.023 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26

Tgt Ion:113 Resp: 15461
Ion Ratio Lower Upper
113 100
55 166.1 141.4 212.2
56 110.4 103.9 155.9









4-Chloro-3-methylphenol
Concen: 4.000 ng/ul
RT: 11.916 min Scan# 1518
Delta R.T. -0.005 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:107 Resp: 76853
Ion Ratio Lower Upper
107 100
144 25.7 20.6 30.8
142 77.6 64.1 96.1



2-Methylnaphthalene
Concen: 4.662 ng/ul
RT: 12.298 min Scan# 1583
Delta R.T. -0.005 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26

Tgt Ion:142 Resp: 193838
Ion Ratio Lower Upper
142 100
141 87.7 70.2 105.2









1-Methylnaphthalene
Concen: 4.499 ng/u1
RT: 12.516 min Scan# 1620
Delta R.T. -0.005 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:142 Resp: 192681
Ion Ratio Lower Upper
142 100
141 95.8 72.2 108.2
116 4.0 3.0 4.6



1,2,4,5-Tetrachlorobenzene
Concen: 4.778 ng/u1
RT: 12.669 min Scan# 1646
Delta R.T. -0.005 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26

Tgt Ion:216 Resp: 98268
Ion Ratio Lower Upper
216 100
214 73.4 65.6 98.4
179 22.4 17.4 26.0
108 19.4 14.0 21.0









Hexachlorocyclopentadiene
 Concen: 3.630 ng/ul
 RT: 12.657 min Scan# 1644
 Delta R.T. -0.005 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument :
 BNA_N
 ClientSampleId :
 SST005245


Tgt Ion:237 Resp: 47439
 Ion Ratio Lower Upper
 237 100
 235 62.0 51.9 77.9
 272 11.1 10.9 16.3



2,4,6-Trichlorophenol
 Concen: 3.736 ng/ul
 RT: 12.910 min Scan# 1687
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:196 Resp: 49818
 Ion Ratio Lower Upper
 196 100
 198 97.2 75.8 113.8
 200 28.5 24.0 36.0









2,4,5-Trichlorophenol
Concen: 3.856 ng/uI
RT: 12.975 min Scan# 1698
Delta R.T. -0.005 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:	196	Resp:	57252
Ion	Ratio	Lower	Upper
196	100		
198	90.7	76.4	114.6
200	29.2	24.4	36.6



1,1'-Biphenyl
Concen: 4.830 ng/uI
RT: 13.310 min Scan# 1755
Delta R.T. -0.005 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26

Tgt Ion:	154	Resp:	263872
Ion	Ratio	Lower	Upper
154	100		
153	39.1	32.1	48.1
76	13.6	10.6	15.8









2-Chloronaphthalene
Concen: 4.794 ng/u1
RT: 13.351 min Scan# 1744
Delta R.T. 0.000 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:	162	Resp:	199467
Ion Ratio	Lower	Upper	
162	100		
164	32.0	25.6	38.4
127	40.9	31.1	46.7



2-Nitroaniline
Concen: 3.474 ng/u1
RT: 13.545 min Scan# 1795
Delta R.T. -0.005 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26

Tgt Ion:	65	Resp:	39348
Ion Ratio	Lower	Upper	
65	100		
92	65.3	56.9	85.3
138	101.2	89.3	133.9









Dimethylphthalate
Concen: 4.410 ng/ul
RT: 13.934 min Scan# 1861
Delta R.T. -0.005 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:163 Resp: 233926
Ion Ratio Lower Upper
163 100
194 4.5 3.3 4.9
164 9.8 7.8 11.8



2,6-Dinitrotoluene
Concen: 3.184 ng/ul
RT: 14.045 min Scan# 1880
Delta R.T. -0.005 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26

Tgt Ion:165 Resp: 33171
Ion Ratio Lower Upper
165 100
89 69.5 42.3 63.5#
121 25.9 16.9 25.3#









Acenaphthylene
 Concen: 4.635 ng/u1
 RT: 14.193 min Scan# 1905
 Delta R.T. -0.005 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005245


Tgt Ion:152 Resp: 316110
 Ion Ratio Lower Upper
 152 100
 151 19.9 15.9 23.9
 153 12.9 10.6 15.8



3-Nitroaniline
 Concen: 3.106 ng/u1
 RT: 14.369 min Scan# 1935
 Delta R.T. -0.005 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:138 Resp: 37444
 Ion Ratio Lower Upper
 138 100
 108 12.0 8.5 12.7
 92 121.9 89.4 134.0





Acenaphthene

Concen: 4.774 ng/u1

RT: 14.540 min Scan# 1965

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion:153 Resp: 210548

Ion Ratio Lower Upper

153 100

152 46.3 37.6 56.4

154 89.7 72.2 108.4



2,4-Dinitrophenol

Concen: 1.635 ng/u1

RT: 14.575 min Scan# 1970

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26




Tgt Ion:184 Resp: 10896


Ion Ratio Lower Upper

184 100

63 72.9 54.2 81.4

154 62.3 50.2 75.2









4-Nitrophenol
 Concen: 3.581 ng/u1
 RT: 14.669 min Scan# 1986
 Delta R.T. -0.011 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument :
 BNA_N
 ClientSampleId :
 SST005245


Tgt Ion:109 Resp: 26339
 Ion Ratio Lower Upper
 109 100
 139 130.7 121.5 182.3
 65 135.1 109.4 164.0



Dibenzofuran
 Concen: 4.900 ng/u1
 RT: 14.875 min Scan# 2021
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:168 Resp: 308513
 Ion Ratio Lower Upper
 168 100
 139 37.7 29.9 44.9









2,4-Dinitrotoluene
 Concen: 3.369 ng/u1
 RT: 14.828 min Scan# 2057
 Delta R.T. -0.005 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument :
 BNA_N
 ClientSampleId :
 SST005245


Tgt Ion:165	Resp:	50612
Ion Ratio	Lower	Upper
165	100	
63	51.2	35.4 53.0
182	2.5	2.3 3.5



2,3,4,6-Tetrachlorophenol
 Concen: 3.417 ng/u1
 RT: 15.098 min Scan# 2059
 Delta R.T. -0.006 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:232	Resp:	41882
Ion Ratio	Lower	Upper
232	100	
131	50.6	40.1 60.1
130	2.4	2.0 3.0
166	34.6	26.2 39.4





Diethylphthalate

Concen: 4.368 ng/uI

RT: 15.298 min Scan# 2

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion:149 Resp: 231069

Ion Ratio Lower Upper

149 100

177 20.7 17.6 26.4

150 12.4 10.0 15.0



Fluorene

Concen: 4.990 ng/uI

RT: 15.522 min Scan# 2131

Delta R.T. -0.005 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26




Tgt Ion:166 Resp: 247642


Ion Ratio Lower Upper

166 100

165 92.1 77.4 116.0

167 13.1 11.0 16.6









4-Chlorophenyl-phenylether
Concen: 4.937 ng/ul
RT: 15.522 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:204 Resp: 120479
Ion Ratio Lower Upper
204 100
206 33.6 27.0 40.4
141 66.7 47.2 70.8



4-Nitroaniline
Concen: 3.113 ng/ul
RT: 15.528 min Scan# 2132
Delta R.T. -0.011 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26

Tgt Ion:138 Resp: 37091
Ion Ratio Lower Upper
138 100
92 52.6 39.0 58.4
108 80.7 56.7 85.1









4,6-Dinitro-2-methylphenol
 Concen: 2.425 ng/ul
 RT: 15.592 min Scan# 2144
 Delta R.T. -0.005 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005245


Tgt Ion:198 Resp: 23041
 Ion Ratio Lower Upper
 198 100
 182 5.4 3.1 4.7#
 77 29.5 17.4 26.2#



N-Nitrosodiphenylamine
 Concen: 4.477 ng/ul
 RT: 15.728 min Scan# 2166
 Delta R.T. -0.006 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:169 Resp: 200439
 Ion Ratio Lower Upper
 169 100
 168 64.9 50.8 76.2
 167 36.9 29.0 43.6









4-Bromophenyl-phenylether
 Concen: 4.094 ng/ul
 RT: 16.416 min Scan# 2283
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument :
 BNA_N
 ClientSampleId :
 SST005245


Tgt Ion:248 Resp: 64174
 Ion Ratio Lower Upper
 248 100
 250 97.6 78.9 118.3
 141 90.4 59.8 89.6#



Hexachlorobenzene
 Concen: 3.964 ng/ul
 RT: 16.534 min Scan# 2303
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:284 Resp: 72302
 Ion Ratio Lower Upper
 284 100
 142 44.4 28.2 42.2#
 249 32.3 24.5 36.7









Atrazine
Concen: 3.750 ng/u1
RT: 16.681 min Scan# 2329
Delta R.T. -0.006 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:200 Resp: 61605
Ion Ratio Lower Upper
200 100
173 26.5 21.6 32.4
215 50.1 38.8 58.2



Pentachlorophenol
Concen: 2.542 ng/u1
RT: 16.869 min Scan# 2360
Delta R.T. -0.005 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26

Tgt Ion:266 Resp: 27838
Ion Ratio Lower Upper
266 100
264 63.1 52.2 78.2
268 60.2 51.4 77.0









Scan 2427 (17.263 min): BN017541.D\data.ms (-2422) (-)
 Phenanthrene
 Concen: 4.843 ng/u1
 RT: 17.263 min Scan# 2427
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument : BNA_N
 ClientSampleId : SST005245


Tgt Ion:178	Resp:	392480	
Ion	Ratio	Lower	Upper
178	100		
179	15.1	12.5	18.7
176	18.8	14.6	22.0



Scan 2443 (17.357 min): BN017541.D\data.ms (-2444) (-)
 Anthracene
 Concen: 4.739 ng/u1
 RT: 17.351 min Scan# 2442
 Delta R.T. -0.005 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:178	Resp:	388908	
Ion	Ratio	Lower	Upper
178	100		
179	15.7	12.2	18.2
176	17.2	14.2	21.2









Scan 1750 (13.280 min): BN017541.D\data.ms (-1715) (-)
 1,2,3,4-Tetrachlorobenzene
 Concen: 4.558 ng/uL
 RT: 13.275 min Scan# 1715
 Delta R.T. -0.005 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument : BNA_N
 ClientSampleId : SST005245


Tgt Ion:	216	Resp:	98011
Ion Ratio	Lower	Upper	
216	100		
214	71.6	61.5	92.3
218	44.1	40.6	61.0



Scan 2009 (14.804 min): BN017541.D\data.ms (-2008) (-)
 Pentachlorobenzene
 Concen: 4.424 ng/uL
 RT: 14.798 min Scan# 2008
 Delta R.T. -0.006 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26


Tgt Ion:	250	Resp:	96322
Ion Ratio	Lower	Upper	
250	100		
248	63.9	53.2	79.8
252	61.8	52.5	78.7








Scan 2488 (17.621 min): BN017541.D\data.ms (-2487) (-)
 Carbazole
 Concen: 4.670 ng/u1
 RT: 17.622 min Scan# 2487
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Instrument : BNA_N
 ClientSampleId : SST005245



Tgt Ion:167 Resp: 338412




Ion	Ratio	Lower	Upper
167	100		
166	20.8	17.6	26.4
139	13.0	10.6	16.0




Scan 2586 (18.198 min): BN017541.D\data.ms (-2578) (-)
 Di-n-butylphthalate
 Concen: 3.910 ng/u1
 RT: 18.198 min Scan# 2586
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:149 Resp: 334913

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.4	11.2
104	5.7	3.8	5.6#









Fluoranthene
 Concen: 4.485 ng/u1
 RT: 19.280 min Scan# 2770
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005245


Tgt Ion:	202	Resp:	444494
Ion	Ratio	Lower	Upper
202	100		
101	12.0	9.6	14.4
100	8.7	7.3	10.9



Pyrene
 Concen: 4.629 ng/u1
 RT: 19.639 min Scan# 2831
 Delta R.T. -0.005 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:	202	Resp:	469605
Ion	Ratio	Lower	Upper
202	100		
101	15.1	11.8	17.6
100	12.4	9.6	14.4









Butylbenzylphthalate
 Concen: 3.362 ng/u1
 RT: 20.551 min Scan# 2986
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Instrument :
 BNA_N
 ClientSampleId :
 SSTD005245

Tgt Ion:149 Resp: 126997




Ion	Ratio	Lower	Upper
149	100		
91	78.0	54.5	81.7
206	20.8	17.7	26.5




3,3'-Dichlorobenzidine
 Concen: 3.148 ng/u1
 RT: 21.327 min Scan# 3118
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:252 Resp: 99167

Ion	Ratio	Lower	Upper
252	100		
254	62.9	47.6	71.4
126	14.1	10.1	15.1









Benzo(a)anthracene
 Concen: 4.708 ng/u1
 RT: 21.392 min Scan# 3129
 Delta R.T. -0.006 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument :
 BNA_N
 ClientSampleId :
 SST005245


Tgt Ion:228 Resp: 438627
 Ion Ratio Lower Upper
 228 100
 229 18.6 15.8 23.8
 226 25.1 21.8 32.8



Bis(2-ethylhexyl)phthalate
 Concen: 3.684 ng/u1
 RT: 21.339 min Scan# 3120
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:149 Resp: 210017
 Ion Ratio Lower Upper
 149 100
 167 27.8 22.6 33.8
 279 3.4 3.8 5.6#





Chrysene

Concen: 4.844 ng/u1

RT: 21.445 min Scan# 3138

Delta R.T. 0.000 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion:228 Resp: 444804

Ion Ratio Lower Upper

228 100

226 30.4 24.1 36.1

229 19.2 15.6 23.4



Di-n-octyl phthalate

Concen: 3.395 ng/u1




RT: 22.251 min Scan# 3275


Delta R.T. 0.000 min

Lab File: BN017539.D

Acq: 22 Nov 2021 10:26

Tgt Ion:149 Resp: 307687









Benzo(b)fluoranthene
Concen: 4.734 ng/u1
RT: 23.057 min Scan# 3412
Delta R.T. 0.000 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:252 Resp: 396334
Ion Ratio Lower Upper
252 100
253 23.2 18.4 27.6
125 12.2 9.4 14.0



Benzo(k)fluoranthene
Concen: 4.742 ng/u1
RT: 23.104 min Scan# 3420
Delta R.T. 0.000 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26

Tgt Ion:252 Resp: 386187
Ion Ratio Lower Upper
252 100
253 20.3 17.4 26.2
125 10.9 8.6 13.0









Benzo(a)pyrene
 Concen: 4.667 ng/u1
 RT: 23.674 min Scan# 3517
 Delta R.T. 0.000 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26




Instrument :
 BNA_N
 ClientSampleId :
 SST005245


Tgt Ion:252 Resp: 375985
 Ion Ratio Lower Upper
 252 100
 253 19.5 17.5 26.3
 125 11.9 9.7 14.5



Indeno(1,2,3-cd)pyrene
 Concen: 4.294 ng/u1
 RT: 26.215 min Scan# 3949
 Delta R.T. -0.011 min
 Lab File: BN017539.D
 Acq: 22 Nov 2021 10:26

Tgt Ion:276 Resp: 352917
 Ion Ratio Lower Upper
 276 100
 138 26.5 21.4 32.2
 277 24.5 19.5 29.3









Dibenzo(a,h)anthracene
Concen: 4.289 ng/ul
RT: 26.233 min Scan# 3995
Delta R.T. -0.005 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26

Instrument :
BNA_N
ClientSampleId :
SSTD005245

Tgt Ion:278 Resp: 297968
Ion Ratio Lower Upper
278 100
139 19.9 14.7 22.1
279 24.1 19.7 29.5



Benzo(g,h,i)perylene
Concen: 4.238 ng/ul
RT: 26.968 min Scan# 4077
Delta R.T. 0.000 min
Lab File: BN017539.D
Acq: 22 Nov 2021 10:26

Tgt Ion:276 Resp: 291617
Ion Ratio Lower Upper
276 100
138 22.5 18.2 27.4
277 23.7 17.4 26.0

