

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112223\
 Data File : BN028824.D
 Acq On : 21 Nov 2023 23:16
 Operator : MA/JU
 Sample : PB156981BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK981

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/23/2023
 Supervised By :mohammad ahmed 11/25/2023

Quant Time: Nov 21 23:53:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 21 23:49:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.719	152	4803m	0.400	ng/ul	-0.02
4) Naphthalene-d8	10.457	136	16854m	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.302	164	10310	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.060	188	20247	0.400	ng/ul	0.00
17) Chrysene-d12	21.286	240	14019	0.400	ng/ul	# 0.02
23) Perylene-d12	23.543	264	20565	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.099	96	45501	8.572	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.123	152	5878m	0.232	ng/ul	0.05
18) Fluoranthene-d10	19.094	212	20499	0.454	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112223\
Data File : BN028824.D
Acq On : 21 Nov 2023 23:16
Operator : MA/JU
Sample : PB156981BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK981

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/23/2023
Supervised By :mohammad ahmed 11/25/2023

Quant Time: Nov 21 23:53:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 21 23:49:43 2023
Response via : Initial Calibration

