



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112223\
Data File : BN028814.D
Acq On : 21 Nov 2023 17:19
Operator : MA/JU
Sample : DFTPP390
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP390

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Nov 21 23:49:43 2023



AutoFind: Scans 2568, 2569, 2570; Background Corrected with Scan 2561

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.9	141666	PASS
68	69	0.00	2	1.7	2365	PASS
69	198	0.00	100	35.3	143025	PASS
70	69	0.00	2	0.4	551	PASS
127	198	10	80	47.4	192171	PASS
197	198	0.00	2	0.5	2231	PASS
198	198	100	100	100.0	405739	PASS
199	198	5	9	6.6	26643	PASS
275	198	10	60	26.7	108517	PASS
365	198	1	100	3.4	13742	PASS
441	198	0.01	100	13.5	54651	PASS
442	198	50	100	86.9	352597	PASS
443	442	15	24	20.1	70731	PASS

m/z	Abundance
35.90	170.0
39.10	442.0
39.90	1315.0
41.00	693.0
42.00	245.0
43.00	928.0
44.00	4140.0
45.00	938.0
46.90	209.0
48.90	465.0
50.00	214.0
51.00	521.0
51.90	312.0
52.80	301.0
54.90	497.0
56.00	298.0
57.00	365.0
58.00	470.0
59.00	935.0
59.90	332.0
60.90	186.0
67.10	222.0
68.00	180.0
69.00	396.0
69.90	203.0
71.00	181.0
72.10	176.0
73.00	9081.0
74.00	1096.0
74.90	1234.0
77.00	573.0
78.00	636.0
78.90	233.0
81.00	292.0
82.90	280.0
83.90	427.0
85.00	272.0
85.90	324.0
86.90	203.0
88.90	303.0
90.90	417.0
93.00	264.0
94.90	224.0
95.90	1168.0
96.80	431.0
103.00	449.0
103.80	154.0
104.90	317.0
107.00	304.0
115.00	337.0
117.00	426.0
118.00	164.0
118.90	520.0
121.10	277.0
129.00	159.0
131.10	289.0
132.90	1845.0
133.80	232.0
135.00	2377.0
136.00	291.0
136.90	200.0
144.90	300.0
147.00	3900.0
148.00	786.0
148.90	690.0
161.00	194.0
162.90	404.0
164.90	208.0
176.90	589.0
177.90	156.0
179.00	234.0
189.00	237.0
191.00	2010.0
191.90	640.0
193.00	3268.0
194.00	548.0
194.90	327.0
195.10	331.0
204.90	294.0
207.00	13083.0
208.00	2991.0
209.10	1909.0
209.90	551.0
221.10	4448.0
222.10	1111.0
223.00	762.0
234.90	264.0
248.90	364.0
250.90	374.0
253.00	351.0
265.00	497.0
266.10	204.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390

267.00	1416.0
267.90	430.0
269.00	570.0
281.00	4802.0
282.00	1292.0
283.10	1297.0
284.00	271.0
285.00	164.0
295.10	1054.0
296.00	422.0
296.90	412.0
324.90	316.0
326.90	650.0
327.90	171.0
338.90	158.0
341.00	1293.0
342.00	420.0
343.00	524.0
344.00	169.0
355.10	1785.0
356.00	662.0
357.00	567.0
369.00	306.0
370.00	190.0
400.90	497.0
401.90	184.0
415.00	482.0
416.00	274.0
417.00	169.0
429.00	525.0
430.00	260.0
431.10	220.0
475.00	326.0
476.10	160.0
489.00	236.0
503.10	187.0
549.00	161.0

Instrument :

BNA_N

ClientSampleId :

DFTPP390

m/z	Abundance
35.90	167.0
37.00	599.0
38.00	1507.0
39.00	8619.0
39.90	1577.0
41.00	722.0
42.00	372.0
43.00	1094.0
44.00	3748.0
45.00	1385.0
46.90	325.0
49.00	1301.0
50.00	29480.0
51.00	115648.0
52.00	6412.0
53.00	595.0
55.00	1277.0
56.00	3633.0
57.00	8687.0
58.00	880.0
59.00	1023.0
59.90	386.0
61.00	1824.0
62.00	1599.0
63.00	4568.0
64.00	726.0
65.00	2156.0
66.00	347.0
66.90	311.0
68.00	2215.0
69.00	117944.0
70.00	722.0
71.00	176.0
72.00	305.0
73.00	10320.0
74.00	12885.0
75.00	19208.0
76.10	6641.0
77.00	130880.0
78.00	9812.0
79.00	9333.0
80.00	6574.0
81.00	9328.0
82.00	2468.0
83.00	2538.0
84.00	516.0
85.00	1833.0
86.00	2843.0
87.00	1453.0
88.00	494.0
88.90	463.0
90.90	2366.0
92.00	2390.0
93.00	14573.0
94.00	1014.0
95.00	383.0
95.90	1238.0
97.00	536.0
98.00	11845.0
99.00	9572.0
100.00	855.0
101.00	5481.0
102.00	466.0
102.90	2045.0
104.00	4020.0
105.00	3629.0
106.00	962.0
107.00	41144.0
108.00	6356.0
109.00	1179.0
110.00	73576.0
111.00	12043.0
111.90	1512.0
113.10	481.0
113.80	208.0
115.00	553.0
116.00	2354.0
117.00	32080.0
118.00	2592.0
119.00	684.0
120.00	568.0
121.00	486.0
121.90	2976.0
123.00	4363.0
123.90	1883.0
124.90	2088.0
126.10	892.0
127.00	145984.0
128.00	10909.0
129.00	55248.0
130.00	5010.0
131.00	1405.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390

132.00	619.0
133.00	1924.0
134.00	1913.0
135.00	7086.0
136.00	1903.0
136.90	2478.0
138.00	502.0
139.00	388.0
140.00	982.0
141.00	7333.0
142.00	2560.0
142.90	1386.0
144.00	478.0
144.90	636.0
145.90	1309.0
147.00	7298.0
148.00	8659.0
149.00	2700.0
150.00	539.0
151.10	1143.0
152.00	431.0
153.00	2375.0
153.90	1667.0
155.00	3929.0
156.00	6014.0
157.00	1187.0
157.90	1426.0
159.00	1086.0
160.00	2630.0
161.00	3478.0
162.00	1014.0
162.90	460.0
164.00	316.0
164.90	2903.0
166.00	2283.0
167.00	12678.0
168.00	7558.0
169.00	1334.0
169.90	519.0
170.90	552.0
171.90	1218.0
173.00	1519.0
174.00	2985.0
175.00	5489.0
176.00	2043.0
176.90	2889.0
178.00	886.0
179.00	10455.0
180.00	7743.0
181.00	3541.0
182.20	578.0
182.90	204.0
184.00	608.0
185.00	4893.0
186.00	36536.0
187.00	11468.0
188.00	1317.0
188.90	2701.0
189.90	513.0
191.00	2760.0
192.00	3559.0
193.00	5851.0
194.00	1438.0
195.00	894.0
196.00	7297.0
197.00	1460.0
198.00	269376.0
199.00	18608.0
199.90	1360.0
201.60	1121.0
203.00	1881.0
204.00	8415.0
205.00	16338.0
206.10	70464.0
207.00	24744.0
208.00	5635.0
209.00	3103.0
210.10	1561.0
211.00	2612.0
211.90	621.0
213.00	274.0
213.80	178.0
215.10	895.0
216.00	1660.0
217.00	18832.0
218.00	2580.0
219.10	303.0
221.00	15053.0
222.00	3216.0
223.00	4641.0
224.00	39896.0
225.00	9975.0
226.00	1243.0
227.00	16301.0
228.00	2378.0
229.00	3554.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390

230.00	493.0
231.00	1412.0
232.00	319.0
233.00	345.0
234.00	992.0
234.90	1416.0
235.90	660.0
237.00	2105.0
238.00	225.0
239.00	716.0
239.90	582.0
240.90	870.0
242.00	2157.0
243.10	1889.0
244.00	29336.0
245.00	4032.0
246.00	6030.0
247.00	1304.0
248.10	267.0
249.00	1207.0
250.00	246.0
251.00	618.0
252.00	377.0
252.90	1181.0
254.10	1117.0
255.00	148992.0
256.00	22544.0
257.00	2011.0
258.00	8183.0
259.00	1448.0
260.00	247.0
261.00	256.0
264.00	218.0
264.90	3651.0
266.00	835.0
267.00	1371.0
267.90	418.0
269.10	721.0
269.90	300.0
270.90	445.0
271.90	295.0
273.00	4774.0
274.00	11422.0
275.00	67984.0
276.00	8651.0
277.00	5424.0
278.00	1042.0
278.80	446.0
281.00	4722.0
282.00	1471.0
283.00	2159.0
284.00	782.0
285.00	1123.0
285.90	260.0
288.90	239.0
290.00	232.0
290.90	208.0
292.00	291.0
293.00	1418.0
294.00	276.0
295.00	1574.0
296.00	17240.0
297.00	3165.0
297.90	304.0
301.00	167.0
302.20	193.0
303.00	2242.0
304.00	449.0
308.00	335.0
308.90	274.0
309.90	197.0
311.10	216.0
313.00	314.0
313.90	862.0
315.00	1962.0
315.90	1028.0
316.80	226.0
320.90	567.0
322.00	307.0
323.00	5752.0
324.00	1091.0
324.90	352.0
326.00	178.0
327.00	1447.0
327.90	559.0
329.00	263.0
332.00	452.0
332.90	640.0
334.00	3213.0
335.00	909.0
339.10	222.0
341.00	2081.0
342.00	407.0
343.10	401.0
343.90	163.0
345.90	1164.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390

346.80	207.0
351.90	1562.0
353.00	1057.0
354.00	1606.0
355.00	2376.0
356.00	636.0
357.00	570.0
358.00	160.0
358.90	203.0
365.00	7609.0
366.00	1132.0
366.90	165.0
369.10	340.0
369.90	245.0
371.00	493.0
372.00	2693.0
372.90	653.0
382.90	755.0
384.00	361.0
385.00	202.0
387.00	188.0
390.00	420.0
390.90	169.0
392.00	195.0
400.90	591.0
401.90	1169.0
403.00	1872.0
404.00	575.0
404.90	153.0
415.10	674.0
416.00	254.0
420.90	1604.0
421.90	1070.0
423.00	10047.0
424.00	2769.0
425.00	393.0
429.10	617.0
430.00	219.0
431.20	178.0
441.00	23992.0
442.00	169984.0
443.00	35400.0
444.00	3811.0
444.90	272.0
461.10	197.0
475.00	300.0
489.10	321.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390

m/z	Abundance
35.90	200.0
37.00	782.0
38.00	2177.0
39.00	12957.0
40.00	1506.0
41.00	860.0
42.00	247.0
43.00	1195.0
44.00	3733.0
45.00	1133.0
46.90	216.0
48.00	173.0
49.00	1449.0
50.00	43696.0
51.00	168000.0
52.00	8787.0
53.00	670.0
54.90	1194.0
56.00	5011.0
57.00	11651.0
58.00	959.0
59.00	1040.0
60.00	347.0
61.00	2428.0
62.00	2386.0
63.00	6352.0
64.00	1057.0
65.00	3593.0
66.00	383.0
66.90	413.0
68.00	2894.0
69.00	168704.0
70.00	810.0
71.10	228.0
72.20	199.0
73.00	9792.0
74.00	18264.0
75.00	27576.0
76.00	9911.0
77.00	192000.0
78.00	13461.0
79.00	13630.0
80.00	10271.0
81.00	13514.0
82.00	3687.0
83.00	3330.0
84.00	711.0
85.00	2815.0
86.00	3927.0
87.00	2019.0
88.00	831.0
88.90	570.0
91.00	3434.0
91.90	3688.0
93.00	22808.0
94.00	1677.0
95.00	578.0
96.00	2001.0
97.20	569.0
98.00	16282.0
99.00	14124.0
100.00	1290.0
101.00	8295.0
102.00	435.0
103.00	3382.0
104.00	5527.0
104.90	4757.0
106.00	1766.0
107.00	63896.0
108.00	9591.0
109.00	1663.0
110.00	114816.0
111.00	17536.0
112.00	2306.0
112.90	824.0
114.10	195.0
115.00	624.0
115.90	3745.0
117.00	50192.0
118.00	3469.0
119.00	847.0
120.00	1079.0
120.90	697.0
122.00	4528.0
123.00	6310.0
124.00	3047.0
125.00	3018.0
126.20	1407.0
127.00	225408.0
128.00	17216.0
129.00	88088.0
130.00	7439.0

Instrument :

BNA_N

ClientSampleId :

DFTPP390

130.90	1789.0
131.90	906.0
132.90	2343.0
133.90	3050.0
135.00	9713.0
136.00	3089.0
137.00	3794.0
138.00	923.0
138.90	711.0
139.90	1281.0
141.00	10795.0
142.00	3632.0
143.00	2547.0
144.00	626.0
144.90	1019.0
145.90	2413.0
147.00	9609.0
148.00	13642.0
149.00	3247.0
150.10	755.0
151.00	1475.0
151.60	874.0
153.00	3823.0
154.00	2753.0
155.00	6732.0
156.00	8858.0
157.00	1846.0
157.90	2113.0
159.00	1820.0
159.90	3424.0
161.00	4703.0
162.00	1669.0
163.00	706.0
164.00	694.0
165.00	4582.0
166.00	3235.0
167.00	23008.0
168.00	11484.0
169.00	2045.0
170.00	852.0
170.90	1370.0
172.00	2110.0
173.00	2423.0
174.00	5011.0
175.00	9499.0
175.90	2407.0
177.00	5011.0
177.90	1707.0
179.00	17744.0
180.00	12212.0
181.00	5947.0
182.00	859.0
182.90	392.0
183.90	1429.0
185.00	7972.0
186.00	62968.0
187.00	18408.0
188.00	2037.0
189.00	4330.0
190.00	743.0
191.00	3332.0
191.90	5807.0
193.00	8431.0
194.00	1822.0
195.00	881.0
196.00	12331.0
197.10	2702.0
197.90	478720.0
199.00	32552.0
200.00	2871.0
201.60	1676.0
203.00	3391.0
204.00	16688.0
205.00	29104.0
206.00	119240.0
207.00	28640.0
208.00	6935.0
209.00	3659.0
210.10	2074.0
211.00	4634.0
211.70	1043.0
213.00	336.0
214.90	1389.0
216.00	2909.0
216.90	32128.0
218.00	4281.0
218.90	474.0
221.00	22232.0
221.90	4587.0
223.00	7851.0
224.00	66392.0
225.00	18232.0
226.00	2062.0
227.00	25400.0
228.00	4094.0
229.00	6037.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390

230.00	912.0
231.10	2722.0
232.00	470.0
233.00	687.0
234.00	2031.0
234.90	2121.0
235.90	1361.0
237.00	2943.0
238.00	428.0
239.00	1135.0
240.00	910.0
240.90	1517.0
242.00	3717.0
243.00	3481.0
244.00	52200.0
245.00	7141.0
246.00	10530.0
247.00	2187.0
247.80	462.0
249.00	2355.0
250.00	544.0
251.00	858.0
252.00	491.0
253.00	1286.0
254.10	2343.0
255.00	287808.0
256.00	39432.0
257.00	3527.0
258.00	15082.0
259.00	2487.0
259.90	499.0
261.00	356.0
263.10	258.0
263.80	385.0
264.90	7233.0
266.00	1080.0
267.00	1535.0
267.90	494.0
269.00	490.0
269.90	383.0
271.00	538.0
272.00	670.0
273.00	9053.0
274.00	20944.0
275.00	125152.0
276.00	16448.0
277.00	9826.0
278.00	1940.0
278.90	473.0
281.00	5384.0
282.00	1701.0
283.10	2399.0
284.00	1053.0
285.00	1662.0
286.00	435.0
289.00	437.0
289.90	274.0
291.00	245.0
291.90	427.0
292.90	2080.0
293.90	571.0
295.10	1929.0
296.00	34336.0
296.90	4855.0
298.00	544.0
299.00	218.0
301.00	481.0
302.00	538.0
303.00	3830.0
304.00	1015.0
304.90	301.0
307.90	536.0
308.90	279.0
310.00	524.0
310.90	205.0
312.00	205.0
312.80	248.0
314.10	1585.0
314.90	3597.0
316.00	1954.0
316.90	304.0
321.00	1034.0
322.10	545.0
323.00	8905.0
323.90	1630.0
325.00	605.0
326.00	446.0
327.00	2638.0
327.90	1338.0
329.00	369.0
331.90	803.0
333.00	1094.0
334.00	7029.0
335.00	1662.0
336.00	321.0
339.10	301.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390

340.00	188.0
341.10	2399.0
342.00	759.0
343.00	515.0
344.00	252.0
345.00	153.0
345.90	2431.0
346.80	535.0
351.00	237.0
352.00	2891.0
353.00	2584.0
354.00	3064.0
355.10	2405.0
356.00	694.0
356.90	684.0
358.00	253.0
359.10	312.0
365.00	15161.0
365.90	2333.0
366.90	257.0
369.10	390.0
370.10	453.0
370.90	1106.0
372.10	6280.0
373.00	1287.0
376.90	250.0
383.00	1567.0
383.90	541.0
385.00	321.0
389.90	791.0
390.90	582.0
392.00	303.0
396.90	179.0
401.00	829.0
402.00	2578.0
403.00	3492.0
403.90	1303.0
405.00	308.0
415.00	639.0
416.00	328.0
417.20	259.0
421.00	2712.0
422.00	3238.0
423.00	20896.0
424.00	5999.0
425.10	719.0
429.00	521.0
430.10	240.0
431.00	241.0
441.00	61336.0
441.90	404288.0
443.10	70608.0
444.00	7693.0
445.00	354.0
460.90	198.0
475.10	302.0
489.10	260.0
490.90	196.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390

m/z	Abundance
35.00	267.0
37.00	745.0
38.10	1745.0
39.00	10189.0
40.00	1705.0
41.10	617.0
41.90	435.0
43.00	1231.0
44.00	3530.0
45.00	1189.0
46.00	186.0
46.90	262.0
49.00	1545.0
50.00	35288.0
51.00	142912.0
52.00	7720.0
52.90	571.0
55.00	1255.0
56.00	4438.0
57.00	10323.0
58.00	1003.0
59.00	941.0
60.10	527.0
61.00	2052.0
62.00	2074.0
63.00	6099.0
64.00	907.0
65.00	2823.0
66.00	236.0
66.90	426.0
68.00	2527.0
69.00	143616.0
70.00	729.0
70.90	258.0
72.10	215.0
73.00	10506.0
74.00	15727.0
75.00	24384.0
76.00	7852.0
77.00	167360.0
78.00	11584.0
79.00	11436.0
80.00	8334.0
81.00	12326.0
82.00	2866.0
83.00	3073.0
84.00	692.0
85.00	2206.0
86.00	3751.0
86.90	1527.0
88.00	706.0
89.00	393.0
91.00	3132.0
92.00	3279.0
93.00	20072.0
94.00	1433.0
95.00	393.0
96.00	1947.0
96.90	513.0
98.00	15265.0
99.00	11872.0
99.90	1053.0
101.00	7425.0
102.00	399.0
103.00	2789.0
104.00	4698.0
105.00	4269.0
105.90	1358.0
107.00	55224.0
108.00	8648.0
109.00	1544.0
110.00	102224.0
111.00	16370.0
112.00	1387.0
113.00	724.0
114.90	498.0
116.00	3210.0
117.00	44584.0
118.00	3523.0
118.90	1016.0
119.90	627.0
121.00	538.0
122.00	3774.0
123.00	5933.0
123.90	2778.0
125.00	2519.0
127.00	205120.0
128.00	15696.0
129.00	80048.0
130.00	6950.0
131.00	1561.0
131.90	865.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390

133.00	1645.0
133.90	2579.0
135.00	8827.0
135.90	2973.0
137.00	3955.0
137.90	720.0
138.90	421.0
139.90	890.0
141.00	10101.0
142.00	3400.0
143.00	2355.0
143.90	589.0
144.90	647.0
145.90	2059.0
147.00	8212.0
148.00	12214.0
149.00	3152.0
150.00	1020.0
151.00	1253.0
151.80	787.0
153.00	3186.0
154.00	2820.0
155.00	5309.0
156.00	9479.0
157.10	1593.0
157.90	2060.0
159.00	1611.0
160.00	3347.0
161.00	5044.0
162.00	1446.0
163.00	725.0
163.90	538.0
165.00	4116.0
165.90	3153.0
167.00	21448.0
168.00	10362.0
169.00	1754.0
170.00	769.0
170.90	854.0
171.90	1987.0
173.00	2421.0
174.00	4659.0
175.00	8495.0
176.00	2739.0
177.00	4260.0
178.00	1702.0
179.00	16204.0
180.00	11338.0
181.00	5279.0
182.00	1125.0
183.00	440.0
184.00	1350.0
185.00	7577.0
186.00	61552.0
187.00	18432.0
188.00	2039.0
189.00	4516.0
189.90	766.0
191.00	3260.0
192.00	5900.0
193.00	8820.0
194.00	1727.0
195.10	1034.0
196.00	12568.0
197.10	2530.0
198.00	469120.0
199.00	28768.0
200.00	2497.0
201.40	1819.0
203.00	3084.0
204.00	16061.0
205.00	26840.0
206.00	117760.0
207.00	28168.0
208.00	7190.0
209.00	3775.0
210.10	1932.0
211.00	4695.0
213.00	305.0
214.00	251.0
214.90	1432.0
216.00	2757.0
217.00	33152.0
218.00	4024.0
218.90	554.0
221.00	23016.0
221.90	4912.0
223.00	8163.0
224.00	66136.0
225.00	17144.0
226.10	1777.0
227.00	27344.0
228.00	3885.0
229.00	5917.0
229.90	848.0
231.00	2430.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390

231.90	318.0
233.00	622.0
234.00	1897.0
235.00	2124.0
236.00	1277.0
237.00	2662.0
237.90	507.0
239.00	1276.0
240.00	832.0
241.00	1423.0
242.00	3635.0
243.10	3507.0
244.00	51976.0
245.00	7805.0
246.00	10329.0
247.00	2102.0
248.00	541.0
248.90	2597.0
250.00	558.0
250.90	803.0
252.00	611.0
253.00	1524.0
255.00	292416.0
256.00	41088.0
257.00	3077.0
257.90	15739.0
259.00	2241.0
260.00	434.0
260.90	498.0
263.00	297.0
264.00	462.0
264.90	6468.0
265.90	830.0
267.00	1657.0
267.90	494.0
269.00	709.0
270.00	620.0
271.00	555.0
271.90	813.0
273.00	8378.0
274.00	20848.0
275.00	132416.0
276.00	17792.0
277.00	10307.0
278.00	1773.0
278.90	610.0
280.00	165.0
281.00	5499.0
282.10	1535.0
283.00	2578.0
284.10	1251.0
285.00	2262.0
286.00	448.0
289.00	370.0
289.90	372.0
291.00	365.0
291.90	590.0
293.00	2626.0
294.00	701.0
295.10	1940.0
296.00	35512.0
297.00	5028.0
298.10	476.0
299.00	194.0
301.00	583.0
302.00	620.0
303.00	3962.0
304.00	1132.0
304.90	174.0
307.90	557.0
308.90	398.0
309.90	533.0
310.90	215.0
311.90	192.0
313.00	380.0
314.00	1616.0
315.00	4241.0
316.00	2454.0
317.00	430.0
318.90	184.0
320.90	1106.0
322.00	842.0
323.00	11513.0
324.00	2368.0
324.90	784.0
325.80	369.0
327.00	2940.0
328.00	1263.0
328.90	398.0
332.00	1053.0
333.00	1303.0
334.00	8472.0
335.00	2155.0
335.90	232.0
339.00	410.0
340.00	264.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390

341.00	2641.0
342.00	771.0
343.10	556.0
346.00	2722.0
347.00	542.0
351.10	222.0
352.00	3570.0
353.00	2808.0
354.00	4330.0
355.00	2993.0
356.10	912.0
357.10	708.0
358.00	266.0
358.90	422.0
364.90	18456.0
365.90	2769.0
366.90	312.0
369.20	241.0
370.00	640.0
370.90	1034.0
372.00	7152.0
373.00	1387.0
373.90	192.0
376.90	155.0
382.90	2015.0
384.00	523.0
385.10	332.0
390.00	960.0
391.00	570.0
391.90	434.0
401.00	986.0
402.00	2864.0
403.00	4954.0
404.00	1713.0
405.00	328.0
415.00	731.0
416.00	417.0
417.00	282.0
421.00	3536.0
422.00	3394.0
423.00	27448.0
424.00	6358.0
425.00	785.0
429.00	532.0
430.10	262.0
431.20	189.0
441.00	78624.0
442.00	483520.0
443.00	106184.0
444.00	9522.0
444.90	533.0
460.90	231.0
475.10	343.0
476.20	242.0
488.90	205.0
490.00	190.0
535.10	153.0

Instrument :
BNA_N
ClientSampleId :
DFTPP390