

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112322\
 Data File : BN022856.D
 Acq On : 23 Nov 2022 21:37
 Operator : CG/JU
 Sample : PB149182BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB149182BL

Quant Time: Nov 24 04:08:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 24 03:45:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.078	152	3783	0.400	ng	0.00
7) Naphthalene-d8	10.893	136	9669	0.400	ng	0.00
13) Acenaphthene-d10	14.709	164	5310	0.400	ng	0.00
19) Phenanthrene-d10	17.452	188	11524	0.400	ng	0.00
29) Chrysene-d12	21.638	240	7092	0.400	ng	0.00
35) Perylene-d12	24.133	264	5634	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.608	112	4159	0.368	ng	0.00
5) Phenol-d6	7.219	99	5204	0.357	ng	0.00
8) Nitrobenzene-d5	9.238	82	2644	0.352	ng	0.00
11) 2-Methylnaphthalene-d10	12.480	152	6923	0.319	ng	0.00
14) 2,4,6-Tribromophenol	16.198	330	951	0.305	ng	0.00
15) 2-Fluorobiphenyl	13.351	172	9357	0.400	ng	0.00
27) Fluoranthene-d10	19.481	212	11027	0.325	ng	0.00
31) Terphenyl-d14	20.080	244	7590	0.472	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112322\
Data File : BN022856.D
Acq On : 23 Nov 2022 21:37
Operator : CG/JU
Sample : PB149182BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB149182BL

Quant Time: Nov 24 04:08:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 24 03:45:05 2022
Response via : Initial Calibration

