

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112322\
 Data File : BN022851.D
 Acq On : 23 Nov 2022 17:53
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Nov 24 03:19:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 24 03:17:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.078	152	3605	0.400	ng	0.00
7) Naphthalene-d8	10.893	136	9990	0.400	ng	0.00
13) Acenaphthene-d10	14.709	164	6413	0.400	ng	0.00
19) Phenanthrene-d10	17.452	188	14281	0.400	ng	0.00
29) Chrysene-d12	21.643	240	12423	0.400	ng	0.00
35) Perylene-d12	24.135	264	11993	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.608	112	8520	0.767	ng	0.00
5) Phenol-d6	7.219	99	10898	0.790	ng	0.00
8) Nitrobenzene-d5	9.238	82	5825	0.571	ng	0.00
11) 2-Methylnaphthalene-d10	12.480	152	18739	1.130	ng	0.00
14) 2,4,6-Tribromophenol	16.198	330	2800	0.786	ng	0.00
15) 2-Fluorobiphenyl	13.351	172	21634	0.840	ng	0.00
27) Fluoranthene-d10	19.486	212	32533	0.784	ng	0.00
31) Terphenyl-d14	20.085	244	21845	0.577	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.449	88	3774	0.745	ng	98
3) n-Nitrosodimethylamine	3.774	42	4331	0.745	ng	# 95
6) bis(2-Chloroethyl)ether	7.493	93	10213	0.862	ng	98
9) Naphthalene	10.947	128	24360	0.592	ng	98
10) Hexachlorobutadiene	11.246	225	4993	0.925	ng	# 100
12) 2-Methylnaphthalene	12.163	142	5868	0.839	ng	# 87
16) Acenaphthylene	14.431	152	28025	0.921	ng	100
17) Acenaphthene	14.773	154	17352	0.850	ng	99
18) Fluorene	15.751	166	22356	0.739	ng	99
20) 4,6-Dinitro-2-methylph...	15.826	198	941	0.336	ng	# 53
21) 4-Bromophenyl-phenylether	16.645	248	8172	0.962	ng	98
22) Hexachlorobenzene	16.757	284	9537	1.012	ng	100
23) Atrazine	16.918	200	6328	0.780	ng	97
24) Pentachlorophenol	17.104	266	2809	0.650	ng	99
25) Phenanthrene	17.501	178	37512	0.757	ng	100
26) Anthracene	17.588	178	35455	0.861	ng	100
28) Fluoranthene	19.515	202	42885	0.823	ng	100
30) Pyrene	19.878	202	43803	0.842	ng	100
32) Benzo(a)anthracene	21.625	228	39861	0.837	ng	99
33) Chrysene	21.679	228	38970	0.826	ng	100
34) Bis(2-ethylhexyl)phtha...	21.563	149	22337	0.642	ng	99
36) Indeno(1,2,3-cd)pyrene	26.752	276	42911	0.839	ng	100
37) Benzo(b)fluoranthene	23.372	252	38358	0.805	ng	97
38) Benzo(k)fluoranthene	23.422	252	38926	0.804	ng	98
39) Benzo(a)pyrene	24.024	252	31336	0.803	ng	96
40) Dibenzo(a,h)anthracene	26.778	278	33742	0.824	ng	98
41) Benzo(g,h,i)perylene	27.553	276	35666	0.841	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112322\
Data File : BN022851.D
Acq On : 23 Nov 2022 17:53
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Nov 24 03:19:03 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 24 03:17:21 2022
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 8.078 min Scan# 692
 Delta R.T. 0.000 min
 Lab File: BN022851.D
 Acq: 23 Nov 2022 17:53




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:152 Resp: 3605
 Ion Ratio Lower Upper
 152 100
 150 152.2 120.4 180.6
 115 65.5 53.8 80.8



#2
 1,4-Dioxane
 Concen: 0.745 ng
 RT: 3.449 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: BN022851.D
 Acq: 23 Nov 2022 17:53

Tgt Ion: 88 Resp: 3774
 Ion Ratio Lower Upper
 88 100
 43 32.6 27.8 41.8
 58 76.8 61.1 91.7









#3
n-Nitrosodimethylamine
Concen: 0.745 ng
RT: 3.774 min Scan# 90
Delta R.T. -0.007 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 4331
Ion Ratio Lower Upper
42 100
74 116.4 89.9 134.9
44 7.2 10.5 15.7#



#4
2-Fluorophenol
Concen: 0.767 ng
RT: 5.608 min Scan# 350
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion: 112 Resp: 8520
Ion Ratio Lower Upper
112 100
64 60.3 47.9 71.9
63 32.5 25.1 37.7









#5
Phenol-d6
Concen: 0.790 ng
RT: 7.219 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 99 Resp: 10898
Ion Ratio Lower Upper
99 100
42 22.1 19.0 28.6
71 34.6 27.4 41.0



#6
bis(2-Chloroethyl)ether
Concen: 0.862 ng
RT: 7.493 min Scan# 611
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion: 93 Resp: 10213
Ion Ratio Lower Upper
93 100
63 76.0 59.0 88.6
95 31.0 24.3 36.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.893 min Scan# 966
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	9990
Ion	Ratio	Lower	Upper
136	100		
137	11.9	9.8	14.6
54	12.8	11.1	16.7
68	9.9	9.0	13.6



#8
Nitrobenzene-d5
Concen: 0.571 ng
RT: 9.238 min Scan# 811
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion:	82	Resp:	5825
Ion	Ratio	Lower	Upper
82	100		
128	43.4	35.8	53.8
54	55.2	46.6	70.0









#9
Naphthalene
Concen: 0.592 ng
RT: 10.947 min Scan# 971
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	128	Resp:	24360
Ion Ratio	Lower	Upper	
128	100		
129	11.4	10.1	15.1
127	13.5	11.4	17.0



#10
Hexachlorobutadiene
Concen: 0.925 ng
RT: 11.246 min Scan# 999
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion:	225	Resp:	4993
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.6	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 1.130 ng
RT: 12.480 min Scan# 1196
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 18739
Ion Ratio Lower Upper
152 100
151 20.5 18.6 27.8



#12
2-Methylnaphthalene
Concen: 0.839 ng
RT: 12.163 min Scan# 1196
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion:142 Resp: 5868
Ion Ratio Lower Upper
142 100
141 13.0 13.7 20.5#
115 6.7 10.6 16.0#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.709 min Scan# 1493
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 6413
Ion Ratio Lower Upper
164 100
162 106.2 84.2 126.4
160 51.2 41.6 62.4



#14
2,4,6-Tribromophenol
Concen: 0.786 ng
RT: 16.198 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion:330 Resp: 2800
Ion Ratio Lower Upper
330 100
332 97.9 77.4 116.2
141 36.4 29.3 43.9









#15
2-Fluorobiphenyl
Concen: 0.840 ng
RT: 13.351 min Scan# 1366
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 21634
Ion Ratio Lower Upper
172 100
171 34.2 27.8 41.8
170 22.6 18.9 28.3



#16
Acenaphthylene
Concen: 0.921 ng
RT: 14.431 min Scan# 1467
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion:152 Resp: 28025
Ion Ratio Lower Upper
152 100
151 19.8 15.9 23.9
153 13.1 10.4 15.6









#17
Acenaphthene
Concen: 0.850 ng
RT: 14.773 min Scan# 1499
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:154 Resp: 17352
Ion Ratio Lower Upper
154 100
153 111.5 89.6 134.4
152 54.7 44.3 66.5



#18
Fluorene
Concen: 0.739 ng
RT: 15.751 min Scan# 1590
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion:166 Resp: 22356
Ion Ratio Lower Upper
166 100
165 99.1 79.9 119.9
167 13.7 11.3 16.9





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.452 min Scan# 1727

Delta R.T. 0.000 min

Lab File: BN022851.D

Acq: 23 Nov 2022 17:53





Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8

Tgt Ion:188	Resp:	14281
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	15.9	13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.336 ng




RT: 15.826 min Scan# 1596


Delta R.T. 0.000 min

Lab File: BN022851.D

Acq: 23 Nov 2022 17:53

Tgt Ion:198	Resp:	941
Ion Ratio	Lower	Upper
198	100	
51	75.0	104.4
105	71.4	100.2









#21
4-Bromophenyl-phenylether
Concen: 0.962 ng
RT: 16.645 min Scan# 1662
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 8172
Ion Ratio Lower Upper
248 100
250 91.6 72.8 109.2
141 78.6 65.6 98.4



#22
Hexachlorobenzene
Concen: 1.012 ng
RT: 16.757 min Scan# 1671
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion:284 Resp: 9537
Ion Ratio Lower Upper
284 100
142 36.2 29.0 43.4
249 30.1 24.3 36.5









#23
 Atrazine
 Concen: 0.780 ng
 RT: 16.918 min Scan# 1684
 Delta R.T. 0.000 min
 Lab File: BN022851.D
 Acq: 23 Nov 2022 17:53




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:200 Resp: 6328
 Ion Ratio Lower Upper
 200 100
 173 22.2 20.1 30.1
 215 47.6 39.1 58.7



#24
 Pentachlorophenol
 Concen: 0.650 ng
 RT: 17.104 min Scan# 1699
 Delta R.T. 0.000 min
 Lab File: BN022851.D
 Acq: 23 Nov 2022 17:53

Tgt Ion:266 Resp: 2809
 Ion Ratio Lower Upper
 266 100
 264 62.1 48.6 73.0
 268 63.8 50.3 75.5





#25

Phenanthrene

Concen: 0.757 ng

RT: 17.501 min Scan# 1731

Delta R.T. 0.000 min

Lab File: BN022851.D

Acq: 23 Nov 2022 17:53

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:178 Resp: 37512

Ion Ratio Lower Upper

178 100

176 19.8 15.7 23.5

179 15.2 12.2 18.2



#26

Anthracene

Concen: 0.861 ng

RT: 17.588 min Scan# 1738

Delta R.T. 0.000 min

Lab File: BN022851.D

Acq: 23 Nov 2022 17:53




Tgt Ion:178 Resp: 35455


Ion Ratio Lower Upper

178 100

176 19.0 15.1 22.7

179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.784 ng
RT: 19.486 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 32533
Ion Ratio Lower Upper
212 100
106 16.2 13.0 19.4
104 9.1 7.4 11.0



#28
Fluoranthene
Concen: 0.823 ng
RT: 19.515 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion:202 Resp: 42885
Ion Ratio Lower Upper
202 100
101 12.5 10.0 15.0
203 17.3 13.8 20.8





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.643 min Scan# 2428

Delta R.T. 0.000 min

Lab File: BN022851.D

Acq: 23 Nov 2022 17:53

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240 Resp: 12423

Ion Ratio Lower Upper

240 100

120 14.0 11.6 17.4

236 28.3 22.8 34.2



#30

Pyrene

Concen: 0.842 ng

RT: 19.878 min Scan# 2215

Delta R.T. 0.000 min

Lab File: BN022851.D

Acq: 23 Nov 2022 17:53




Tgt Ion:202 Resp: 43803


Ion Ratio Lower Upper

202 100

200 21.3 17.1 25.7

203 17.8 14.3 21.5





#31

Terphenyl-d14

Concen: 0.577 ng

RT: 20.085 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN022851.D

Acq: 23 Nov 2022 17:53

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:244 Resp: 21845

Ion Ratio Lower Upper

244 100

212 7.0 6.2 9.4

122 9.1 8.6 12.8



#32

Benzo(a)anthracene

Concen: 0.837 ng

RT: 21.625 min Scan# 2426

Delta R.T. 0.000 min

Lab File: BN022851.D

Acq: 23 Nov 2022 17:53




Tgt Ion:228 Resp: 39861


Ion Ratio Lower Upper

228 100

226 27.7 22.6 33.8

229 19.3 15.8 23.6









#33
Chrysene
Concen: 0.826 ng
RT: 21.679 min Scan# 2432
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:228 Resp: 38970
Ion Ratio Lower Upper
228 100
226 31.3 25.4 38.0
229 19.6 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.642 ng
RT: 21.563 min Scan# 2419
Delta R.T. -0.009 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion:149 Resp: 22337
Ion Ratio Lower Upper
149 100
167 26.6 21.6 32.4
279 2.9 2.5 3.7









#35
Perylene-d12
Concen: 0.400 ng
RT: 24.135 min Scan# 3105
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:264	Resp:	11993
Ion Ratio	Lower	Upper
264	100	
260	26.9	21.2 31.8
265	28.4	22.8 34.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.839 ng
RT: 26.752 min Scan# 4000
Delta R.T. -0.003 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion:276	Resp:	42911
Ion Ratio	Lower	Upper
276	100	
138	28.4	22.7 34.1
277	24.8	19.8 29.6









#37
Benzo(b)fluoranthene
Concen: 0.805 ng
RT: 23.372 min Scan# 2844
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 38358
Ion Ratio Lower Upper
252 100
253 22.1 18.7 28.1
125 12.2 10.7 16.1



#38
Benzo(k)fluoranthene
Concen: 0.804 ng
RT: 23.422 min Scan# 2861
Delta R.T. -0.003 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53


Tgt Ion:252 Resp: 38926
Ion Ratio Lower Upper
252 100
253 22.2 18.3 27.5
125 12.7 11.4 17.0








#39
Benzo(a)pyrene
Concen: 0.803 ng
RT: 24.024 min Scan# 3067
Delta R.T. 0.000 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8







Tgt Ion:252 Resp: 31336
Ion Ratio Lower Upper
252 100
253 22.1 19.1 28.7
125 13.9 12.2 18.4



#40
Dibenzo(a,h)anthracene
Concen: 0.824 ng
RT: 26.778 min Scan# 4009
Delta R.T. -0.003 min
Lab File: BN022851.D
Acq: 23 Nov 2022 17:53

Tgt Ion:278 Resp: 33742
Ion Ratio Lower Upper
278 100
139 20.2 17.2 25.8
279 23.9 19.8 29.8





#41

Benzo(g,h,i)perylene

Concen: 0.841 ng

RT: 27.553 min Scan# 41

Delta R.T. 0.003 min

Lab File: BN022851.D


Acq: 23 Nov 2022 17:53

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8



Tgt Ion:276 Resp: 35666

Ion Ratio Lower Upper

276 100

277 24.0 19.6 29.4

138 25.8 22.0 33.0

