

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112521\
 Data File : BN017593.D
 Acq On : 24 Nov 2021 23:16
 Operator : CG/JU
 Sample : PB141021BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB141021BL

Quant Time: Nov 25 03:28:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 25 03:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.754	152	24460	0.400	ng	0.00
7) Naphthalene-d8	10.534	136	104499	0.400	ng	0.00
13) Acenaphthene-d10	14.371	164	55288	0.400	ng	0.00
19) Phenanthrene-d10	17.116	188	108866	0.400	ng	0.00
29) Chrysene-d12	21.313	240	84654	0.400	ng	0.00
35) Perylene-d12	23.622	264	61668	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.375	112	20561	0.319	ng	0.03
5) Phenol-d6	6.933	99	21728	0.298	ng	0.00
8) Nitrobenzene-d5	8.887	82	13279	0.241	ng	0.00
11) 2-Methylnaphthalene-d10	12.124	152	54833	0.303	ng	0.00
14) 2,4,6-Tribromophenol	15.868	330	5175	0.181	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	67403	0.316	ng	0.00
27) Fluoranthene-d10	19.153	212	105845	0.289	ng	0.00
31) Terphenyl-d14	19.759	244	70932	0.357	ng	0.00
Target Compounds						Qvalue
6) bis(2-Chloroethyl)ether	7.201	93	40	0.001	ng	# 68
9) Naphthalene	10.573	128	206	0.001	ng	# 16
10) Hexachlorobutadiene	10.902	225	9	0.000	ng	# 68
12) 2-Methylnaphthalene	12.195	142	99	0.001	ng	# 51
16) Acenaphthylene	14.092	152	97	0.000	ng	# 72
17) Acenaphthene	14.435	154	36	0.000	ng	# 65
18) Fluorene	15.424	166	57	0.000	ng	# 91
20) 4,6-Dinitro-2-methylph...	15.501	198	9	Below Cal		# 7
21) 4-Bromophenyl-phenylether	16.325	248	14	0.000	ng	# 1
22) Hexachlorobenzene	16.434	284	12	0.000	ng	# 1
23) Atrazine	16.593	200	16	0.100	ng	# 3
24) Pentachlorophenol	16.751	266	9	0.089	ng	# 54
25) Phenanthrene	17.165	178	164	0.001	ng	# 83
26) Anthracene	17.250	178	54	0.000	ng	# 21
28) Fluoranthene	19.183	202	126	0.000	ng	# 65
30) Pyrene	19.541	202	122	0.000	ng	# 81
32) Benzo(a)anthracene	21.303	228	574	0.002	ng	# 52
33) Chrysene	21.303	228	417	0.002	ng	# 56
34) Bis(2-ethylhexyl)phtha...	21.239	149	890	0.011	ng	94
36) Indeno(1,2,3-cd)pyrene	25.981	276	66	0.001	ng	# 68
37) Benzo(b)fluoranthene	22.955	252	866	0.004	ng	# 1
38) Benzo(k)fluoranthene	22.955	252	603	0.003	ng	# 1
39) Benzo(a)pyrene	23.520	252	87	0.000	ng	# 19
40) Dibenzo(a,h)anthracene	26.008	278	126	0.001	ng	# 1
41) Benzo(g,h,i)perylene	26.703	276	23	0.000	ng	# 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112521\
 Data File : BN017593.D
 Acq On : 24 Nov 2021 23:16
 Operator : CG/JU
 Sample : PB141021BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB141021BL

Quant Time: Nov 25 03:28:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 25 03:24:52 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.754 min Scan# 518
Delta R.T. 0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16




Instrument :
BNA_N
ClientSampleId :
PB141021BL


Tgt Ion:152 Resp: 24460
Ion Ratio Lower Upper
152 100
150 151.8 122.5 183.7
115 57.8 47.0 70.4



#4
2-Fluorophenol
Concen: 0.319 ng
RT: 5.375 min Scan# 260
Delta R.T. 0.027 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Tgt Ion:112 Resp: 20561
Ion Ratio Lower Upper
112 100
64 64.3 51.5 77.3
63 32.8 26.6 39.8









#5
Phenol-d6
Concen: 0.298 ng
RT: 6.933 min Scan# 41
Delta R.T. 0.009 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16




Instrument :
BNA_N
ClientSampleId :
PB141021BL


Tgt Ion: 99 Resp: 21728
Ion Ratio Lower Upper
99 100
42 19.1 16.2 24.2
71 34.7 27.6 41.4



#6
bis(2-Chloroethyl)ether
Concen: 0.001 ng
RT: 7.201 min Scan# 458
Delta R.T. 0.018 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Tgt Ion: 93 Resp: 40
Ion Ratio Lower Upper
93 100
63 60.0 63.0 94.6#
95 0.0 25.6 38.4#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.534 min Scan# 754
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16




Instrument :
BNA_N
ClientSampleId :
PB141021BL


Tgt Ion:136 Resp: 104499
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 6.4 5.2 7.8
68 5.7 4.6 6.8



#8
Nitrobenzene-d5
Concen: 0.241 ng
RT: 8.887 min Scan# 624
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Tgt Ion: 82 Resp: 13279
Ion Ratio Lower Upper
82 100
128 32.3 25.3 37.9
54 52.5 43.7 65.5









#9
Naphthalene

Concen: 0.001 ng
RT: 10.573 min Scan# 71
Delta R.T. -0.013 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Instrument :
BNA_N
ClientSampleId :
PB141021BL



Tgt Ion:128 Resp: 206
Ion Ratio Lower Upper
128 100
129 47.9 8.8 13.2#
127 41.8 10.2 15.4#




#10

Hexachlorobutadiene
Concen: 0.000 ng
RT: 10.902 min Scan# 783
Delta R.T. 0.025 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Tgt Ion:225 Resp: 9
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 88.9 51.1 76.7#





#11

2-Methylnaphthalene-d10

Concen: 0.303 ng

RT: 12.124 min Scan# 9

Delta R.T. -0.000 min

Lab File: BN017593.D

Acq: 24 Nov 2021 23:16

Instrument :

BNA_N

ClientSampleId :





PB141021BL

Tgt Ion:152 Resp: 54833

Ion Ratio Lower Upper

152 100

151 20.2 16.2 24.4



#12

2-Methylnaphthalene

Concen: 0.001 ng

RT: 12.195 min Scan# 985

Delta R.T. -0.005 min

Lab File: BN017593.D

Acq: 24 Nov 2021 23:16




Tgt Ion:142 Resp: 99


Ion Ratio Lower Upper

142 100

141 96.3 69.4 104.0

115 118.5 27.8 41.6#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.371 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16




Instrument :
BNA_N
ClientSampleId :
PB141021BL


Tgt Ion:164 Resp: 55288
Ion Ratio Lower Upper
164 100
162 102.5 82.5 123.7
160 48.6 39.6 59.4



#14
2,4,6-Tribromophenol
Concen: 0.181 ng
RT: 15.868 min Scan# 1320
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Tgt Ion:330 Resp: 5175
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 30.5 24.1 36.1









#15
2-Fluorobiphenyl
Concen: 0.316 ng
RT: 13.001 min Scan# 1094
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16




Instrument :
BNA_N
ClientSampleId :
PB141021BL


Tgt Ion:172 Resp: 67403
Ion Ratio Lower Upper
172 100
171 34.2 27.6 41.4
170 22.8 18.3 27.5



#16
Acenaphthylene
Concen: 0.000 ng
RT: 14.092 min Scan# 1180
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Tgt Ion:152 Resp: 97
Ion Ratio Lower Upper
152 100
151 27.8 15.5 23.3#
153 29.9 10.5 15.7#









#17
Acenaphthene
Concen: 0.000 ng
RT: 14.435 min Scan# 1207
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16




Instrument :
BNA_N
ClientSampleId :
PB141021BL


Tgt Ion:154 Resp: 36
Ion Ratio Lower Upper
154 100
153 158.3 89.8 134.8#
152 72.2 45.9 68.9#



#18
Fluorene
Concen: 0.000 ng
RT: 15.424 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Tgt Ion:166 Resp: 57
Ion Ratio Lower Upper
166 100
165 87.7 77.6 116.4
167 17.5 10.9 16.3#





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.116 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BN017593.D

Acq: 24 Nov 2021 23:16

Instrument :

BNA_N

ClientSampleId :

PB141021BL





Tgt Ion:188 Resp: 108866

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.2 9.0 13.4



#20

4,6-Dinitro-2-methylphenol

Concen: Below Cal

RT: 15.501 min Scan# 1291

Delta R.T. -0.000 min

Lab File: BN017593.D

Acq: 24 Nov 2021 23:16




Tgt Ion:198 Resp: 9


Ion Ratio Lower Upper

198 100

182 100.0 33.1 49.7#

77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.000 ng
RT: 16.325 min Scan# 1357
Delta R.T. 0.012 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16




Instrument :
BNA_N
ClientSampleId :
PB141021BL


Tgt Ion:248 Resp: 14
Ion Ratio Lower Upper
248 100
250 275.6 76.6 114.8#
141 93.6 50.5 75.7#



#22
Hexachlorobenzene
Concen: 0.000 ng
RT: 16.434 min Scan# 1366
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Tgt Ion:284 Resp: 12
Ion Ratio Lower Upper
284 100
142 0.0 23.2 34.8#
249 141.7 24.2 36.2#









#23
 Atrazine
 Concen: 0.100 ng
 RT: 16.593 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BN017593.D
 Acq: 24 Nov 2021 23:16




Instrument :
 BNA_N
 ClientSampleId :
 PB141021BL


Tgt Ion:200 Resp: 16
 Ion Ratio Lower Upper
 200 100
 173 103.7 17.4 26.0#
 215 96.3 40.9 61.3#



#24
 Pentachlorophenol
 Concen: 0.089 ng
 RT: 16.751 min Scan# 1392
 Delta R.T. -0.024 min
 Lab File: BN017593.D
 Acq: 24 Nov 2021 23:16

Tgt Ion:266 Resp: 9
 Ion Ratio Lower Upper
 266 100
 264 0.0 50.2 75.2#
 268 55.6 51.7 77.5









#25
Phenanthrene
Concen: 0.001 ng
RT: 17.165 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16




Instrument :
BNA_N
ClientSampleId :
PB141021BL


Tgt Ion:178 Resp: 164
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.6
179 0.0 12.2 18.4#



#26
Anthracene
Concen: 0.000 ng
RT: 17.250 min Scan# 1433
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Tgt Ion:178 Resp: 54
Ion Ratio Lower Upper
178 100
176 63.0 14.6 21.8#
179 37.0 12.2 18.4#









#27
Fluoranthene-d10
Concen: 0.289 ng
RT: 19.153 min Scan# 1725
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16




Instrument :
BNA_N
ClientSampleId :
PB141021BL


Tgt Ion:212 Resp: 105845
Ion Ratio Lower Upper
212 100
106 12.1 9.6 14.4
104 6.6 5.3 7.9



#28
Fluoranthene
Concen: 0.000 ng
RT: 19.183 min Scan# 1731
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Tgt Ion:202 Resp: 126
Ion Ratio Lower Upper
202 100
101 0.0 8.3 12.5#
203 0.0 13.9 20.9#









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.313 min Scan# 20
Delta R.T. -0.000 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16




Instrument :
BNA_N
ClientSampleId :
PB141021BL


Tgt Ion:240 Resp: 84654
Ion Ratio Lower Upper
240 100
120 6.2 6.1 9.1
236 26.2 21.2 31.8



#30
Pyrene
Concen: 0.000 ng
RT: 19.541 min Scan# 1803
Delta R.T. -0.010 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16


Tgt Ion:202 Resp: 122
Ion Ratio Lower Upper
202 100
200 27.0 16.4 24.6#
203 28.7 13.9 20.9#








#31
 Terphenyl-d14
 Concen: 0.357 ng
 RT: 19.759 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BN017593.D
 Acq: 24 Nov 2021 23:16

Instrument :
 BNA_N
 ClientSampleId :
 PB141021BL







Tgt Ion:244 Resp: 70932
 Ion Ratio Lower Upper
 244 100
 212 7.5 6.2 9.2
 122 10.3 8.9 13.3



#32
 Benzo(a)anthracene
 Concen: 0.002 ng
 RT: 21.303 min Scan# 2019
 Delta R.T. 0.010 min
 Lab File: BN017593.D
 Acq: 24 Nov 2021 23:16

Tgt Ion:228 Resp: 574
 Ion Ratio Lower Upper
 228 100
 226 42.6 22.4 33.6#
 229 53.9 15.4 23.2#





#33

Chrysene

Concen: 0.002 ng

RT: 21.303 min Scan# 2019

Delta R.T. -0.043 min

Lab File: BN017593.D

Acq: 24 Nov 2021 23:16

Instrument :

BNA_N

ClientSampleId :

PB141021BL





Tgt Ion:228 Resp: 417

Ion Ratio Lower Upper

228 100

226 42.6 24.0 36.0#

229 53.9 15.6 23.4#



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.011 ng

RT: 21.239 min Scan# 2013

Delta R.T. 0.000 min

Lab File: BN017593.D

Acq: 24 Nov 2021 23:16




Tgt Ion:149 Resp: 890


Ion Ratio Lower Upper

149 100

167 25.4 23.2 34.8

279 4.5 3.4 5.0





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.622 min Scan# 2479

Delta R.T. -0.004 min

Lab File: BN017593.D


Acq: 24 Nov 2021 23:16

Instrument :

BNA_N

ClientSampleId :

PB141021BL






Tgt Ion:264 Resp: 61668

Ion Ratio Lower Upper

264 100

260 24.8 20.1 30.1

265 25.2 20.6 30.8



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.001 ng

RT: 25.981 min Scan# 3167

Delta R.T. -0.017 min

Lab File: BN017593.D

Acq: 24 Nov 2021 23:16




Tgt Ion:276 Resp: 66


Ion Ratio Lower Upper

276 100

138 0.0 16.4 24.6#


277 12.1 17.9 26.9#








#37
Benzo(b)fluoranthene
Concen: 0.004 ng
RT: 22.955 min Scan# 21
Delta R.T. 0.027 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16


Instrument :
BNA_N
ClientSampleId :
PB141021BL





Tgt Ion:252 Resp: 866
Ion Ratio Lower Upper
252 100
253 87.3 17.5 26.3#
125 317.3 7.1 10.7#




#38
Benzo(k)fluoranthene
Concen: 0.003 ng
RT: 22.955 min Scan# 2283
Delta R.T. -0.017 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16



Tgt Ion:252 Resp: 603
Ion Ratio Lower Upper
252 100
253 87.3 17.7 26.5#
125 317.3 7.9 11.9#









#39
Benzo(a)pyrene
Concen: 0.000 ng
RT: 23.520 min Scan# 2449
Delta R.T. -0.004 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16




Instrument :
BNA_N
ClientSampleId :
PB141021BL


Tgt Ion:252 Resp: 87
Ion Ratio Lower Upper
252 100
253 51.9 17.7 26.5#
125 55.4 8.0 12.0#



#40
Dibenzo(a,h)anthracene
Concen: 0.001 ng
RT: 26.008 min Scan# 3175
Delta R.T. -0.010 min
Lab File: BN017593.D
Acq: 24 Nov 2021 23:16

Tgt Ion:278 Resp: 126
Ion Ratio Lower Upper
278 100
139 522.3 15.2 22.8#
279 133.8 19.5 29.3#





#41

Benzo(g,h,i)perylene

Concen: 0.000 ng

RT: 26.703 min Scan# 31

Delta R.T. -0.021 min

Lab File: BN017593.D


Acq: 24 Nov 2021 23:16

Instrument :

BNA_N

ClientSampleId :

PB141021BL



Tgt Ion:276 Resp: 23

Ion Ratio Lower Upper

276 100

277 93.3 20.0 30.0#

138 106.7 15.7 23.5#

