

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112521\
 Data File : BN017591.D
 Acq On : 24 Nov 2021 21:28
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN112521

Quant Time: Nov 25 03:26:31 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 25 03:24:52 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.195	0.206	-5.6	115	0.00
3	n-Nitrosodimethylamine	0.397	0.401	-1.0	108	0.00
4 S	2-Fluorophenol	1.054	1.081	-2.6	107	0.00
5 S	Phenol-d6	1.190	1.223	-2.8	107	0.00
6	bis(2-Chloroethyl)ether	1.171	1.221	-4.3	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.211	0.192	9.0	114	0.00
9	Naphthalene	1.006	1.022	-1.6	108	-0.01
10	Hexachlorobutadiene	0.253	0.256	-1.2	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.693	0.700	-1.0	106	0.00
12	2-Methylnaphthalene	0.666	0.686	-3.0	107	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.206	0.188	8.7	107	0.00
15 S	2-Fluorobiphenyl	1.544	1.581	-2.4	107	0.00
16	Acenaphthylene	1.690	1.633	3.4	105	0.00
17	Acenaphthene	1.085	1.091	-0.6	106	0.00
18	Fluorene	1.369	1.377	-0.6	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4,6-Dinitro-2-methylphenol	0.004	0.002	50.0#	106	0.00
21	4-Bromophenyl-phenylether	0.250	0.251	-0.4	106	0.00
22	Hexachlorobenzene	0.288	0.298	-3.5	106	0.00
23	Atrazine	0.115	0.130	-13.0	125	-0.01
24	Pentachlorophenol	0.087	0.076	12.6	98	0.00
25	Phenanthrene	1.087	1.124	-3.4	105	-0.01
26	Anthracene	0.997	0.994	0.3	104	0.00
27 SURR	Fluoranthene-d10	1.347	1.337	0.7	103	0.00
28	Fluoranthene	1.313	1.361	-3.7	105	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
30	Pyrene	1.436	1.490	-3.8	104	0.00
31 S	Terphenyl-d14	0.938	0.961	-2.5	104	0.00
32	Benzo(a)anthracene	1.330	1.329	0.1	102	0.00
33	Chrysene	1.309	1.352	-3.3	104	0.00
34	Bis(2-ethylhexyl)phthalate	0.366	0.368	-0.5	122	0.00
35 I	Perylene-d12	1.000	1.000	0.0	102	0.00
36	Indeno(1,2,3-cd)pyrene	0.854	0.842	1.4	101	0.00
37	Benzo(b)fluoranthene	1.546	1.564	-1.2	102	0.00
38	Benzo(k)fluoranthene	1.464	1.500	-2.5	102	0.00
39 C	Benzo(a)pyrene	1.233	1.205	2.3	100	0.00
40	Dibenzo(a,h)anthracene	0.718	0.701	2.4	101	0.00
41	Benzo(g,h,i)perylene	0.637	0.629	1.3	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN112521