

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112521\
 Data File : BN017590.D
 Acq On : 24 Nov 2021 20:52
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Nov 25 02:38:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 25 02:34:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.754	152	25338	0.400	ng	0.00
7) Naphthalene-d8	10.535	136	98425	0.400	ng	0.00
13) Acenaphthene-d10	14.372	164	55539	0.400	ng	0.00
19) Phenanthrene-d10	17.116	188	114028	0.400	ng	0.00
29) Chrysene-d12	21.314	240	107517	0.400	ng	0.00
35) Perylene-d12	23.626	264	79993	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.339	112	302721	4.820	ng	0.00
5) Phenol-d6	6.925	99	345599	4.878	ng	0.00
8) Nitrobenzene-d5	8.887	82	328797	5.073	ng	0.00
11) 2-Methylnaphthalene-d10	12.124	152	806996	4.922	ng	0.00
14) 2,4,6-Tribromophenol	15.869	330	165397	9.695	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	1020293	4.933	ng	0.00
27) Fluoranthene-d10	19.154	212	1851164	5.278	ng	0.00
31) Terphenyl-d14	19.759	244	1225566	5.188	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.301	88	59181	4.870	ng	# 100
3) n-Nitrosodimethylamine	3.614	42	124038	4.962	ng	# 58
6) bis(2-Chloroethyl)ether	7.183	93	327621	4.530	ng	100
9) Naphthalene	10.573	128	1179536	4.723	ng	99
10) Hexachlorobutadiene	10.877	225	294479	5.980	ng	# 100
12) 2-Methylnaphthalene	12.195	142	759593	4.727	ng	100
16) Acenaphthylene	14.092	152	1188339	5.363	ng	100
17) Acenaphthene	14.435	154	734068	4.796	ng	97
18) Fluorene	15.425	166	908658	4.927	ng	100
21) 4-Bromophenyl-phenylether	16.313	248	358731	5.729	ng	95
22) Hexachlorobenzene	16.435	284	383738	5.669	ng	100
25) Phenanthrene	17.153	178	1496931	4.690	ng	100
26) Anthracene	17.250	178	1434425	5.257	ng	100
28) Fluoranthene	19.183	202	1744064	4.942	ng	100
30) Pyrene	19.546	202	1846951	5.077	ng	100
32) Benzo(a)anthracene	21.293	228	1796361	5.367	ng	99
33) Chrysene	21.346	228	1694222	4.741	ng	100
34) Bis(2-ethylhexyl)phtha...	21.239	149	630292	4.884	ng	99
36) Indeno(1,2,3-cd)pyrene	25.995	276	853555	3.496	ng	96
37) Benzo(b)fluoranthene	22.924	252	1491840	5.126	ng	99
38) Benzo(k)fluoranthene	22.972	252	1418151	5.024	ng	98
39) Benzo(a)pyrene	23.523	252	1219237	5.077	ng	99
40) Dibenzo(a,h)anthracene	26.019	278	727043	3.627	ng	# 95
41) Benzo(g,h,i)perylene	26.720	276	628151	3.005	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112521\
Data File : BN017590.D
Acq On : 24 Nov 2021 20:52
Operator : CG/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Nov 25 02:38:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN112521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 25 02:34:04 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.754 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 25338
Ion Ratio Lower Upper
152 100
150 156.9 122.5 183.7
115 58.2 47.0 70.4



#2
1,4-Dioxane
Concen: 4.870 ng
RT: 3.301 min Scan# 35
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion: 88 Resp: 59181
Ion Ratio Lower Upper
88 100
43 30.6 0.0 0.0#
58 78.7 0.0 0.0#









#3
 n-Nitrosodimethylamine
 Concen: 4.962 ng
 RT: 3.614 min Scan# 61
 Delta R.T. -0.009 min
 Lab File: BN017590.D
 Acq: 24 Nov 2021 20:52




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion: 42 Resp: 124038
 Ion Ratio Lower Upper
 42 100
 74 130.5 72.1 108.1#
 44 6.8 2.6 4.0#



#4
 2-Fluorophenol
 Concen: 4.820 ng
 RT: 5.339 min Scan# 256
 Delta R.T. -0.009 min
 Lab File: BN017590.D
 Acq: 24 Nov 2021 20:52

Tgt Ion: 112 Resp: 302721
 Ion Ratio Lower Upper
 112 100
 64 64.0 51.5 77.3
 63 32.6 26.6 39.8









#5
Phenol-d6
Concen: 4.878 ng
RT: 6.925 min Scan# 41
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 345599
Ion Ratio Lower Upper
99 100
42 19.7 16.2 24.2
71 34.2 27.6 41.4



#6
bis(2-Chloroethyl)ether
Concen: 4.530 ng
RT: 7.183 min Scan# 456
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion: 93 Resp: 327621
Ion Ratio Lower Upper
93 100
63 78.6 63.0 94.6
95 32.2 25.6 38.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.535 min Scan# 754
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	98425
Ion Ratio	Lower	Upper	
136	100		
137	11.1	8.8	13.2
54	5.6	5.2	7.8
68	5.2	4.6	6.8



#8
Nitrobenzene-d5
Concen: 5.073 ng
RT: 8.887 min Scan# 624
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:	82	Resp:	328797
Ion Ratio	Lower	Upper	
82	100		
128	32.5	25.3	37.9
54	53.1	43.7	65.5









#9
Naphthalene
Concen: 4.723 ng
RT: 10.573 min Scan# 758
Delta R.T. -0.013 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 1179536
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.4 10.2 15.4



#10
Hexachlorobutadiene
Concen: 5.980 ng
RT: 10.877 min Scan# 781
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:225 Resp: 294479
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 51.1 76.7





#11

2-Methylnaphthalene-d10

Concen: 4.922 ng

RT: 12.124 min Scan# 9

Delta R.T. 0.000 min

Lab File: BN017590.D

Acq: 24 Nov 2021 20:52

Instrument :

BNA_N

ClientSampleId :





SSTDICC5.0

Tgt Ion:152 Resp: 806996

Ion Ratio Lower Upper

152 100

151 20.1 16.2 24.4



#12

2-Methylnaphthalene

Concen: 4.727 ng

RT: 12.195 min Scan# 985

Delta R.T. -0.004 min

Lab File: BN017590.D

Acq: 24 Nov 2021 20:52




Tgt Ion:142 Resp: 759593


Ion Ratio Lower Upper

142 100

141 87.1 69.4 104.0

115 35.0 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.372 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 55539
Ion Ratio Lower Upper
164 100
162 101.8 82.5 123.7
160 47.5 39.6 59.4



#14
2,4,6-Tribromophenol
Concen: 9.695 ng
RT: 15.869 min Scan# 1320
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:330 Resp: 165397
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 28.9 24.1 36.1









#15
2-Fluorobiphenyl
Concen: 4.933 ng
RT: 13.001 min Scan# 1094
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 1020293
Ion Ratio Lower Upper
172 100
171 33.8 27.6 41.4
170 22.3 18.3 27.5



#16
Acenaphthylene
Concen: 5.363 ng
RT: 14.092 min Scan# 1180
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:152 Resp: 1188339
Ion Ratio Lower Upper
152 100
151 19.2 15.5 23.3
153 13.2 10.5 15.7









#17
Acenaphthene
Concen: 4.796 ng
RT: 14.435 min Scan# 1207
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 734068
Ion Ratio Lower Upper
154 100
153 110.7 89.8 134.8
152 53.2 45.9 68.9



#18
Fluorene
Concen: 4.927 ng
RT: 15.425 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:166 Resp: 908658
Ion Ratio Lower Upper
166 100
165 96.5 77.6 116.4
167 13.6 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.116 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188 Resp: 114028
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.1 9.0 13.4



#21
4-Bromophenyl-phenylether
Concen: 5.729 ng
RT: 16.313 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:248 Resp: 358731
Ion Ratio Lower Upper
248 100
250 94.0 76.6 114.8
141 55.2 50.5 75.7









#22
Hexachlorobenzene
Concen: 5.669 ng
RT: 16.435 min Scan# 1366
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:284 Resp: 383738
Ion Ratio Lower Upper
284 100
142 29.0 23.2 34.8
249 30.4 24.2 36.2



#25
Phenanthrene
Concen: 4.690 ng
RT: 17.153 min Scan# 1425
Delta R.T. -0.012 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:178 Resp: 1496931
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.6
179 15.4 12.2 18.4









#26
 Anthracene
 Concen: 5.257 ng
 RT: 17.250 min Scan# 1433
 Delta R.T. 0.000 min
 Lab File: BN017590.D
 Acq: 24 Nov 2021 20:52




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:178 Resp: 1434425
 Ion Ratio Lower Upper
 178 100
 176 18.1 14.6 21.8
 179 15.6 12.2 18.4



#27
 Fluoranthene-d10
 Concen: 5.278 ng
 RT: 19.154 min Scan# 1725
 Delta R.T. 0.000 min
 Lab File: BN017590.D
 Acq: 24 Nov 2021 20:52

Tgt Ion:212 Resp: 1851164
 Ion Ratio Lower Upper
 212 100
 106 12.1 9.6 14.4
 104 6.7 5.3 7.9









#28
Fluoranthene
Concen: 4.942 ng
RT: 19.183 min Scan# 1731
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 1744064
Ion Ratio Lower Upper
202 100
101 10.5 8.3 12.5
203 17.6 13.9 20.9



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.314 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:240 Resp: 107517
Ion Ratio Lower Upper
240 100
120 7.2 6.1 9.1
236 26.4 21.2 31.8









#30
Pyrene
Concen: 5.077 ng
RT: 19.546 min Scan# 1804
Delta R.T. -0.005 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 1846951
Ion Ratio Lower Upper
202 100
200 20.6 16.4 24.6
203 17.7 13.9 20.9



#31
Terphenyl-d14
Concen: 5.188 ng
RT: 19.759 min Scan# 1847
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:244 Resp: 1225566
Ion Ratio Lower Upper
244 100
212 7.6 6.2 9.2
122 10.6 8.9 13.3









#32
Benzo(a)anthracene
Concen: 5.367 ng
RT: 21.293 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 1796361
Ion Ratio Lower Upper
228 100
226 27.4 22.4 33.6
229 19.7 15.4 23.2



#33
Chrysene
Concen: 4.741 ng
RT: 21.346 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:228 Resp: 1694222
Ion Ratio Lower Upper
228 100
226 30.1 24.0 36.0
229 19.8 15.6 23.4









#34
Bis(2-ethylhexyl)phthalate
Concen: 4.884 ng
RT: 21.239 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:149 Resp: 630292
Ion Ratio Lower Upper
149 100
167 29.5 23.2 34.8
279 4.2 3.4 5.0



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.626 min Scan# 2479
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:264 Resp: 79993
Ion Ratio Lower Upper
264 100
260 25.0 20.1 30.1
265 24.2 20.6 30.8









#36
Indeno(1,2,3-cd)pyrene
Concen: 3.496 ng
RT: 25.995 min Scan# 3171
Delta R.T. -0.003 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:276 Resp: 853555
Ion Ratio Lower Upper
276 100
138 21.7 16.4 24.6
277 25.1 17.9 26.9



#37
Benzo(b)fluoranthene
Concen: 5.126 ng
RT: 22.924 min Scan# 2274
Delta R.T. -0.003 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:252 Resp: 1491840
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 8.3 7.1 10.7









#38
Benzo(k)fluoranthene
Concen: 5.024 ng
RT: 22.972 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 1418151
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 8.3 7.9 11.9



#39
Benzo(a)pyrene
Concen: 5.077 ng
RT: 23.523 min Scan# 2449
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:252 Resp: 1219237
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 9.1 8.0 12.0









#40
Dibenzo(a,h)anthracene
Concen: 3.627 ng
RT: 26.019 min Scan# 3178
Delta R.T. 0.000 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:278 Resp: 727043
Ion Ratio Lower Upper
278 100
139 14.4 15.2 22.8#
279 23.6 19.5 29.3



#41
Benzo(g,h,i)perylene
Concen: 3.005 ng
RT: 26.720 min Scan# 3383
Delta R.T. -0.003 min
Lab File: BN017590.D
Acq: 24 Nov 2021 20:52

Tgt Ion:276 Resp: 628151
Ion Ratio Lower Upper
276 100
277 23.9 20.0 30.0
138 18.3 15.7 23.5

