

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112522\
 Data File : BN022871.D
 Acq On : 25 Nov 2022 22:48
 Operator : CG/JU
 Sample : N5317-12DL2 100X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PT-BNA-SOILDL2

Quant Time: Nov 25 23:18:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 25 22:56:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.071	152	9897	0.400	ng	0.00
7) Naphthalene-d8	10.893	136	26855	0.400	ng	0.00
13) Acenaphthene-d10	14.709	164	13863	0.400	ng	0.00
19) Phenanthrene-d10	17.452	188	27584	0.400	ng	0.00
29) Chrysene-d12	21.634	240	19720	0.400	ng	# 0.00
35) Perylene-d12	24.126	264	16984	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.601	112	23843	0.927	ng	0.00
5) Phenol-d6	7.212	99	31568	0.964	ng	0.00
8) Nitrobenzene-d5	9.228	82	12479	0.598	ng	-0.01
11) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng	
14) 2,4,6-Tribromophenol	16.198	330	5706	0.853	ng	0.00
15) 2-Fluorobiphenyl	13.341	172	41527	0.655	ng	0.00
27) Fluoranthene-d10	0.000	212	0d	0.000	ng	
31) Terphenyl-d14	20.076	244	32256	0.696	ng	0.00
Target Compounds						Qvalue
6) bis(2-Chloroethyl)ether	7.486	93	43556	1.416	ng	99
10) Hexachlorobutadiene	11.235	225	19033	1.121	ng	# 99
12) 2-Methylnaphthalene	12.159	142	13199	0.782	ng	# 96
16) Acenaphthylene	14.431	152	28095	0.383	ng	100
17) Acenaphthene	14.773	154	20537	0.425	ng	98
18) Fluorene	15.751	166	67018	1.084	ng	100
21) 4-Bromophenyl-phenylether	16.645	248	13337	0.670	ng	# 91
25) Phenanthrene	17.489	178	55470	0.578	ng	100
28) Fluoranthene	19.511	202	58575	0.574	ng	100
30) Pyrene	19.873	202	88174	0.859	ng	97
32) Benzo(a)anthracene	21.616	228	29116	0.365	ng	99
33) Chrysene	21.670	228	41419	0.501	ng	98
34) Bis(2-ethylhexyl)phtha...	21.553	149	41148	1.286	ng	99
37) Benzo(b)fluoranthene	23.366	252	39414	0.520	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112522\
Data File : BN022871.D
Acq On : 25 Nov 2022 22:48
Operator : CG/JU
Sample : N5317-12DL2 100X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PT-BNA-SOILD2

Quant Time: Nov 25 23:18:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 25 22:56:22 2022
Response via : Initial Calibration

