

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112720\
 Data File : BN012749.D
 Acq On : 27 Nov 2020 12:31
 Operator : CG/JU
 Sample : PB133234BL
 Misc : MDL-WATER-BL-05
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK234

Quant Time: Nov 27 13:00:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN112520.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 27 10:53:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.42	152	100459	20.00	ng/ul	0.00
20) Naphthalene-d8	10.19	136	415358	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.08	164	274506	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.84	188	554107	20.00	ng/ul	0.00
79) Chrysene-d12	21.06	240	553912	20.00	ng/ul	0.00
88) Perylene-d12	23.16	264	632438	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	2.96	96	21752	7.80	ng/uL	0.00
4) Pyridine-d5	3.36	84	301132	39.94	ng/ul	0.00
7) Phenol-d5	6.61	99	351280	39.06	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.78	67	225455	40.19	ng/ul	0.00
11) 2-Chlorophenol-d4	6.96	132	286843	40.57	ng/ul	0.00
15) 4-Methylphenol-d8	8.13	113	283269	38.67	ng/ul	0.00
21) Nitrobenzene-d5	8.58	128	137659	41.42	ng/ul	0.00
24) 2-Nitrophenol-d4	9.29	143	150339	40.98	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.82	165	271842	39.00	ng/ul	0.00
31) 4-Chloroaniline-d4	10.33	131	369889	37.94	ng/ul	0.00
46) Dimethylphthalate-d6	13.50	166	929288	41.63	ng/ul	0.00
49) Acenaphthylene-d8	13.77	160	1017666	39.07	ng/ul	0.00
54) 4-Nitrophenol-d4	14.30	143	135325	34.27	ng/ul	0.00
60) Fluorene-d10	15.09	176	760683	40.37	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.22	200	133389	35.21	ng/ul	0.00
73) Anthracene-d10	16.93	188	1188260	42.91	ng/ul	0.00
81) Pyrene-d10	19.25	212	1312603	47.53	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.04	264	1510460	44.21	ng/ul	0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN112720\
Data File : BN012749.D
Acq On : 27 Nov 2020 12:31
Operator : CG/JU
Sample : PB133234BL
Misc : MDL-WATER-BL-05
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK234

Quant Time: Nov 27 13:00:43 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN112520.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 27 10:53:49 2020
Response via : Initial Calibration

