


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112720\
 Data File : BN012745.D
 Acq On : 27 Nov 2020 09:39
 Operator : CG/JU
 Sample : DFTPP21
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP121

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN112520.M
 Title : SVOA CALIBRATION
 Last Update : Fri Nov 27 10:53:49 2020



AutoFind: Scans 2517, 2518, 2519; Background Corrected with Scan 2512

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.2	108826	PASS
68	69	0.00	2	2.0	2047	PASS
69	198	0.00	100	48.2	102557	PASS
70	69	0.00	2	0.8	780	PASS
127	198	10	80	49.0	104293	PASS
197	198	0.00	2	0.4	908	PASS
198	198	100	100	100.0	212693	PASS
199	198	5	9	6.4	13716	PASS
275	198	10	60	27.4	58213	PASS
365	198	1	100	4.2	8967	PASS
441	198	0.01	100	13.5	28687	PASS
442	198	50	100	81.2	172653	PASS
443	442	15	24	18.9	32696	PASS

Scan 2512 (17.263 min): BN012745.D
DFTPP21

m/z	Abundance
39.90	480.0
44.00	1977.0
73.00	449.0
207.00	1353.0
208.10	387.0

Instrument :
BNA_N
ClientSampleId :
DFTPP121

m/z	Abundance
37.00	608.0
38.10	1809.0
39.10	7764.0
40.00	963.0
44.00	2025.0
45.00	373.0
49.00	1246.0
50.00	26808.0
51.00	100792.0
52.00	4970.0
55.10	731.0
56.00	2917.0
57.00	5868.0
61.00	1352.0
62.00	2049.0
63.10	3644.0
64.00	660.0
64.90	1627.0
68.00	2151.0
69.00	90736.0
70.00	676.0
73.00	1462.0
74.00	12492.0
75.00	15895.0
76.10	6144.0
77.10	100216.0
78.10	7102.0
79.00	8716.0
80.00	6070.0
81.00	7761.0
82.00	1686.0
83.00	1149.0
83.90	322.0
85.10	1135.0
86.00	2915.0
87.00	722.0
88.00	909.0
91.00	1408.0
92.00	2269.0
93.00	15276.0
94.00	946.0
96.00	935.0
98.00	12849.0
99.00	8060.0
100.00	608.0
101.00	3366.0
102.90	1265.0
104.00	2356.0
104.90	2558.0
105.90	869.0
107.00	27768.0
108.00	3820.0
109.00	507.0
110.00	47464.0
111.00	8035.0
112.00	1078.0
112.90	371.0
116.00	1529.0
117.00	33800.0
118.00	2962.0
118.90	446.0
119.90	305.0
122.10	1827.0
123.00	3172.0
123.90	1517.0
125.00	1388.0
127.00	83648.0
128.00	7267.0
129.00	35688.0
130.00	2649.0
131.10	641.0
131.80	380.0
132.90	400.0
134.00	1676.0
135.10	3081.0
135.90	1486.0
137.00	2082.0
138.80	306.0
139.80	607.0
141.00	6236.0
142.00	1199.0
143.00	1147.0
145.10	340.0
145.90	954.0
147.00	2651.0
148.00	7141.0
149.00	1379.0
150.00	552.0
151.10	574.0
151.80	600.0
152.90	1607.0
153.90	1387.0

Instrument :
BNA_N
ClientSampleId :
DFTPP121

154.90	2856.0
156.00	4205.0
156.90	860.0
157.90	1124.0
158.90	552.0
159.90	1545.0
161.00	2268.0
162.10	745.0
165.00	1936.0
166.00	1724.0
167.00	10712.0
168.00	7416.0
169.10	1330.0
169.90	484.0
171.10	459.0
171.90	658.0
172.90	908.0
174.00	1886.0
175.00	3558.0
176.10	1044.0
177.00	1960.0
178.10	500.0
178.90	6257.0
180.00	4569.0
181.10	2282.0
184.00	490.0
185.00	3085.0
186.00	21864.0
187.00	7087.0
188.00	936.0
189.00	2152.0
189.90	449.0
191.10	804.0
192.00	1959.0
193.00	2161.0
194.00	417.0
195.90	4260.0
197.10	1190.0
197.90	154048.0
199.00	9785.0
200.10	827.0
201.30	903.0
202.90	1318.0
204.00	6925.0
205.00	9861.0
206.00	42680.0
207.10	7017.0
208.00	1987.0
209.00	979.0
210.00	933.0
211.00	2080.0
211.90	371.0
215.10	489.0
216.00	1019.0
217.00	11497.0
217.90	1759.0
221.00	4946.0
222.00	1597.0
223.10	2436.0
224.00	22408.0
225.00	6153.0
226.00	983.0
227.00	11944.0
227.90	2090.0
229.00	2181.0
229.90	351.0
231.00	853.0
234.10	852.0
235.10	1285.0
235.90	527.0
236.90	1277.0
238.90	467.0
240.00	431.0
241.00	583.0
242.00	1361.0
243.00	1190.0
244.00	17880.0
245.10	2190.0
246.00	4072.0
246.90	789.0
249.10	669.0
252.90	558.0
253.90	682.0
255.00	83304.0
256.00	14492.0
257.00	1278.0
258.00	6375.0
259.10	1017.0
265.00	2779.0
265.80	330.0
271.00	326.0
271.90	350.0
273.00	2165.0
274.00	8159.0
275.00	39840.0
276.00	6733.0

Instrument :
BNA_N
ClientSampleId :
DFTPP121

277.00	3827.0
277.90	690.0
281.20	604.0
283.00	560.0
285.00	509.0
291.90	418.0
293.00	1145.0
296.00	13970.0
297.00	2091.0
301.90	400.0
302.90	1130.0
314.00	525.0
315.00	1623.0
316.00	1038.0
320.90	501.0
323.00	3783.0
323.90	330.0
327.00	842.0
327.80	320.0
334.00	2189.0
335.00	683.0
341.00	740.0
345.90	878.0
352.00	886.0
353.00	851.0
354.00	1435.0
354.80	325.0
364.90	5522.0
366.00	762.0
371.00	488.0
371.90	1462.0
373.10	427.0
382.90	425.0
402.00	796.0
402.90	837.0
421.00	496.0
422.10	817.0
423.10	4142.0
424.00	2201.0
425.10	450.0
441.00	13013.0
442.00	96136.0
443.00	17328.0
444.10	1786.0

Instrument :
BNA_N
ClientSampleId :
DFTPP121

m/z	Abundance
37.00	669.0
38.10	2061.0
39.10	10470.0
40.10	981.0
44.00	2280.0
45.00	466.0
48.90	928.0
50.00	37832.0
51.00	132608.0
52.00	6512.0
53.10	425.0
55.10	626.0
56.00	3523.0
57.00	8208.0
58.00	504.0
61.00	1547.0
62.00	2258.0
63.00	5248.0
64.00	938.0
65.10	2131.0
68.00	2261.0
69.00	122416.0
70.00	970.0
73.10	2202.0
74.00	14506.0
75.00	21240.0
76.10	6716.0
77.00	137152.0
78.00	9439.0
79.00	10559.0
80.00	8082.0
81.00	12155.0
82.00	2783.0
83.00	1895.0
84.00	591.0
84.90	1899.0
86.00	3145.0
87.10	1339.0
87.90	542.0
89.10	471.0
91.00	2243.0
92.00	2746.0
93.00	20136.0
94.00	718.0
96.10	620.0
98.00	17280.0
99.00	11432.0
100.00	1235.0
101.00	5577.0
101.90	583.0
103.00	1926.0
104.00	3887.0
105.00	3703.0
106.00	1289.0
107.00	40504.0
108.00	7205.0
109.20	1270.0
110.00	67072.0
111.00	10557.0
111.90	1085.0
112.90	581.0
116.00	2668.0
117.00	53584.0
118.00	3505.0
119.10	391.0
120.00	603.0
121.90	3806.0
123.00	5580.0
124.00	2329.0
125.00	2294.0
127.00	131328.0
128.00	10136.0
129.00	55672.0
129.90	5103.0
131.00	848.0
132.10	573.0
132.90	313.0
134.00	1724.0
135.00	4898.0
135.90	1710.0
137.00	2822.0
139.90	711.0
141.00	8501.0
142.00	2617.0
143.00	2218.0
144.00	315.0
146.00	1760.0
147.00	5367.0
148.00	11032.0
149.00	2311.0
149.90	759.0
151.10	1354.0

Instrument :
BNA_N
ClientSampleId :
DFTPP121

152.00	448.0
153.00	3083.0
154.00	1820.0
155.00	4459.0
156.00	5685.0
157.00	1019.0
157.80	1577.0
158.90	1095.0
160.00	2692.0
161.00	4236.0
162.00	1175.0
164.00	416.0
165.00	3008.0
166.00	1991.0
167.00	16960.0
168.00	9251.0
169.00	1454.0
170.20	543.0
170.90	589.0
171.90	1594.0
173.10	1661.0
174.00	3801.0
175.10	5527.0
176.00	2045.0
176.90	2857.0
178.00	787.0
179.00	11651.0
180.00	9435.0
181.00	3775.0
181.90	620.0
184.00	854.0
185.00	5477.0
186.00	38528.0
187.00	11276.0
188.00	1125.0
188.90	2989.0
189.90	417.0
191.10	975.0
192.00	3431.0
193.10	3885.0
194.00	1135.0
195.00	428.0
196.00	8255.0
196.90	1536.0
198.00	260288.0
199.00	17784.0
199.90	1575.0
201.50	1138.0
202.20	545.0
203.00	3054.0
204.00	10777.0
205.10	18824.0
206.00	71792.0
207.10	10236.0
208.00	3671.0
208.90	1581.0
210.00	1435.0
210.90	3760.0
212.00	1218.0
215.00	827.0
215.90	2502.0
217.00	19736.0
218.00	2377.0
221.00	9155.0
222.00	2218.0
223.00	4538.0
224.10	39704.0
225.10	10705.0
226.10	1501.0
227.00	20712.0
227.90	3118.0
229.00	3476.0
230.00	554.0
230.90	1522.0
234.00	928.0
235.00	1114.0
236.00	1497.0
236.90	1588.0
238.90	726.0
239.90	824.0
240.90	1179.0
241.90	1990.0
243.00	2907.0
244.00	28776.0
245.00	4153.0
245.90	6676.0
247.00	1415.0
248.00	361.0
249.00	1188.0
253.00	1155.0
255.00	167360.0
256.00	21832.0
257.00	2105.0
258.00	11363.0
259.00	1912.0
260.00	329.0

Instrument :
BNA_N
ClientSampleId :
DFTPP121

263.90	474.0
265.10	4284.0
265.90	754.0
270.80	525.0
272.00	520.0
273.00	5220.0
274.00	11710.0
275.00	67168.0
276.10	8296.0
277.00	7144.0
278.00	871.0
280.80	327.0
283.00	891.0
284.00	320.0
285.00	1300.0
288.80	496.0
290.10	406.0
292.00	371.0
293.00	1078.0
293.90	372.0
296.00	21712.0
297.00	3355.0
301.00	363.0
302.00	665.0
302.90	2485.0
304.00	600.0
308.00	479.0
308.70	315.0
314.00	1176.0
314.90	2626.0
315.90	1440.0
321.10	611.0
322.10	306.0
323.00	6505.0
324.10	1470.0
326.90	1137.0
328.00	472.0
332.00	646.0
333.00	727.0
334.00	4311.0
335.00	1122.0
341.00	1200.0
345.90	1563.0
352.10	1410.0
353.00	905.0
354.00	1766.0
355.00	717.0
365.00	10744.0
365.90	1593.0
370.90	391.0
372.00	3568.0
373.00	820.0
382.90	1044.0
383.90	402.0
389.90	503.0
390.90	403.0
402.10	901.0
402.90	1836.0
403.90	999.0
420.90	1483.0
422.00	1676.0
422.90	10779.0
423.90	2372.0
424.90	360.0
441.00	35808.0
442.00	185088.0
443.00	37080.0
444.00	3267.0

Instrument :
BNA_N
ClientSampleId :
DFTPP121

m/z	Abundance
36.90	336.0
38.00	1819.0
39.10	7099.0
40.00	987.0
44.00	2164.0
45.10	405.0
49.00	1057.0
50.10	26912.0
51.10	93080.0
52.00	3944.0
55.00	423.0
56.00	2834.0
57.00	6155.0
61.10	1101.0
62.00	1535.0
62.90	4436.0
64.10	526.0
65.00	1566.0
68.10	1729.0
69.00	94520.0
70.00	695.0
73.10	1717.0
74.10	11584.0
75.00	16464.0
76.10	5687.0
77.10	102696.0
78.00	5881.0
79.00	8890.0
80.00	5433.0
81.00	6936.0
82.00	2185.0
83.00	1185.0
85.10	1368.0
86.00	3026.0
86.80	765.0
88.10	414.0
91.10	1881.0
92.00	2275.0
93.00	16337.0
94.00	1151.0
96.00	578.0
98.00	11526.0
99.00	9180.0
99.90	500.0
101.00	4009.0
102.90	1662.0
104.00	2557.0
105.00	2812.0
106.10	1117.0
107.00	32216.0
108.00	4741.0
109.10	1021.0
110.00	53016.0
111.00	9066.0
112.00	961.0
113.00	378.0
113.10	373.0
116.00	2289.0
117.00	40440.0
118.00	2237.0
122.00	2199.0
123.00	3836.0
124.00	2275.0
124.90	1655.0
127.00	97904.0
128.00	7216.0
129.00	45272.0
130.00	4309.0
130.90	820.0
131.90	327.0
132.90	361.0
134.00	1228.0
135.00	3440.0
136.00	1548.0
137.00	1537.0
137.90	321.0
138.90	479.0
139.90	456.0
141.00	6871.0
142.00	1854.0
143.00	1252.0
145.00	472.0
146.00	1175.0
147.00	3916.0
148.00	8016.0
149.00	1544.0
149.90	368.0
151.00	637.0
152.00	407.0
153.00	2243.0
153.90	1178.0
155.00	3226.0

Instrument :
BNA_N
ClientSampleId :
DFTPP121

156.00	5419.0
157.10	997.0
158.00	1107.0
158.90	591.0
159.90	1819.0
161.00	2454.0
162.00	879.0
163.00	496.0
164.00	372.0
164.90	2202.0
166.10	1762.0
167.00	14774.0
168.00	7607.0
169.00	1608.0
170.90	488.0
171.90	1410.0
172.90	1757.0
173.90	3001.0
175.00	4426.0
175.90	835.0
177.00	2234.0
178.00	1024.0
179.00	10377.0
180.00	5848.0
181.00	2890.0
182.00	485.0
184.10	941.0
185.00	4768.0
186.00	32016.0
187.00	10669.0
188.00	1052.0
188.90	3110.0
190.00	600.0
190.90	963.0
192.00	3139.0
193.10	3200.0
193.90	884.0
194.80	548.0
196.00	6180.0
198.00	223744.0
199.00	13581.0
200.00	909.0
201.20	576.0
201.50	599.0
202.10	526.0
202.90	2236.0
204.00	9574.0
205.00	14886.0
206.00	55384.0
207.00	9819.0
208.00	3030.0
209.00	964.0
210.30	1182.0
211.00	2485.0
211.90	687.0
215.00	704.0
216.00	1533.0
217.00	16349.0
218.00	1901.0
220.20	684.0
221.00	7920.0
222.00	2654.0
223.00	3279.0
224.00	35344.0
225.00	8138.0
225.90	1224.0
227.00	16936.0
228.00	2774.0
228.90	3432.0
231.00	1280.0
233.90	1132.0
235.00	1026.0
236.00	860.0
237.00	1385.0
239.00	798.0
240.00	771.0
241.00	1079.0
241.90	1954.0
243.00	2149.0
244.00	27072.0
245.00	4111.0
246.00	6089.0
247.00	1420.0
248.90	952.0
251.10	525.0
253.10	773.0
255.00	137024.0
256.00	19696.0
257.00	1891.0
258.00	10465.0
259.00	1711.0
260.90	446.0
264.10	437.0
265.00	3571.0
266.00	965.0
271.00	544.0

Instrument :
BNA_N
ClientSampleId :
DFTPP121

272.00	438.0
272.90	4396.0
274.00	11288.0
275.00	67632.0
276.00	7878.0
277.00	6225.0
277.90	1174.0
279.00	325.0
281.00	666.0
283.00	1096.0
283.90	475.0
285.10	954.0
289.00	318.0
290.00	345.0
291.90	327.0
293.00	1975.0
294.90	370.0
296.00	22056.0
297.00	3077.0
301.90	397.0
303.00	3120.0
304.00	753.0
307.80	576.0
308.80	315.0
314.10	889.0
315.00	2315.0
316.00	1537.0
320.80	618.0
322.00	449.0
323.00	5378.0
323.90	873.0
327.00	1523.0
328.00	553.0
332.00	525.0
333.00	728.0
334.00	4150.0
335.00	1278.0
341.00	844.0
341.80	355.0
346.00	1568.0
352.00	2232.0
353.00	1313.0
354.00	1745.0
355.00	354.0
365.00	10636.0
366.00	1388.0
371.00	739.0
371.90	3044.0
373.00	811.0
382.90	735.0
390.00	695.0
390.90	365.0
402.10	1146.0
403.00	1873.0
404.00	748.0
421.00	1555.0
422.10	1926.0
423.00	11603.0
424.00	3301.0
424.90	447.0
441.00	37240.0
442.00	236736.0
443.00	43680.0
444.00	4152.0

Instrument :
BNA_N
ClientSampleId :
DFTPP121