

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112724\
 Data File : BN035358.D
 Acq On : 27 Nov 2024 20:57
 Operator : RC/JU
 Sample : PB165224BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB165224BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/29/2024
 Supervised By :mohammad ahmed 12/03/2024

Quant Time: Nov 28 00:30:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 27 23:03:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.308	152	1607	0.400	ng	0.00
7) Naphthalene-d8	10.063	136	3806	0.400	ng	0.01
13) Acenaphthene-d10	13.967	164	2822	0.400	ng	0.00
19) Phenanthrene-d10	16.736	188	7341	0.400	ng	# 0.00
29) Chrysene-d12	20.974	240	7132	0.400	ng	0.00
35) Perylene-d12	23.067	264	8113	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.968	112	1264	0.314	ng	0.00
5) Phenol-d6	6.513	99	1363	0.282	ng	0.00
8) Nitrobenzene-d5	8.450	82	887m	0.382	ng	0.01
11) 2-Methylnaphthalene-d10	11.661	152	2528	0.424	ng	0.00
14) 2,4,6-Tribromophenol	15.485	330	349	0.174	ng	0.01
15) 2-Fluorobiphenyl	12.579	172	4240	0.397	ng	0.00
27) Fluoranthene-d10	18.785	212	8826	0.424	ng	0.00
31) Terphenyl-d14	19.412	244	6801	0.483	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112724\
Data File : BN035358.D
Acq On : 27 Nov 2024 20:57
Operator : RC/JU
Sample : PB165224BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB165224BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024

Quant Time: Nov 28 00:30:01 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 27 23:03:24 2024
Response via : Initial Calibration

