

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112724\
 Data File : BN035368.D
 Acq On : 28 Nov 2024 04:51
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 28 05:19:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 27 23:03:24 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.393	1.8	100	0.00
3	n-Nitrosodimethylamine	0.400	0.411	-2.7	100	0.00
4 S	2-Fluorophenol	0.400	0.412	-3.0	101	0.00
5 S	Phenol-d6	0.400	0.423	-5.7	107	0.00
6	bis(2-Chloroethyl)ether	0.400	0.390	2.5	100	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	103	0.00
8 S	Nitrobenzene-d5	0.400	0.395	1.3	105	0.00
9	Naphthalene	0.400	0.395	1.3	102	0.00
10	Hexachlorobutadiene	0.400	0.412	-3.0	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.394	1.5	102	0.00
12	2-Methylnaphthalene	0.400	0.397	0.8	104	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.400	0.394	1.5	112	0.00
15 S	2-Fluorobiphenyl	0.400	0.402	-0.5	104	0.00
16	Acenaphthylene	0.400	0.388	3.0	105	0.00
17	Acenaphthene	0.400	0.392	2.0	104	0.00
18	Fluorene	0.400	0.392	2.0	104	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	103	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.364	9.0	104	0.00
21	4-Bromophenyl-phenylether	0.400	0.388	3.0	103	0.00
22	Hexachlorobenzene	0.400	0.405	-1.3	105	0.00
23	Atrazine	0.400	0.362	9.5	101	0.00
24	Pentachlorophenol	0.400	0.380	5.0	122	0.00
25	Phenanthrene	0.400	0.389	2.8	103	0.00
26	Anthracene	0.400	0.380	5.0	103	0.00
27 SURR	Fluoranthene-d10	0.400	0.376	6.0	102	0.00
28	Fluoranthene	0.400	0.375	6.3	101	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
30	Pyrene	0.400	0.404	-1.0	101	0.00
31 S	Terphenyl-d14	0.400	0.399	0.3	99	0.00
32	Benzo(a)anthracene	0.400	0.381	4.8	98	0.00
33	Chrysene	0.400	0.403	-0.8	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.365	8.8	98	0.00
35 I	Perylene-d12	0.400	0.400	0.0	96	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.392	2.0	96	0.00
37	Benzo(b)fluoranthene	0.400	0.431	-7.7	115	0.00
38	Benzo(k)fluoranthene	0.400	0.391	2.3	97	0.00
39 C	Benzo(a)pyrene	0.400	0.378	5.5	96	0.00
40	Dibenzo(a,h)anthracene	0.400	0.390	2.5	97	0.00
41	Benzo(g,h,i)perylene	0.400	0.383	4.3	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4