

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112820\
 Data File : BN012753.D
 Acq On : 27 Nov 2020 15:50
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 27 16:39:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 27 15:36:21 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	135	0.00
2	1,4-Dioxane	0.400	0.412	-3.0	147	0.00
3	n-Nitrosodimethylamine	0.400	0.188	53.0#	70	0.00
4 S	2-Fluorophenol	0.400	0.310	22.5	112	0.00
5 S	Phenol-d6	0.400	0.319	20.3	113	0.00
6	bis(2-Chloroethyl)ether	0.400	0.351	12.3	135	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	128	0.00
8 S	Nitrobenzene-d5	0.400	0.277	30.8#	99	0.00
9	Naphthalene	0.400	0.375	6.3	134	0.00
10	Hexachlorobutadiene	0.400	0.371	7.3	130	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.392	2.0	143	0.00
12	2-Methylnaphthalene	0.400	0.393	1.8	142	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	139	0.00
14 S	2,4,6-Tribromophenol	0.400	0.235	41.3#	95	0.00
15 S	2-Fluorobiphenyl	0.400	0.390	2.5	149	0.00
16	Acenaphthylene	0.400	0.340	15.0	136	-0.01
17	Acenaphthene	0.400	0.356	11.0	142	0.00
18	Fluorene	0.400	0.368	8.0	146	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	147	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.229	42.8#	102	0.00
21	4-Bromophenyl-phenylether	0.400	0.302	24.5	127	0.00
22	Hexachlorobenzene	0.400	0.355	11.3	144	0.00
23	Atrazine	0.400	0.304	24.0	127	0.00
24	Pentachlorophenol	0.400	0.251	37.3#	111	-0.01
25	Phenanthrene	0.400	0.379	5.3	157	-0.01
26	Anthracene	0.400	0.331	17.3	139	0.00
27 SURR	Fluoranthene-d10	0.400	0.355	11.3	149	0.00
28	Fluoranthene	0.400	0.372	7.0	156	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	152	0.00
30	Pyrene	0.400	0.371	7.3	157	0.00
31 S	Terphenyl-d14	0.400	0.325	18.8	136	0.00
32	Benzo(a)anthracene	0.400	0.336	16.0	141	0.01
33	Chrysene	0.400	0.386	3.5	163	0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.261	34.8#	114	0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.330	17.5	130	0.00
36 I	Perylene-d12	0.400	0.400	0.0	128	0.00
37	Benzo(b)fluoranthene	0.400	0.417	-4.2	152	0.00
38	Benzo(k)fluoranthene	0.400	0.415	-3.7	147	0.00
39 C	Benzo(a)pyrene	0.400	0.393	1.8	141	0.00
40	Dibenzo(a,h)anthracene	0.400	0.372	7.0	126	0.00
41	Benzo(g,h,i)perylene	0.400	0.385	3.8	130	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4