

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113020\
 Data File : BN012764.D
 Acq On : 30 Nov 2020 13:02
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 30 13:40:30 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 27 15:36:21 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	156	0.00
2	1,4-Dioxane	0.400	0.437	-9.2	181	0.00
3	n-Nitrosodimethylamine	0.400	0.194	51.5#	84	0.00
4 S	2-Fluorophenol	0.400	0.343	14.2	143	-0.01
5 S	Phenol-d6	0.400	0.359	10.3	148	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.371	7.3	166	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	158	0.00
8 S	Nitrobenzene-d5	0.400	0.289	27.8#	128	0.00
9	Naphthalene	0.400	0.376	6.0	166	-0.01
10	Hexachlorobutadiene	0.400	0.359	10.3	156	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.395	1.3	178	0.00
12	2-Methylnaphthalene	0.400	0.390	2.5	174	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	177	0.00
14 S	2,4,6-Tribromophenol	0.400	0.256	36.0#	132	0.00
15 S	2-Fluorobiphenyl	0.400	0.371	7.3	182	0.00
16	Acenaphthylene	0.400	0.324	19.0	166	-0.01
17	Acenaphthene	0.400	0.344	14.0	176	0.00
18	Fluorene	0.400	0.361	9.8	184	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	189	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.206	48.5#	118	-0.01
21	4-Bromophenyl-phenylether	0.400	0.277	30.8#	150	0.00
22	Hexachlorobenzene	0.400	0.338	15.5	176	0.00
23	Atrazine	0.400	0.296	26.0#	159	0.00
24	Pentachlorophenol	0.400	0.292	27.0#	166	-0.01
25	Phenanthrene	0.400	0.367	8.3	195	-0.01
26	Anthracene	0.400	0.318	20.5	172	0.00
27 SURR	Fluoranthene-d10	0.400	0.341	14.8	184	0.00
28	Fluoranthene	0.400	0.345	13.8	186	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	180	0.00
30	Pyrene	0.400	0.366	8.5	184	0.00
31 S	Terphenyl-d14	0.400	0.329	17.8	162	0.00
32	Benzo(a)anthracene	0.400	0.345	13.8	172	0.00
33	Chrysene	0.400	0.400	0.0	200	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.260	35.0#	134	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.309	22.8	144	0.01
36 I	Perylene-d12	0.400	0.400	0.0	156	0.00
37	Benzo(b)fluoranthene	0.400	0.386	3.5	172	0.00
38	Benzo(k)fluoranthene	0.400	0.403	-0.8	175	0.00
39 C	Benzo(a)pyrene	0.400	0.359	10.3	158	0.00
40	Dibenzo(a,h)anthracene	0.400	0.338	15.5	140	0.00
41	Benzo(g,h,i)perylene	0.400	0.344	14.0	142	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4