

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113022\
 Data File : BN023004.D
 Acq On : 01 Dec 2022 04:44
 Operator : CG/JU
 Sample : N5756-08
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT172S1-20221118

Quant Time: Dec 01 05:18:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 30 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.057	152	5761	0.400	ng	0.00
7) Naphthalene-d8	10.872	136	15524	0.400	ng	#-0.01
13) Acenaphthene-d10	14.698	164	7298	0.400	ng	0.00
19) Phenanthrene-d10	17.439	188	13697	0.400	ng	0.00
29) Chrysene-d12	21.625	240	8144	0.400	ng	# 0.00
35) Perylene-d12	24.109	264	5902	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.594	112	1484	0.099	ng	0.00
5) Phenol-d6	7.204	99	1083	0.057	ng	0.00
8) Nitrobenzene-d5	9.217	82	3981	0.330	ng	-0.01
11) 2-Methylnaphthalene-d10	12.461	152	6901	0.205	ng	0.00
14) 2,4,6-Tribromophenol	16.186	330	856	0.243	ng	0.00
15) 2-Fluorobiphenyl	13.330	172	12226	0.366	ng	0.00
27) Fluoranthene-d10	19.469	212	9491	0.255	ng	0.00
31) Terphenyl-d14	20.058	244	7535	0.394	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.442	88	10727	1.467	ng	99
34) Bis(2-ethylhexyl)phtha...	21.527	149	1244	0.094	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113022\
Data File : BN023004.D
Acq On : 01 Dec 2022 04:44
Operator : CG/JU
Sample : N5756-08
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT172S1-20221118

Quant Time: Dec 01 05:18:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 30 00:45:21 2022
Response via : Initial Calibration

