

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113022\
 Data File : BN023010.D
 Acq On : 01 Dec 2022 08:24
 Operator : CG/JU
 Sample : N5756-09
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW181S-20221117

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 12/02/2022
 Supervised By :mohammad ahmed 12/02/2022

Quant Time: Dec 01 13:38:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 01 11:51:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.056	152	5782	0.400	ng	0.00
7) Naphthalene-d8	10.872	136	15775	0.400	ng	#-0.01
13) Acenaphthene-d10	14.698	164	7437	0.400	ng	0.00
19) Phenanthrene-d10	17.439	188	13903	0.400	ng	0.00
29) Chrysene-d12	21.625	240	7975	0.400	ng	# 0.00
35) Perylene-d12	24.109	264	5907	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.594	112	1544	0.103	ng	0.00
5) Phenol-d6	7.204	99	1181	0.062	ng	0.00
8) Nitrobenzene-d5	9.217	82	3592m	0.293	ng	-0.01
11) 2-Methylnaphthalene-d10	12.457	152	6418	0.188	ng	0.00
14) 2,4,6-Tribromophenol	16.185	330	768	0.214	ng	0.00
15) 2-Fluorobiphenyl	13.330	172	11072	0.325	ng	0.00
27) Fluoranthene-d10	19.469	212	9521	0.252	ng	0.00
31) Terphenyl-d14	20.058	244	7061	0.377	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.441	88	1024	0.140	ng	# 88
34) Bis(2-ethylhexyl)phtha...	21.535	149	882	0.068	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113022\
Data File : BN023010.D
Acq On : 01 Dec 2022 08:24
Operator : CG/JU
Sample : N5756-09
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW181S-20221117

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 12/02/2022
Supervised By : mohammad ahmed 12/02/2022

Quant Time: Dec 01 13:38:03 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 01 11:51:26 2022
Response via : Initial Calibration

