




Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN113022\  
Data File : BN022995.D  
Acq On : 30 Nov 2022 21:55  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 01 04:41:45 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Dec 01 04:40:35 2022  
Response via : Initial Calibration



TIC: BN022995.D\data.ms

**(70) Pentachlorophenol (C)**




17.092min (-0.000) 18582.19 ng

response 462130

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 66.60  |
| 264.00 | 61.60  | 65.32  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN113022\  
Data File : BN022995.D  
Acq On : 30 Nov 2022 21:55  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 01 04:41:45 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Dec 01 04:40:35 2022  
Response via : Initial Calibration



TIC: BN022995.D\data.ms

**(77) Benzidine**

19.686min (-0.000) 0.00 ng

response 2763971

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 14.09  |
| 183.00 | 13.20  | 11.56  |
| 0.00   | 0.00   | 0.00   |