

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113022\
 Data File : BN022979.D
 Acq On : 30 Nov 2022 10:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 30 22:58:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 30 00:45:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	74	0.00
2	1,4-Dioxane	0.400	0.341	14.8	61	0.00
3	n-Nitrosodimethylamine	0.400	0.327	18.3	59	0.00
4 S	2-Fluorophenol	0.400	0.318	20.5	57	0.00
5 S	Phenol-d6	0.400	0.323	19.3	59	0.00
6	bis(2-Chloroethyl)ether	0.400	0.376	6.0	66	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	74	0.00
8 S	Nitrobenzene-d5	0.400	0.361	9.8	67	-0.01
9	Naphthalene	0.400	0.360	10.0	65	-0.01
10	Hexachlorobutadiene	0.400	0.376	6.0	67	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.345	13.8	65	0.00
12	2-Methylnaphthalene	0.400	0.288	28.0#	56	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	67	0.00
14 S	2,4,6-Tribromophenol	0.400	0.285	28.8#	48	0.00
15 S	2-Fluorobiphenyl	0.400	0.387	3.3	61	0.00
16	Acenaphthylene	0.400	0.318	20.5	53	0.00
17	Acenaphthene	0.400	0.364	9.0	59	0.00
18	Fluorene	0.400	0.355	11.3	56	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	62	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.336	16.0	55	0.00
21	4-Bromophenyl-phenylether	0.400	0.353	11.8	53	0.00
22	Hexachlorobenzene	0.400	0.384	4.0	57	0.00
23	Atrazine	0.400	0.291	27.3#	46	0.00
24	Pentachlorophenol	0.400	0.293	26.8#	50	0.00
25	Phenanthrene	0.400	0.362	9.5	54	0.00
26	Anthracene	0.400	0.322	19.5	50	0.00
27 SURR	Fluoranthene-d10	0.400	0.364	9.0	56	0.00
28	Fluoranthene	0.400	0.327	18.3	49	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	54	0.00
30	Pyrene	0.400	0.375	6.3	48	0.00
31 S	Terphenyl-d14	0.400	0.378	5.5	47	0.00
32	Benzo(a)anthracene	0.400	0.329	17.8	44	0.00
33	Chrysene	0.400	0.377	5.8	48	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.357	10.8	54	0.00
35 I	Perylene-d12	0.400	0.400	0.0	47	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.416	-4.0	49	0.00
37	Benzo(b)fluoranthene	0.400	0.412	-3.0	48	0.00
38	Benzo(k)fluoranthene	0.400	0.384	4.0	43	0.00
39 C	Benzo(a)pyrene	0.400	0.362	9.5	43	0.00
40	Dibenzo(a,h)anthracene	0.400	0.422	-5.5	50	0.01
41	Benzo(g,h,i)perylene	0.400	0.426	-6.5	49	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4