

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120121\
 Data File : BN017652.D
 Acq On : 01 Dec 2021 15:54
 Operator : CG/JU
 Sample : PB141079BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB141079BL

Quant Time: Dec 01 16:22:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN113021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 30 18:26:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.773	152	20683	0.400	ng	0.00
7) Naphthalene-d8	10.547	136	80751	0.400	ng	# 0.00
13) Acenaphthene-d10	14.384	164	47589	0.400	ng	0.00
19) Phenanthrene-d10	17.128	188	102002	0.400	ng	0.00
29) Chrysene-d12	21.303	240	95098	0.400	ng	#-0.01
35) Perylene-d12	23.612	264	63877	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.394	112	14922	0.318	ng	0.03
5) Phenol-d6	6.952	99	13938	0.281	ng	0.00
8) Nitrobenzene-d5	8.912	82	12010	0.350	ng	0.00
11) 2-Methylnaphthalene-d10	12.142	152	41454	0.294	ng	0.00
14) 2,4,6-Tribromophenol	15.881	330	3200	0.232	ng	0.00
15) 2-Fluorobiphenyl	13.014	172	56558	0.305	ng	0.00
27) Fluoranthene-d10	19.153	212	96768	0.303	ng	0.00
31) Terphenyl-d14	19.759	244	76062	0.365	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120121\
Data File : BN017652.D
Acq On : 01 Dec 2021 15:54
Operator : CG/JU
Sample : PB141079BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB141079BL

Quant Time: Dec 01 16:22:07 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN113021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 30 18:26:44 2021
Response via : Initial Calibration

