

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/1/2021	BNA_N	<u>BN017663.D</u>
Compound Name	Response	Retention Time
DDT	1750548	20.586
DDD	207727	20.145
DDE	732	19.645
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
208459	1959007	10.64