



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120121\
 Data File : BN017650.D
 Acq On : 01 Dec 2021 14:41
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN113021.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Dec 02 01:51:49 2021






Spectrum Information: Average of 17.563 to 17.575 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.4	514815	PASS
68	69	0.00	2	1.9	10100	PASS
69	198	0.00	100	35.1	541547	PASS
70	69	0.00	2	0.5	2938	PASS
127	198	10	80	37.3	575275	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	1541120	PASS
199	198	5	9	6.7	103621	PASS
275	198	10	60	28.1	433728	PASS
365	198	1	100	3.7	56416	PASS
441	198	0.01	100	12.8	197184	PASS
442	442	50	100	100.0	1312597	PASS
443	442	15	24	18.5	243029	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120121\
Data File : BN017650.D
Acq On : 01 Dec 2021 14:41
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 02 02:31:25 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 02 02:31:15 2021
Response via : Initial Calibration



TIC: BN017650.D\data.ms

(70) Pentachlorophenol (C)

16.757min (0.000) 31334.70 ng




response 2127832

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	66.20
264.00	61.60	66.99
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120121\
Data File : BN017650.D
Acq On : 01 Dec 2021 14:41
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 02 02:31:25 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 02 02:31:15 2021
Response via : Initial Calibration



TIC: BN017650.D\data.ms

(77) Benzidine

19.351min (0.000) 2729508.43 ng

response 9760021

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.67
183.00	13.20	11.22
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/1/2021	BNA_N	<u>BN017650.D</u>
Compound Name	Response	Retention Time
DDT	2335908	20.59
DDD	1753587	20.15
DDE	2873	19.64
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
1756460	4092368	42.92

Instrument :
BNA_N
ClientSampleId :
DFTPP