



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120122\
 Data File : BN023028.D
 Acq On : 01 Dec 2022 22:15
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Dec 02 02:30:10 2022






AutoFind: Scans 2514, 2515, 2516; Background Corrected with Scan 2507

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.1	143736	PASS
68	69	0.00	2	1.8	2605	PASS
69	198	0.00	100	44.6	148672	PASS
70	69	0.00	2	0.5	784	PASS
127	198	10	80	53.6	178451	PASS
197	198	0.00	2	0.5	1723	PASS
198	198	100	100	100.0	333120	PASS
199	198	5	9	6.8	22741	PASS
275	198	10	60	26.2	87400	PASS
365	198	1	100	3.5	11523	PASS
441	198	0.01	100	11.8	39421	PASS
442	442	50	100	100.0	229760	PASS
443	442	15	24	21.1	48389	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120122\
Data File : BN023028.D
Acq On : 01 Dec 2022 22:15
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 02 06:08:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 02 06:06:28 2022
Response via : Initial Calibration



TIC: BN023028.D\data.ms

(70) Pentachlorophenol (C)

17.086min (-0.000) 16055.75 ng




response 331510

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	66.27
264.00	61.60	64.63
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120122\
 Data File : BN023028.D
 Acq On : 01 Dec 2022 22:15
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Dec 02 06:08:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 02 06:06:28 2022
 Response via : Initial Calibration



TIC: BN023028.D\data.ms

(77) Benzidine

19.680min (-0.006) 0.00 ng

response 2133664

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.35
183.00	13.20	11.80
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/1/2022	BNA_N	<u>BN023028.D</u>
Compound Name	Response	Retention Time
DDT	1001073	20.916
DDD	12870	20.469
DDE	641	19.969
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
13511	1014584	1.33

Instrument :
BNA_N
ClientSampleId :
DFTPP