

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120221\
 Data File : BN017682.D
 Acq On : 02 Dec 2021 14:22
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 02 15:28:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN113021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 02 01:51:49 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	75	0.00
2	1,4-Dioxane	0.400	0.321	19.8	59	0.00
3	n-Nitrosodimethylamine	0.400	0.440	-10.0	83	-0.05
4 S	2-Fluorophenol	0.400	0.431	-7.7	81	-0.03
5 S	Phenol-d6	0.400	0.433	-8.2	84	0.00
6	bis(2-Chloroethyl)ether	0.400	0.374	6.5	69	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	78	0.00
8 S	Nitrobenzene-d5	0.400	0.475	-18.7	99	-0.01
9	Naphthalene	0.400	0.392	2.0	75	-0.01
10	Hexachlorobutadiene	0.400	0.412	-3.0	79	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.410	-2.5	79	0.00
12	2-Methylnaphthalene	0.400	0.410	-2.5	77	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	82	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.428	-7.0	93	0.00
15 S	2-Fluorobiphenyl	0.400	0.391	2.3	78	-0.01
16	Acenaphthylene	0.400	0.390	2.5	84	-0.01
17	Acenaphthene	0.400	0.379	5.3	78	-0.01
18	Fluorene	0.400	0.399	0.3	81	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	83	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.174	56.5#	27	0.00
21	4-Bromophenyl-phenylether	0.400	0.384	4.0	81	0.00
22	Hexachlorobenzene	0.400	0.394	1.5	80	-0.01
23	Atrazine	0.400	0.439	-9.7	110	0.00
24	Pentachlorophenol	0.400	0.414	-3.5	92	0.00
25	Phenanthrene	0.400	0.378	5.5	78	0.00
26	Anthracene	0.400	0.386	3.5	83	-0.01
27 SURR	Fluoranthene-d10	0.400	0.404	-1.0	86	0.00
28	Fluoranthene	0.400	0.407	-1.7	83	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	92	0.01
30	Pyrene	0.400	0.370	7.5	84	0.00
31 S	Terphenyl-d14	0.400	0.382	4.5	86	0.00
32	Benzo(a)anthracene	0.400	0.389	2.8	93	0.00
33	Chrysene	0.400	0.367	8.3	82	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.539	-34.8#	135	0.00
35 I	Perylene-d12	0.400	0.400	0.0	104	0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.399	0.3	103	0.01
37	Benzo(b)fluoranthene	0.400	0.361	9.8	95	0.01
38	Benzo(k)fluoranthene	0.400	0.369	7.8	94	0.00
39 C	Benzo(a)pyrene	0.400	0.377	5.8	99	0.01
40	Dibenzo(a,h)anthracene	0.400	0.388	3.0	100	0.01
41	Benzo(g,h,i)perylene	0.400	0.420	-5.0	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4