

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120222\
 Data File : BN023046.D
 Acq On : 02 Dec 2022 12:40
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 05 00:33:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 02 02:30:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	74	0.00
2	1,4-Dioxane	0.400	0.377	5.8	68	0.00
3	n-Nitrosodimethylamine	0.400	0.342	14.5	62	0.00
4 S	2-Fluorophenol	0.400	0.342	14.5	62	0.00
5 S	Phenol-d6	0.400	0.346	13.5	64	0.00
6	bis(2-Chloroethyl)ether	0.400	0.406	-1.5	72	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	75	0.00
8 S	Nitrobenzene-d5	0.400	0.394	1.5	74	0.00
9	Naphthalene	0.400	0.392	2.0	71	-0.01
10	Hexachlorobutadiene	0.400	0.403	-0.8	73	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.360	10.0	68	0.00
12	2-Methylnaphthalene	0.400	0.318	20.5	63	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	71	0.00
14 S	2,4,6-Tribromophenol	0.400	0.327	18.3	59	-0.01
15 S	2-Fluorobiphenyl	0.400	0.408	-2.0	69	0.00
16	Acenaphthylene	0.400	0.335	16.3	60	0.00
17	Acenaphthene	0.400	0.392	2.0	67	0.00
18	Fluorene	0.400	0.377	5.8	64	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	66	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.428	-7.0	75	0.00
21	4-Bromophenyl-phenylether	0.400	0.370	7.5	59	-0.01
22	Hexachlorobenzene	0.400	0.405	-1.3	64	0.00
23	Atrazine	0.400	0.314	21.5	53	0.00
24	Pentachlorophenol	0.400	0.366	8.5	67	0.00
25	Phenanthrene	0.400	0.389	2.8	62	0.00
26	Anthracene	0.400	0.342	14.5	56	-0.01
27 SURR	Fluoranthene-d10	0.400	0.400	0.0	65	0.00
28	Fluoranthene	0.400	0.360	10.0	58	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	59	0.00
30	Pyrene	0.400	0.400	0.0	56	0.00
31 S	Terphenyl-d14	0.400	0.408	-2.0	55	0.00
32	Benzo(a)anthracene	0.400	0.348	13.0	51	0.00
33	Chrysene	0.400	0.406	-1.5	57	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.402	-0.5	66	0.00
35 I	Perylene-d12	0.400	0.400	0.0	56	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.468	-17.0	66	0.00
37	Benzo(b)fluoranthene	0.400	0.430	-7.5	60	0.00
38	Benzo(k)fluoranthene	0.400	0.387	3.3	52	0.00
39 C	Benzo(a)pyrene	0.400	0.403	-0.8	57	0.00
40	Dibenzo(a,h)anthracene	0.400	0.463	-15.8	65	0.00
41	Benzo(g,h,i)perylene	0.400	0.453	-13.2	62	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4