

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120222\
 Data File : BN023054.D
 Acq On : 02 Dec 2022 19:34
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 05 00:35:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 02 02:30:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	65	0.00
2	1,4-Dioxane	0.400	0.367	8.3	58	0.00
3	n-Nitrosodimethylamine	0.400	0.352	12.0	55	0.00
4 S	2-Fluorophenol	0.400	0.331	17.3	52	0.00
5 S	Phenol-d6	0.400	0.343	14.2	55	0.00
6	bis(2-Chloroethyl)ether	0.400	0.410	-2.5	63	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	65	0.00
8 S	Nitrobenzene-d5	0.400	0.416	-4.0	68	0.00
9	Naphthalene	0.400	0.397	0.8	63	0.00
10	Hexachlorobutadiene	0.400	0.412	-3.0	64	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.346	13.5	57	0.00
12	2-Methylnaphthalene	0.400	0.315	21.3	54	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	59	0.00
14 S	2,4,6-Tribromophenol	0.400	0.321	19.8	48	-0.01
15 S	2-Fluorobiphenyl	0.400	0.416	-4.0	58	0.00
16	Acenaphthylene	0.400	0.347	13.3	51	0.00
17	Acenaphthene	0.400	0.394	1.5	56	0.00
18	Fluorene	0.400	0.378	5.5	53	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	51	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.461	-15.3	63	0.00
21	4-Bromophenyl-phenylether	0.400	0.374	6.5	47	0.00
22	Hexachlorobenzene	0.400	0.431	-7.7	53	0.00
23	Atrazine	0.400	0.324	19.0	43	0.00
24	Pentachlorophenol	0.400	0.348	13.0	49	0.00
25	Phenanthrene	0.400	0.399	0.3	50	0.00
26	Anthracene	0.400	0.351	12.3	45	-0.01
27 SURR	Fluoranthene-d10	0.400	0.387	3.3	49	0.00
28	Fluoranthene	0.400	0.351	12.3	44	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	43	0.00
30	Pyrene	0.400	0.426	-6.5	44	0.00
31 S	Terphenyl-d14	0.400	0.392	2.0	39	0.00
32	Benzo(a)anthracene	0.400	0.368	8.0	39	0.00
33	Chrysene	0.400	0.421	-5.2	43	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.445	-11.2	53	0.00
35 I	Perylene-d12	0.400	0.400	0.0	42	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.509	-27.2#	53	0.00
37	Benzo(b)fluoranthene	0.400	0.463	-15.8	48	0.00
38	Benzo(k)fluoranthene	0.400	0.420	-5.0	42	0.00
39 C	Benzo(a)pyrene	0.400	0.434	-8.5	45	0.00
40	Dibenzo(a,h)anthracene	0.400	0.502	-25.5#	52	0.00
41	Benzo(g,h,i)perylene	0.400	0.471	-17.7	48	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4