

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
 Data File : BN017696.D
 Acq On : 03 Dec 2021 17:19
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Dec 03 22:25:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 03 17:44:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.754	152	20983	0.400	ng	0.00
7) Naphthalene-d8	10.535	136	83137	0.400	ng	# 0.00
13) Acenaphthene-d10	14.372	164	49459	0.400	ng	0.00
19) Phenanthrene-d10	17.116	188	99903	0.400	ng	0.00
29) Chrysene-d12	21.303	240	107539	0.400	ng	# 0.00
35) Perylene-d12	23.612	264	85659	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	84040	1.763	ng	0.00
5) Phenol-d6	6.925	99	88462	1.757	ng	0.00
8) Nitrobenzene-d5	8.900	82	80061	2.269	ng	0.00
11) 2-Methylnaphthalene-d10	12.124	152	249354	1.718	ng	0.00
14) 2,4,6-Tribromophenol	15.869	330	44917	1.784	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	318915	1.656	ng	0.00
27) Fluoranthene-d10	19.149	212	578973	1.854	ng	0.00
31) Terphenyl-d14	19.754	244	393773	1.670	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.310	88	25317	1.796	ng	95
3) n-Nitrosodimethylamine	3.633	42	35559	1.805	ng	98
6) bis(2-Chloroethyl)ether	7.183	93	89577	1.533	ng	95
9) Naphthalene	10.585	128	344311	1.642	ng	100
10) Hexachlorobutadiene	10.890	225	100947	1.746	ng	# 100
12) 2-Methylnaphthalene	12.200	142	232770	1.683	ng	97
16) Acenaphthylene	14.092	152	340086	1.763	ng	100
17) Acenaphthene	14.435	154	220340	1.652	ng	98
18) Fluorene	15.425	166	278174	1.716	ng	100
20) 4,6-Dinitro-2-methylph...	15.501	198	7954	1.292	ng	# 90
21) 4-Bromophenyl-phenylether	16.313	248	107397	1.750	ng	# 74
22) Hexachlorobenzene	16.435	284	124886	1.678	ng	# 92
23) Atrazine	16.581	200	75023	2.172	ng	97
24) Pentachlorophenol	16.776	266	38690	1.549	ng	99
25) Phenanthrene	17.153	178	455689	1.663	ng	100
26) Anthracene	17.250	178	415933	1.767	ng	100
28) Fluoranthene	19.178	202	557303	1.797	ng	# 97
30) Pyrene	19.541	202	564941	1.627	ng	100
32) Benzo(a)anthracene	21.293	228	555863	1.702	ng	99
33) Chrysene	21.346	228	575150	1.619	ng	99
34) Bis(2-ethylhexyl)phtha...	21.239	149	194540	2.511	ng	96
36) Indeno(1,2,3-cd)pyrene	25.978	276	371747	1.336	ng	# 93
37) Benzo(b)fluoranthene	22.914	252	497745	1.752	ng	# 97
38) Benzo(k)fluoranthene	22.959	252	489306	1.680	ng	# 97
39) Benzo(a)pyrene	23.510	252	428965	1.687	ng	# 96
40) Dibenzo(a,h)anthracene	25.995	278	281888	1.240	ng	# 96
41) Benzo(g,h,i)perylene	26.700	276	337037	1.441	ng	# 94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
Data File : BN017696.D
Acq On : 03 Dec 2021 17:19
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Dec 03 22:25:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 03 17:44:19 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.754 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: BN017696.D
 Acq: 03 Dec 2021 17:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:152 Resp: 20983
 Ion Ratio Lower Upper
 152 100
 150 155.6 123.0 184.4
 115 58.6 45.6 68.4



#2
 1,4-Dioxane
 Concen: 1.796 ng
 RT: 3.310 min Scan# 36
 Delta R.T. -0.009 min
 Lab File: BN017696.D
 Acq: 03 Dec 2021 17:19

Tgt Ion: 88 Resp: 25317
 Ion Ratio Lower Upper
 88 100
 43 35.5 24.7 37.1
 58 75.6 63.0 94.6









#3
n-Nitrosodimethylamine
Concen: 1.805 ng
RT: 3.633 min Scan# 71
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	42	Resp:	35559
Ion	Ratio	Lower	Upper
42	100		
74	120.1	97.8	146.6
44	6.2	5.5	8.3



#4
2-Fluorophenol
Concen: 1.763 ng
RT: 5.348 min Scan# 257
Delta R.T. -0.009 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Tgt Ion:	112	Resp:	84040
Ion	Ratio	Lower	Upper
112	100		
64	61.9	47.6	71.4
63	33.6	24.6	37.0





#5

Phenol-d6

Concen: 1.757 ng

RT: 6.925 min Scan# 41

Delta R.T. 0.000 min

Lab File: BN017696.D

Acq: 03 Dec 2021 17:19

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion: 99 Resp: 88462

Ion Ratio Lower Upper

99 100

42 21.3 17.0 25.6

71 37.8 24.8 37.2#



#6

bis(2-Chloroethyl)ether

Concen: 1.533 ng

RT: 7.183 min Scan# 456

Delta R.T. 0.000 min

Lab File: BN017696.D

Acq: 03 Dec 2021 17:19




Tgt Ion: 93 Resp: 89577


Ion Ratio Lower Upper

93 100

63 76.0 65.8 98.6

95 32.3 26.6 39.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.535 min Scan# 754
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	83137
Ion Ratio	Lower	Upper	
136	100		
137	11.2	8.8	13.2
54	6.6	4.0	6.0#
68	5.2	3.0	4.6#



#8
Nitrobenzene-d5
Concen: 2.269 ng
RT: 8.900 min Scan# 625
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Tgt Ion:	82	Resp:	80061
Ion Ratio	Lower	Upper	
82	100		
128	41.3	27.7	41.5
54	45.8	49.2	73.8#









#9
Naphthalene
Concen: 1.642 ng
RT: 10.585 min Scan# 758
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 344311
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.1 10.6 15.8



#10
Hexachlorobutadiene
Concen: 1.746 ng
RT: 10.890 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Tgt Ion:225 Resp: 100947
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 51.1 76.7





#11

2-Methylnaphthalene-d10

Concen: 1.718 ng

RT: 12.124 min Scan# 969

Delta R.T. 0.000 min

Lab File: BN017696.D

Acq: 03 Dec 2021 17:19

Instrument :

BNA_N

ClientSampleId :





SSTDICC1.6

Tgt Ion:152 Resp: 249354

Ion Ratio Lower Upper

152 100

151 20.3 16.2 24.4



#12

2-Methylnaphthalene

Concen: 1.683 ng

RT: 12.200 min Scan# 986

Delta R.T. 0.000 min

Lab File: BN017696.D

Acq: 03 Dec 2021 17:19




Tgt Ion:142 Resp: 232770


Ion Ratio Lower Upper

142 100

141 87.2 70.6 106.0

115 35.9 25.8 38.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.372 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 49459
Ion Ratio Lower Upper
164 100
162 102.2 81.4 122.2
160 46.3 36.8 55.2



#14
2,4,6-Tribromophenol
Concen: 1.784 ng
RT: 15.869 min Scan# 1320
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Tgt Ion:330 Resp: 44917
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 27.5 28.9 43.3#









#15
2-Fluorobiphenyl
Concen: 1.656 ng
RT: 13.001 min Scan# 1094
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 318915
Ion Ratio Lower Upper
172 100
171 34.2 28.7 43.1
170 22.5 19.4 29.2



#16
Acenaphthylene
Concen: 1.763 ng
RT: 14.092 min Scan# 1180
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Tgt Ion:152 Resp: 340086
Ion Ratio Lower Upper
152 100
151 19.5 15.5 23.3
153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 1.652 ng
RT: 14.435 min Scan# 1207
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 220340
Ion Ratio Lower Upper
154 100
153 112.1 91.2 136.8
152 54.4 44.3 66.5



#18
Fluorene
Concen: 1.716 ng
RT: 15.425 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Tgt Ion:166 Resp: 278174
Ion Ratio Lower Upper
166 100
165 97.1 77.9 116.9
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.116 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 99903
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.6 6.1 9.1



#20
4,6-Dinitro-2-methylphenol
Concen: 1.292 ng
RT: 15.501 min Scan# 1291
Delta R.T. -0.012 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Tgt Ion:198 Resp: 7954
Ion Ratio Lower Upper
198 100
182 10.2 11.4 17.2#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.750 ng
RT: 16.313 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 107397
Ion Ratio Lower Upper
248 100
250 96.5 71.8 107.8
141 46.3 70.9 106.3#



#22
Hexachlorobenzene
Concen: 1.678 ng
RT: 16.435 min Scan# 1366
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Tgt Ion:284 Resp: 124886
Ion Ratio Lower Upper
284 100
142 26.4 28.2 42.2#
249 30.4 24.4 36.6









#23
 Atrazine
 Concen: 2.172 ng
 RT: 16.581 min Scan# 1378
 Delta R.T. 0.000 min
 Lab File: BN017696.D
 Acq: 03 Dec 2021 17:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:200 Resp: 75023
 Ion Ratio Lower Upper
 200 100
 173 25.0 17.6 26.4
 215 48.8 40.2 60.4



#24
 Pentachlorophenol
 Concen: 1.549 ng
 RT: 16.776 min Scan# 1394
 Delta R.T. 0.000 min
 Lab File: BN017696.D
 Acq: 03 Dec 2021 17:19

Tgt Ion:266 Resp: 38690
 Ion Ratio Lower Upper
 266 100
 264 63.5 49.8 74.8
 268 64.3 51.3 76.9









#25
Phenanthrene
Concen: 1.663 ng
RT: 17.153 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 455689
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.6
179 15.3 12.3 18.5



#26
Anthracene
Concen: 1.767 ng
RT: 17.250 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Tgt Ion:178 Resp: 415933
Ion Ratio Lower Upper
178 100
176 18.1 14.6 22.0
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 1.854 ng
RT: 19.149 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 578973
Ion Ratio Lower Upper
212 100
106 9.7 10.5 15.7#
104 5.4 5.8 8.6#



#28
Fluoranthene
Concen: 1.797 ng
RT: 19.178 min Scan# 1730
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Tgt Ion:202 Resp: 557303
Ion Ratio Lower Upper
202 100
101 8.5 8.9 13.3#
203 17.5 13.8 20.6









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.303 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 107539
Ion Ratio Lower Upper
240 100
120 7.6 4.2 6.2#
236 26.3 20.0 30.0



#30
Pyrene
Concen: 1.627 ng
RT: 19.541 min Scan# 1803
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Tgt Ion:202 Resp: 564941
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.4 14.0 21.0









#31
 Terphenyl-d14
 Concen: 1.670 ng
 RT: 19.754 min Scan# 1846
 Delta R.T. 0.000 min
 Lab File: BN017696.D
 Acq: 03 Dec 2021 17:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244 Resp: 393773
 Ion Ratio Lower Upper
 244 100
 212 7.4 5.7 8.5
 122 8.1 8.6 13.0#



#32
 Benzo(a)anthracene
 Concen: 1.702 ng
 RT: 21.293 min Scan# 2018
 Delta R.T. 0.000 min
 Lab File: BN017696.D
 Acq: 03 Dec 2021 17:19

Tgt Ion:228 Resp: 555863
 Ion Ratio Lower Upper
 228 100
 226 25.7 21.3 31.9
 229 20.0 15.7 23.5









#33
Chrysene
Concen: 1.619 ng
RT: 21.346 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 575150
Ion Ratio Lower Upper
228 100
226 28.4 23.3 34.9
229 20.1 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 2.511 ng
RT: 21.239 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19


Tgt Ion:149 Resp: 194540
Ion Ratio Lower Upper
149 100
167 29.7 21.8 32.6
279 4.7 3.4 5.0








#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.612 min Scan# 24
 Delta R.T. -0.003 min
 Lab File: BN017696.D
 Acq: 03 Dec 2021 17:19

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6







Tgt Ion:264 Resp: 85659
 Ion Ratio Lower Upper
 264 100
 260 24.8 19.4 29.2
 265 24.3 18.6 28.0



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 1.336 ng
 RT: 25.978 min Scan# 3166
 Delta R.T. -0.003 min
 Lab File: BN017696.D
 Acq: 03 Dec 2021 17:19

Tgt Ion:276 Resp: 371747
 Ion Ratio Lower Upper
 276 100
 138 17.9 19.7 29.5#
 277 24.9 20.0 30.0





#37

Benzo(b)fluoranthene

Concen: 1.752 ng

RT: 22.914 min Scan# 2271

Delta R.T. 0.000 min

Lab File: BN017696.D

Acq: 03 Dec 2021 17:19

Instrument :

BN017696.D

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 497745

Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5

125 7.0 7.8 11.8#



#38

Benzo(k)fluoranthene

Concen: 1.680 ng

RT: 22.959 min Scan# 2284

Delta R.T. 0.000 min

Lab File: BN017696.D

Acq: 03 Dec 2021 17:19




Tgt Ion:252 Resp: 489306


Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5


125 6.9 8.2 12.2#








#39
Benzo(a)pyrene
Concen: 1.687 ng
RT: 23.510 min Scan# 2446
Delta R.T. -0.003 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:252 Resp: 428965
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 7.7 9.2 13.8#



#40
Dibenzo(a,h)anthracene
Concen: 1.240 ng
RT: 25.995 min Scan# 3171
Delta R.T. -0.007 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19


Tgt Ion:278 Resp: 281888
Ion Ratio Lower Upper
278 100
139 12.4 13.0 19.6#
279 23.4 19.2 28.8





#41
Benzo(g,h,i)perylene
Concen: 1.441 ng
RT: 26.700 min Scan# 31
Delta R.T. -0.003 min
Lab File: BN017696.D
Acq: 03 Dec 2021 17:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:276 Resp: 337037
Ion Ratio Lower Upper
276 100
277 23.9 19.0 28.4
138 15.6 17.1 25.7#

