

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
 Data File : BN017700.D
 Acq On : 03 Dec 2021 20:20
 Operator : CG/JU
 Sample : PB141149BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB141149BL

Quant Time: Dec 03 23:15:56 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 03 23:07:38 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	25781	0.400	ng	0.00
7) Naphthalene-d8	10.534	136	105088	0.400	ng	# 0.00
13) Acenaphthene-d10	14.371	164	62654	0.400	ng	0.00
19) Phenanthrene-d10	17.116	188	121467	0.400	ng	0.00
29) Chrysene-d12	21.303	240	108276	0.400	ng	# 0.00
35) Perylene-d12	23.612	264	81448	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	17661	0.291	ng	0.03
5) Phenol-d6	6.943	99	15356	0.248	ng	0.02
8) Nitrobenzene-d5	8.899	82	15289	0.301	ng	0.00
11) 2-Methylnaphthalene-d10	12.128	152	48682	0.262	ng	0.00
14) 2,4,6-Tribromophenol	15.881	330	4014	0.235	ng	0.01
15) 2-Fluorobiphenyl	13.001	172	70253	0.294	ng	0.00
27) Fluoranthene-d10	19.153	212	97976	0.247	ng	0.00
31) Terphenyl-d14	19.754	244	83927	0.349	ng	0.00
Target Compounds						
3) n-Nitrosodimethylamine	3.614	42	24	0.001	ng	# 49
6) bis(2-Chloroethyl)ether	7.182	93	30	0.000	ng	# 35
9) Naphthalene	10.585	128	1154	0.004	ng	# 80
10) Hexachlorobutadiene	10.889	225	16	0.000	ng	# 98
12) 2-Methylnaphthalene	12.199	142	270	0.002	ng	# 82
16) Acenaphthylene	14.092	152	267	0.001	ng	# 91
17) Acenaphthene	14.435	154	130	0.001	ng	92
18) Fluorene	15.437	166	153	0.001	ng	95
20) 4,6-Dinitro-2-methylph...	15.526	198	6	0.237	ng	# 1
21) 4-Bromophenyl-phenylether	16.325	248	9	0.000	ng	# 71
22) Hexachlorobenzene	16.434	284	10	0.000	ng	# 70
23) Atrazine	16.593	200	7	0.097	ng	# 7
24) Pentachlorophenol	16.897	266	41	0.284	ng	# 37
25) Phenanthrene	17.165	178	359	0.001	ng	# 81
26) Anthracene	17.250	178	145	0.001	ng	# 29
28) Fluoranthene	19.178	202	407	0.001	ng	# 65
30) Pyrene	19.545	202	294	0.001	ng	# 94
32) Benzo(a)anthracene	21.303	228	829	0.003	ng	# 81
33) Chrysene	21.345	228	911	0.003	ng	# 84
34) Bis(2-ethylhexyl)phtha...	21.239	149	823	0.008	ng	# 99
36) Indeno(1,2,3-cd)pyrene	26.015	276	318	0.002	ng	# 53
37) Benzo(b)fluoranthene	22.938	252	2248	0.008	ng	# 1
38) Benzo(k)fluoranthene	22.938	252	1391	0.005	ng	# 1
39) Benzo(a)pyrene	23.540	252	1283	0.005	ng	# 1
40) Dibenzo(a,h)anthracene	26.005	278	13	0.082	ng	# 1
41) Benzo(g,h,i)perylene	26.720	276	172	0.001	ng	# 18


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
 Data File : BN017700.D
 Acq On : 03 Dec 2021 20:20
 Operator : CG/JU
 Sample : PB141149BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB141149BL

Quant Time: Dec 03 23:15:56 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 03 23:07:38 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.763 min Scan# 51
Delta R.T. 0.009 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20




Instrument :
BNA_N
ClientSampleId :
PB141149BL


Tgt Ion:152 Resp: 25781
Ion Ratio Lower Upper
152 100
150 151.7 123.0 184.4
115 56.6 45.6 68.4



#3
n-Nitrosodimethylamine
Concen: 0.001 ng
RT: 3.614 min Scan# 69
Delta R.T. -0.018 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Tgt Ion: 42 Resp: 24
Ion Ratio Lower Upper
42 100
74 62.5 97.8 146.6#
44 0.0 5.5 8.3#









#4
2-Fluorophenol
Concen: 0.291 ng
RT: 5.384 min Scan# 21
Delta R.T. 0.028 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20




Instrument :
BNA_N
ClientSampleId :
PB141149BL


Tgt Ion:112 Resp: 17661
Ion Ratio Lower Upper
112 100
64 62.5 47.6 71.4
63 34.2 24.6 37.0



#5
Phenol-d6
Concen: 0.248 ng
RT: 6.943 min Scan# 430
Delta R.T. 0.018 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Tgt Ion: 99 Resp: 15356
Ion Ratio Lower Upper
99 100
42 20.9 17.0 25.6
71 38.6 24.8 37.2#









#6
bis(2-Chloroethyl)ether
Concen: 0.000 ng
RT: 7.182 min Scan# 456
Delta R.T. -0.000 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20




Instrument :
BNA_N
ClientSampleId :
PB141149BL


Tgt Ion: 93 Resp: 30
Ion Ratio Lower Upper
93 100
63 143.3 65.8 98.6#
95 0.0 26.6 39.8#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.534 min Scan# 754
Delta R.T. -0.000 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Tgt Ion:136 Resp: 105088
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 8.3 4.0 6.0#
68 6.4 3.0 4.6#









#8
 Nitrobenzene-d5
 Concen: 0.301 ng
 RT: 8.899 min Scan# 61
 Delta R.T. -0.001 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20




Instrument :
 BNA_N
 ClientSampleId :
 PB141149BL


Tgt Ion: 82 Resp: 15289
 Ion Ratio Lower Upper
 82 100
 128 35.2 27.7 41.5
 54 50.8 49.2 73.8



#9
 Naphthalene
 Concen: 0.004 ng
 RT: 10.585 min Scan# 758
 Delta R.T. -0.000 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20

Tgt Ion: 128 Resp: 1154
 Ion Ratio Lower Upper
 128 100
 129 19.3 8.8 13.2#
 127 20.6 10.6 15.8#









#10
Hexachlorobutadiene
Concen: 0.000 ng
RT: 10.889 min Scan# 71
Delta R.T. -0.000 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20




Instrument :
BNA_N
ClientSampleId :
PB141149BL


Tgt Ion:225 Resp: 16
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.5 51.1 76.7



#11
2-Methylnaphthalene-d10
Concen: 0.262 ng
RT: 12.128 min Scan# 970
Delta R.T. 0.004 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Tgt Ion:152 Resp: 48682
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.4





#12

2-Methylnaphthalene

Concen: 0.002 ng

RT: 12.199 min Scan# 986

Delta R.T. -0.000 min

Lab File: BN017700.D

Acq: 03 Dec 2021 20:20

Instrument :

BNA_N

ClientSampleId :

PB141149BL





Tgt Ion:142 Resp: 270

Ion Ratio Lower Upper

142 100

141 84.3 70.6 106.0

115 63.7 25.8 38.6#



#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.371 min Scan# 1202

Delta R.T. -0.000 min

Lab File: BN017700.D

Acq: 03 Dec 2021 20:20




Tgt Ion:164 Resp: 62654


Ion Ratio Lower Upper

164 100

162 101.8 81.4 122.2

160 46.5 36.8 55.2









#14
2,4,6-Tribromophenol
Concen: 0.235 ng
RT: 15.881 min Scan# 11
Delta R.T. 0.013 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20




Instrument :
BNA_N
ClientSampleId :
PB141149BL


Tgt Ion:330 Resp: 4014
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 28.4 28.9 43.3#



#15
2-Fluorobiphenyl
Concen: 0.294 ng
RT: 13.001 min Scan# 1094
Delta R.T. -0.000 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Tgt Ion:172 Resp: 70253
Ion Ratio Lower Upper
172 100
171 35.2 28.7 43.1
170 23.5 19.4 29.2









#16
Acenaphthylene
Concen: 0.001 ng
RT: 14.092 min Scan# 1180
Delta R.T. -0.000 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20




Instrument :
BNA_N
ClientSampleId :
PB141149BL


Tgt Ion:152 Resp: 267
Ion Ratio Lower Upper
152 100
151 24.3 15.5 23.3#
153 15.0 10.4 15.6



#17
Acenaphthene
Concen: 0.001 ng
RT: 14.435 min Scan# 1207
Delta R.T. -0.000 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Tgt Ion:154 Resp: 130
Ion Ratio Lower Upper
154 100
153 106.9 91.2 136.8
152 62.3 44.3 66.5









#18
Fluorene
Concen: 0.001 ng
RT: 15.437 min Scan# 1286
Delta R.T. 0.013 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20




Instrument :
BNA_N
ClientSampleId :
PB141149BL


Tgt Ion:166 Resp: 153
Ion Ratio Lower Upper
166 100
165 102.0 77.9 116.9
167 11.8 10.9 16.3



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.116 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Tgt Ion:188 Resp: 121467
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.1 6.1 9.1





#20

4,6-Dinitro-2-methylphenol

Concen: 0.237 ng

RT: 15.526 min Scan# 1293

Delta R.T. 0.013 min

Lab File: BN017700.D

Acq: 03 Dec 2021 20:20

Instrument :

BNA_N

ClientSampleId :

PB141149BL





Tgt Ion:198 Resp: 6

Ion Ratio Lower Upper

198 100

182 98.0 11.4 17.2#

77 0.0 0.0 0.0



#21

4-Bromophenyl-phenylether

Concen: 0.000 ng

RT: 16.325 min Scan# 1357

Delta R.T. 0.012 min

Lab File: BN017700.D

Acq: 03 Dec 2021 20:20




Tgt Ion:248 Resp: 9


Ion Ratio Lower Upper

248 100

250 118.2 71.8 107.8#


141 114.5 70.9 106.3#








#22
Hexachlorobenzene
Concen: 0.000 ng
RT: 16.434 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20


Instrument :
BNA_N
ClientSampleId :
PB141149BL





Tgt Ion:284 Resp: 10
Ion Ratio Lower Upper
284 100
142 40.0 28.2 42.2
249 0.0 24.4 36.6#




#23
Atrazine
Concen: 0.097 ng
RT: 16.593 min Scan# 1379
Delta R.T. 0.013 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20



Tgt Ion:200 Resp: 7
Ion Ratio Lower Upper
200 100
173 96.0 17.6 26.4#
215 96.0 40.2 60.4#









#24
 Pentachlorophenol
 Concen: 0.284 ng
 RT: 16.897 min Scan# 1404
 Delta R.T. 0.122 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20




Instrument :
 BNA_N
 ClientSampleId :
 PB141149BL


Tgt Ion:266 Resp: 41
 Ion Ratio Lower Upper
 266 100
 264 0.0 49.8 74.8#
 268 100.0 51.3 76.9#



#25
 Phenanthrene
 Concen: 0.001 ng
 RT: 17.165 min Scan# 1426
 Delta R.T. 0.012 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20

Tgt Ion:178 Resp: 359
 Ion Ratio Lower Upper
 178 100
 176 17.0 15.0 22.6
 179 0.0 12.3 18.5#









#26
 Anthracene
 Concen: 0.001 ng
 RT: 17.250 min Scan# 1433
 Delta R.T. -0.000 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20




Instrument :
 BNA_N
 ClientSampleId :
 PB141149BL


Tgt Ion:178 Resp: 145
 Ion Ratio Lower Upper
 178 100
 176 62.8 14.6 22.0#
 179 0.0 12.2 18.4#



#27
 Fluoranthene-d10
 Concen: 0.247 ng
 RT: 19.153 min Scan# 1725
 Delta R.T. 0.005 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20

Tgt Ion:212 Resp: 97976
 Ion Ratio Lower Upper
 212 100
 106 9.7 10.5 15.7#
 104 5.3 5.8 8.6#









#28
Fluoranthene
Concen: 0.001 ng
RT: 19.178 min Scan# 1730
Delta R.T. -0.000 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20




Instrument :
BNA_N
ClientSampleId :
PB141149BL


Tgt Ion:202 Resp: 407
Ion Ratio Lower Upper
202 100
101 0.0 8.9 13.3#
203 0.0 13.8 20.6#



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.303 min Scan# 2019
Delta R.T. -0.000 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Tgt Ion:240 Resp: 108276
Ion Ratio Lower Upper
240 100
120 10.2 4.2 6.2#
236 26.9 20.0 30.0









#30
 Pyrene
 Concen: 0.001 ng
 RT: 19.545 min Scan# 1803
 Delta R.T. 0.005 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20




Instrument :
 BNA_N
 ClientSampleId :
 PB141149BL


Tgt Ion:202 Resp: 294
 Ion Ratio Lower Upper
 202 100
 200 19.0 16.5 24.7
 203 21.8 14.0 21.0#



#31
 Terphenyl-d14
 Concen: 0.349 ng
 RT: 19.754 min Scan# 1846
 Delta R.T. -0.000 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20

Tgt Ion:244 Resp: 83927
 Ion Ratio Lower Upper
 244 100
 212 7.6 5.7 8.5
 122 9.5 8.6 13.0









#32
Benzo(a)anthracene
Concen: 0.003 ng
RT: 21.303 min Scan# 2019
Delta R.T. 0.011 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20




Instrument :
BNA_N
ClientSampleId :
PB141149BL


Tgt Ion:228 Resp: 829
Ion Ratio Lower Upper
228 100
226 32.6 21.3 31.9#
229 32.8 15.7 23.5#



#33
Chrysene
Concen: 0.003 ng
RT: 21.345 min Scan# 2023
Delta R.T. -0.000 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Tgt Ion:228 Resp: 911
Ion Ratio Lower Upper
228 100
226 35.6 23.3 34.9#
229 29.7 15.6 23.4#









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.008 ng
 RT: 21.239 min Scan# 2013
 Delta R.T. 0.000 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20




Instrument :
 BNA_N
 ClientSampleId :
 PB141149BL


Tgt Ion:149 Resp: 823
 Ion Ratio Lower Upper
 149 100
 167 27.1 21.8 32.6
 279 6.3 3.4 5.0#



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.612 min Scan# 2475
 Delta R.T. -0.003 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20

Tgt Ion:264 Resp: 81448
 Ion Ratio Lower Upper
 264 100
 260 24.8 19.4 29.2
 265 24.5 18.6 28.0









#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.002 ng
 RT: 26.015 min Scan# 3167
 Delta R.T. 0.034 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20




Instrument :
 BNA_N
 ClientSampleId :
 PB141149BL


Tgt Ion: 276 Resp: 318
 Ion Ratio Lower Upper
 276 100
 138 0.0 19.7 29.5#
 277 2.5 20.0 30.0#



#37
 Benzo(b)fluoranthene
 Concen: 0.008 ng
 RT: 22.938 min Scan# 2278
 Delta R.T. 0.024 min
 Lab File: BN017700.D
 Acq: 03 Dec 2021 20:20

Tgt Ion: 252 Resp: 2248
 Ion Ratio Lower Upper
 252 100
 253 130.9 17.7 26.5#
 125 408.8 7.8 11.8#









#38
Benzo(k)fluoranthene
Concen: 0.005 ng
RT: 22.938 min Scan# 21
Delta R.T. -0.021 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20




Instrument :
BNA_N
ClientSampleId :
PB141149BL


Tgt Ion:252 Resp: 1391
Ion Ratio Lower Upper
252 100
253 130.9 17.7 26.5#
125 408.8 8.2 12.2#



#39
Benzo(a)pyrene
Concen: 0.005 ng
RT: 23.540 min Scan# 2454
Delta R.T. 0.027 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Tgt Ion:252 Resp: 1283
Ion Ratio Lower Upper
252 100
253 172.1 17.9 26.9#
125 637.7 9.2 13.8#









#40
Dibenzo(a,h)anthracene
Concen: 0.082 ng
RT: 26.005 min Scan# 3173
Delta R.T. 0.004 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Instrument :
BNA_N
ClientSampleId :
PB141149BL

Tgt Ion: 278 Resp: 13
Ion Ratio Lower Upper
278 100
139 489.0 13.0 19.6#
279 141.0 19.2 28.8#



#41
Benzo(g,h,i)perylene
Concen: 0.001 ng
RT: 26.720 min Scan# 3383
Delta R.T. 0.017 min
Lab File: BN017700.D
Acq: 03 Dec 2021 20:20

Tgt Ion: 276 Resp: 172
Ion Ratio Lower Upper
276 100
277 59.9 19.0 28.4#
138 63.9 17.1 25.7#

