

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
 Data File : BN017703.D
 Acq On : 03 Dec 2021 22:44
 Operator : CG/JU
 Sample : M4917-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-14

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/06/2021
 Supervised By :Sohil Jodhani 12/10/2021

Quant Time: Dec 04 01:01:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 03 23:07:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.754	152	22195	0.40	ng	0.00
7) Naphthalene-d8	10.535	136	88034	0.40	ng	# 0.00
13) Acenaphthene-d10	14.372	164	58600	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	121440	0.40	ng	0.00
29) Chrysene-d12	21.303	240	113587	0.40	ng	# 0.00
35) Perylene-d12	23.609	264	96188	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.366	112	6620	0.13	ng	0.00
5) Phenol-d6	6.934	99	5242	0.10	ng	0.00
8) Nitrobenzene-d5	8.900	82	11200	0.26	ng	0.00
11) 2-Methylnaphthalene-d10	12.124	152	34375	0.22	ng	0.00
14) 2,4,6-Tribromophenol	15.869	330	10232	0.41	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	64505	0.29	ng	0.00
27) Fluoranthene-d10	19.149	212	123271	0.31	ng	0.00
31) Terphenyl-d14	19.749	244	129563	0.51	ng	0.00
Target Compounds						
24) Pentachlorophenol	16.776	266	1022	0.31	ng	Qvalue 96
34) Bis(2-ethylhexyl)phtha...	21.229	149	13674	0.13	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
Data File : BN017703.D
Acq On : 03 Dec 2021 22:44
Operator : CG/JU
Sample : M4917-01
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-14

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/06/2021
Supervised By :Sohil Jodhani 12/10/2021

Quant Time: Dec 04 01:01:09 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 03 23:07:38 2021
Response via : Initial Calibration

