

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
 Data File : BN017713.D
 Acq On : 04 Dec 2021 05:55
 Operator : CG/JU
 Sample : M4917-10
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-26

Quant Time: Dec 06 02:13:16 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 03 23:07:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.754	152	20148	0.40	ng	0.00
7) Naphthalene-d8	10.535	136	83080	0.40	ng	# 0.00
13) Acenaphthene-d10	14.371	164	54632	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	115771	0.40	ng	0.00
29) Chrysene-d12	21.303	240	118603	0.40	ng	# 0.00
35) Perylene-d12	23.609	264	116851	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.375	112	8198	0.17	ng	0.02
5) Phenol-d6	6.934	99	5167	0.11	ng	0.00
8) Nitrobenzene-d5	8.900	82	18386	0.46	ng	0.00
11) 2-Methylnaphthalene-d10	12.124	152	41795	0.28	ng	0.00
14) 2,4,6-Tribromophenol	15.868	330	13436	0.52	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	74318	0.36	ng	0.00
27) Fluoranthene-d10	19.148	212	119492	0.32	ng	0.00
31) Terphenyl-d14	19.754	244	112952	0.43	ng	0.00
Target Compounds						Qvalue
9) Naphthalene	10.585	128	7703	0.04	ng	# 89
17) Acenaphthene	14.435	154	11493	0.08	ng	99
18) Fluorene	15.424	166	3652	0.02	ng	# 93
24) Pentachlorophenol	16.775	266	698	0.30	ng	95
25) Phenanthrene	17.153	178	14154	0.05	ng	# 86
34) Bis(2-ethylhexyl)phtha...	21.229	149	13366	0.12	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
Data File : BN017713.D
Acq On : 04 Dec 2021 05:55
Operator : CG/JU
Sample : M4917-10
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-26

Quant Time: Dec 06 02:13:16 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 03 23:07:38 2021
Response via : Initial Calibration

