

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN120321\  
 Data File : BN017719.D  
 Acq On : 04 Dec 2021 09:29  
 Operator : CG/JU  
 Sample : M4917-16  
 Misc :  
 ALS Vial : 29 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 MW-32

Manual Integrations  
 APPROVED

Reviewed By :Jagrut Upadhyay 12/06/2021  
 Supervised By :Sohil Jodhani 12/10/2021

Quant Time: Dec 06 02:14:03 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN120321.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Dec 03 23:07:38 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|------|-------|----------|
| -----                         |        |      |          |      |       |          |
| Internal Standards            |        |      |          |      |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.754  | 152  | 21002    | 0.40 | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.535 | 136  | 87816    | 0.40 | ng    | # 0.00   |
| 13) Acenaphthene-d10          | 14.372 | 164  | 56078    | 0.40 | ng    | # 0.00   |
| 19) Phenanthrene-d10          | 17.116 | 188  | 119450   | 0.40 | ng    | 0.00     |
| 29) Chrysene-d12              | 21.303 | 240  | 123873   | 0.40 | ng    | # 0.00   |
| 35) Perylene-d12              | 23.612 | 264  | 115096   | 0.40 | ng    | 0.00     |
| System Monitoring Compounds   |        |      |          |      |       |          |
| 4) 2-Fluorophenol             | 5.375  | 112  | 6078     | 0.12 | ng    | 0.02     |
| 5) Phenol-d6                  | 6.943  | 99   | 4127     | 0.08 | ng    | 0.02     |
| 8) Nitrobenzene-d5            | 8.900  | 82   | 17566m   | 0.41 | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10   | 12.124 | 152  | 38461    | 0.25 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 15.869 | 330  | 15155    | 0.56 | ng    | 0.00     |
| 15) 2-Fluorobiphenyl          | 13.001 | 172  | 71641    | 0.34 | ng    | 0.00     |
| 27) Fluoranthene-d10          | 19.149 | 212  | 131901   | 0.34 | ng    | 0.00     |
| 31) Terphenyl-d14             | 19.754 | 244  | 135468   | 0.49 | ng    | 0.00     |
| Target Compounds              |        |      |          |      |       |          |
|                               |        |      |          |      |       | Qvalue   |
| 9) Naphthalene                | 10.585 | 128  | 784709   | 3.50 | ng    | 100      |
| 12) 2-Methylnaphthalene       | 12.200 | 142  | 34805    | 0.24 | ng    | # 94     |
| 17) Acenaphthene              | 14.435 | 154  | 8776     | 0.06 | ng    | # 78     |
| 18) Fluorene                  | 15.425 | 166  | 15992    | 0.09 | ng    | # 89     |
| 24) Pentachlorophenol         | 16.776 | 266  | 208      | 0.29 | ng    | 91       |
| 25) Phenanthrene              | 17.153 | 178  | 37386m   | 0.12 | ng    |          |
| 26) Anthracene                | 17.250 | 178  | 5861     | 0.02 | ng    | # 83     |
| 28) Fluoranthene              | 19.178 | 202  | 16121    | 0.04 | ng    | # 96     |
| 30) Pyrene                    | 19.541 | 202  | 15431    | 0.04 | ng    | # 88     |
| 34) Bis(2-ethylhexyl)phtha... | 21.229 | 149  | 26903    | 0.23 | ng    | # 95     |
| -----                         |        |      |          |      |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN120321\  
Data File : BN017719.D  
Acq On : 04 Dec 2021 09:29  
Operator : CG/JU  
Sample : M4917-16  
Misc :  
ALS Vial : 29 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
MW-32

Manual Integrations  
APPROVED

Reviewed By :Jagrut Upadhyay 12/06/2021  
Supervised By :Sohil Jodhani 12/10/2021

Quant Time: Dec 06 02:14:03 2021  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN120321.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Dec 03 23:07:38 2021  
Response via : Initial Calibration

