

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
 Data File : BN017699.D
 Acq On : 03 Dec 2021 19:08
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN120321

Quant Time: Dec 03 23:11:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 03 23:07:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.297	0.290	2.4	103	0.00
3	n-Nitrosodimethylamine	0.401	0.401	0.0	109	0.00
4 S	2-Fluorophenol	0.942	0.930	1.3	106	0.00
5 S	Phenol-d6	0.962	0.947	1.6	108	0.00
6	bis(2-Chloroethyl)ether	1.025	1.017	0.8	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.193	0.205	-6.2	121	0.00
9	Naphthalene	1.020	1.001	1.9	105	0.00
10	Hexachlorobutadiene	0.294	0.293	0.3	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.708	0.708	0.0	107	0.00
12	2-Methylnaphthalene	0.666	0.678	-1.8	107	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.203	0.169	16.7	114	0.00
15 S	2-Fluorobiphenyl	1.524	1.509	1.0	107	0.00
16	Acenaphthylene	1.539	1.458	5.3	110	0.00
17	Acenaphthene	1.049	1.019	2.9	106	0.00
18	Fluorene	1.302	1.290	0.9	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4,6-Dinitro-2-methylphenol	0.018	0.008	55.6#	132	0.00
21	4-Bromophenyl-phenylether	0.241	0.232	3.7	110	0.00
22	Hexachlorobenzene	0.297	0.296	0.3	109	0.00
23	Atrazine	0.138	0.138	0.0	120	0.00
24	Pentachlorophenol	0.087	0.057	34.5#	119	0.00
25	Phenanthrene	1.063	1.044	1.8	109	0.00
26	Anthracene	0.916	0.873	4.7	111	0.00
27 SURR	Fluoranthene-d10	1.306	1.263	3.3	111	0.00
28	Fluoranthene	1.264	1.271	-0.6	110	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
30	Pyrene	1.274	1.260	1.1	111	0.00
31 S	Terphenyl-d14	0.888	0.887	0.1	112	0.00
32	Benzo(a)anthracene	1.169	1.117	4.4	115	0.00
33	Chrysene	1.308	1.286	1.7	108	0.00
34	Bis(2-ethylhexyl)phthalate	0.380	0.383	-0.8	126	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	112	0.00
36	Indeno(1,2,3-cd)pyrene	0.972	0.866	10.9	110	0.00
37	Benzo(b)fluoranthene	1.316	1.291	1.9	114	0.00
38	Benzo(k)fluoranthene	1.290	1.268	1.7	110	0.00
39 C	Benzo(a)pyrene	1.152	1.106	4.0	114	0.00
40	Dibenzo(a,h)anthracene	0.726	0.628	13.5	110	0.00
41	Benzo(g,h,i)perylene	0.901	0.846	6.1	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN120321