

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
 Data File : BN017698.D
 Acq On : 03 Dec 2021 18:31
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Dec 03 22:25:59 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 03 17:44:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.754	152	23831	0.40	ng	0.00
7) Naphthalene-d8	10.535	136	88148	0.40	ng	# 0.00
13) Acenaphthene-d10	14.371	164	52621	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	109389	0.40	ng	0.00
29) Chrysene-d12	21.303	240	112697	0.40	ng	# 0.00
35) Perylene-d12	23.612	264	103213	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.357	112	265238	4.90	ng	0.00
5) Phenol-d6	6.924	99	288448	5.04	ng	0.00
8) Nitrobenzene-d5	0.000	82	0d	0.00	ng	
11) 2-Methylnaphthalene-d10	12.124	152	747177	4.86	ng	0.00
14) 2,4,6-Tribromophenol	15.868	330	162785	6.08	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	965564	4.71	ng	0.00
27) Fluoranthene-d10	19.148	212	1770645	5.18	ng	0.00
31) Terphenyl-d14	19.754	244	1216436	4.92	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.319	88	79858	4.99	ng	96
3) n-Nitrosodimethylamine	3.633	42	116032	5.19	ng	100
6) bis(2-Chloroethyl)ether	7.182	93	288530	4.35	ng	90
9) Naphthalene	10.585	128	1032351	4.64	ng	100
10) Hexachlorobutadiene	10.889	225	305981	4.99	ng	# 100
12) 2-Methylnaphthalene	12.200	142	680689	4.64	ng	97
16) Acenaphthylene	14.092	152	1075452	5.24	ng	100
17) Acenaphthene	14.435	154	675432	4.76	ng	96
18) Fluorene	15.424	166	843247	4.89	ng	99
20) 4,6-Dinitro-2-methylph...	15.501	198	57304	8.50	ng	# 82
21) 4-Bromophenyl-phenylether	16.313	248	342555	5.10	ng	# 72
22) Hexachlorobenzene	16.435	284	377428	4.63	ng	# 92
24) Pentachlorophenol	16.775	266	178245	6.52	ng	100
25) Phenanthrene	17.153	178	1383383	4.61	ng	100
26) Anthracene	17.250	178	1293932	5.02	ng	100
28) Fluoranthene	19.183	202	1632142	4.81	ng	# 97
30) Pyrene	19.541	202	1720311	4.73	ng	100
32) Benzo(a)anthracene	21.292	228	1772446	5.18	ng	99
33) Chrysene	21.345	228	1723002	4.63	ng	98
36) Indeno(1,2,3-cd)pyrene	25.977	276	1439256	4.29	ng	# 93
37) Benzo(b)fluoranthene	22.914	252	1711376	5.00	ng	# 97
38) Benzo(k)fluoranthene	22.958	252	1576834	4.49	ng	# 97
39) Benzo(a)pyrene	23.510	252	1502066	4.90	ng	# 96
40) Dibenzo(a,h)anthracene	25.998	278	1124317	4.10	ng	# 95
41) Benzo(g,h,i)perylene	26.700	276	1274798	4.52	ng	# 94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
Data File : BN017698.D
Acq On : 03 Dec 2021 18:31
Operator : CG/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Dec 03 22:25:59 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 03 17:44:19 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.754 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 23831
Ion Ratio Lower Upper
152 100
150 155.9 123.0 184.4
115 58.7 45.6 68.4



#2
1,4-Dioxane
Concen: 4.99 ng
RT: 3.319 min Scan# 37
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion: 88 Resp: 79858
Ion Ratio Lower Upper
88 100
43 32.3 24.7 37.1
58 74.5 63.0 94.6









#3
n-Nitrosodimethylamine
Concen: 5.19 ng
RT: 3.633 min Scan# 71
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion: 42 Resp: 116032




Ion	Ratio	Lower	Upper
42	100		
74	121.9	97.8	146.6
44	5.7	5.5	8.3




#4
2-Fluorophenol
Concen: 4.90 ng
RT: 5.357 min Scan# 258
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion: 112 Resp: 265238

Ion	Ratio	Lower	Upper
112	100		
64	62.1	47.6	71.4
63	33.8	24.6	37.0









#5
Phenol-d6
Concen: 5.04 ng
RT: 6.924 min Scan# 41
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 288448
Ion Ratio Lower Upper
99 100
42 21.4 17.0 25.6
71 38.2 24.8 37.2#



#6
bis(2-Chloroethyl)ether
Concen: 4.35 ng
RT: 7.182 min Scan# 456
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion: 93 Resp: 288530
Ion Ratio Lower Upper
93 100
63 71.3 65.8 98.6
95 30.2 26.6 39.8






#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.535 min Scan# 71
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Instrument :




BNA_N

ClientSampleId :

SSTDICC5.0




Tgt Ion:136 Resp: 88148
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 6.5 4.0 6.0#
68 5.2 3.0 4.6#




#8
Nitrobenzene-d5
Concen: 0.00 ng
Expected RT: 8.90 min

Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion: 82
Sig Exp Ratio
82 100
128 34.6
54 61.5









#9
Naphthalene
Concen: 4.64 ng
RT: 10.585 min Scan# 758
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 1032351
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.0 10.6 15.8



#10
Hexachlorobutadiene
Concen: 4.99 ng
RT: 10.889 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:225 Resp: 305981
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.1 76.7









#11
2-Methylnaphthalene-d10
Concen: 4.86 ng
RT: 12.124 min Scan# 969
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 747177
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.4



#12
2-Methylnaphthalene
Concen: 4.64 ng
RT: 12.200 min Scan# 986
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:142 Resp: 680689
Ion Ratio Lower Upper
142 100
141 87.6 70.6 106.0
115 36.9 25.8 38.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.371 min Scan# 1202
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 52621
Ion Ratio Lower Upper
164 100
162 100.9 81.4 122.2
160 45.0 36.8 55.2



#14
2,4,6-Tribromophenol
Concen: 6.08 ng
RT: 15.868 min Scan# 1320
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:330 Resp: 162785
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 27.5 28.9 43.3#









#15
2-Fluorobiphenyl
Concen: 4.71 ng
RT: 13.001 min Scan# 1094
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 965564
Ion Ratio Lower Upper
172 100
171 34.5 28.7 43.1
170 22.9 19.4 29.2



#16
Acenaphthylene
Concen: 5.24 ng
RT: 14.092 min Scan# 1180
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:152 Resp: 1075452
Ion Ratio Lower Upper
152 100
151 19.5 15.5 23.3
153 13.2 10.4 15.6









#17
Acenaphthene
Concen: 4.76 ng
RT: 14.435 min Scan# 1207
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 675432
Ion Ratio Lower Upper
154 100
153 109.8 91.2 136.8
152 52.4 44.3 66.5



#18
Fluorene
Concen: 4.89 ng
RT: 15.424 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:166 Resp: 843247
Ion Ratio Lower Upper
166 100
165 96.5 77.9 116.9
167 13.6 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.116 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188 Resp: 109389
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.4 6.1 9.1



#20
4,6-Dinitro-2-methylphenol
Concen: 8.50 ng
RT: 15.501 min Scan# 1291
Delta R.T. -0.013 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:198 Resp: 57304
Ion Ratio Lower Upper
198 100
182 6.8 11.4 17.2#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 5.10 ng
RT: 16.313 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:248 Resp: 342555
Ion Ratio Lower Upper
248 100
250 97.5 71.8 107.8
141 43.4 70.9 106.3#



#22
Hexachlorobenzene
Concen: 4.63 ng
RT: 16.435 min Scan# 1366
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:284 Resp: 377428
Ion Ratio Lower Upper
284 100
142 26.5 28.2 42.2#
249 30.6 24.4 36.6









#24
 Pentachlorophenol
 Concen: 6.52 ng
 RT: 16.775 min Scan# 1394
 Delta R.T. 0.000 min
 Lab File: BN017698.D
 Acq: 03 Dec 2021 18:31




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:266 Resp: 178245
 Ion Ratio Lower Upper
 266 100
 264 62.6 49.8 74.8
 268 64.0 51.3 76.9



#25
 Phenanthrene
 Concen: 4.61 ng
 RT: 17.153 min Scan# 1425
 Delta R.T. 0.000 min
 Lab File: BN017698.D
 Acq: 03 Dec 2021 18:31

Tgt Ion:178 Resp: 1383383
 Ion Ratio Lower Upper
 178 100
 176 18.6 15.0 22.6
 179 15.5 12.3 18.5









#26
 Anthracene
 Concen: 5.02 ng
 RT: 17.250 min Scan# 1433
 Delta R.T. 0.000 min
 Lab File: BN017698.D
 Acq: 03 Dec 2021 18:31

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:178 Resp: 1293932




Ion	Ratio	Lower	Upper
178	100		
176	18.3	14.6	22.0
179	15.6	12.2	18.4




#27
 Fluoranthene-d10
 Concen: 5.18 ng
 RT: 19.148 min Scan# 1724
 Delta R.T. 0.000 min
 Lab File: BN017698.D
 Acq: 03 Dec 2021 18:31

Tgt Ion:212 Resp: 1770645

Ion	Ratio	Lower	Upper
212	100		
106	9.7	10.5	15.7#
104	5.4	5.8	8.6#









#28
Fluoranthene
Concen: 4.81 ng
RT: 19.183 min Scan# 1
Delta R.T. 0.005 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 1632142
Ion Ratio Lower Upper
202 100
101 8.4 8.9 13.3#
203 17.6 13.8 20.6



#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.303 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:240 Resp: 112697
Ion Ratio Lower Upper
240 100
120 8.0 4.2 6.2#
236 26.7 20.0 30.0









#30
Pyrene
Concen: 4.73 ng
RT: 19.541 min Scan# 1803
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 1720311
Ion Ratio Lower Upper
202 100
200 20.5 16.5 24.7
203 17.8 14.0 21.0



#31
Terphenyl-d14
Concen: 4.92 ng
RT: 19.754 min Scan# 1846
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:244 Resp: 1216436
Ion Ratio Lower Upper
244 100
212 7.5 5.7 8.5
122 8.4 8.6 13.0#









#32
Benzo(a)anthracene
Concen: 5.18 ng
RT: 21.292 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 1772446
Ion Ratio Lower Upper
228 100
226 26.0 21.3 31.9
229 20.4 15.7 23.5



#33
Chrysene
Concen: 4.63 ng
RT: 21.345 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:228 Resp: 1723002
Ion Ratio Lower Upper
228 100
226 28.3 23.3 34.9
229 20.2 15.6 23.4









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.612 min Scan# 24
Delta R.T. -0.003 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:264 Resp: 103213
Ion Ratio Lower Upper
264 100
260 24.8 19.4 29.2
265 23.5 18.6 28.0



#36
Indeno(1,2,3-cd)pyrene
Concen: 4.29 ng
RT: 25.977 min Scan# 3166
Delta R.T. -0.003 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:276 Resp: 1439256
Ion Ratio Lower Upper
276 100
138 17.6 19.7 29.5#
277 25.1 20.0 30.0









#37
Benzo(b)fluoranthene
Concen: 5.00 ng
RT: 22.914 min Scan# 2271
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 1711376
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 6.7 7.8 11.8#



#38
Benzo(k)fluoranthene
Concen: 4.49 ng
RT: 22.958 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31

Tgt Ion:252 Resp: 1576834
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 6.8 8.2 12.2#









#39
Benzo(a)pyrene
Concen: 4.90 ng
RT: 23.510 min Scan# 2446
Delta R.T. -0.003 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 1502066
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 7.4 9.2 13.8#



#40
Dibenzo(a,h)anthracene
Concen: 4.10 ng
RT: 25.998 min Scan# 3172
Delta R.T. -0.003 min
Lab File: BN017698.D
Acq: 03 Dec 2021 18:31


Tgt Ion:278 Resp: 1124317
Ion Ratio Lower Upper
278 100
139 11.9 13.0 19.6#
279 23.6 19.2 28.8





#41
 Benzo(g,h,i)perylene
 Concen: 4.52 ng
 RT: 26.700 min Scan# 31
 Delta R.T. -0.003 min
 Lab File: BN017698.D
 Acq: 03 Dec 2021 18:31

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0



Tgt Ion:276 Resp: 1274798
 Ion Ratio Lower Upper
 276 100
 277 23.7 19.0 28.4
 138 15.2 17.1 25.7#

