


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
 Data File : BN017691.D
 Acq On : 03 Dec 2021 14:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120321.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Dec 03 23:07:38 2021






AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2470

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.8	256067	PASS
68	69	0.00	2	1.9	5079	PASS
69	198	0.00	100	35.1	266005	PASS
70	69	0.00	2	0.5	1458	PASS
127	198	10	80	38.1	288320	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	757461	PASS
199	198	5	9	6.5	48875	PASS
275	198	10	60	26.2	198357	PASS
365	198	1	100	3.5	26515	PASS
441	198	0.01	100	10.9	82240	PASS
442	442	50	100	100.0	555797	PASS
443	442	15	24	18.8	104563	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
 Data File : BN017691.D
 Acq On : 03 Dec 2021 14:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Dec 04 03:34:22 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 04 03:34:17 2021
 Response via : Initial Calibration



TIC: BN017691.D\data.ms

(70) Pentachlorophenol (C)

16.751min (0.000) 25988.28 ng




response 948569

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.17
264.00	61.60	63.32
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120321\
Data File : BN017691.D
Acq On : 03 Dec 2021 14:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 04 03:34:22 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 04 03:34:17 2021
Response via : Initial Calibration



TIC: BN017691.D\data.ms

(77) Benzidine

19.345min (0.000) 0.00 ng

response 5121037

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.18
183.00	13.20	11.67
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/3/2021	BNA_N	<u>BN017691.D</u>
Compound Name	Response	Retention Time
DDT	2010731	20.59
DDD	2688	20.15
DDE	580	19.66
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3268	2013999	0.16

Instrument :
BNA_N
ClientSampleId :
DFTPP