

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120419\
 Data File : BN008836.D
 Acq On : 04 Dec 2019 11:20
 Operator : JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDICC0.1

Quant Time: Dec 04 14:19:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 14:04:21 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	144	0.00
2	1,4-Dioxane	0.384	0.080	79.2#	30#	0.00
3	n-Nitrosodimethylamine	0.531	0.000	100.0#	0#	-3.51#
4 S	2-Fluorophenol	1.083	0.274	74.7#	37#	0.00
5 S	Phenol-d6	1.475	0.366	75.2#	37#	0.00
6	bis(2-Chloroethyl)ether	1.327	0.330	75.1#	37#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	141	0.00
8 S	Nitrobenzene-d5	0.353	0.089	74.8#	38#	0.00
9	Naphthalene	1.067	0.269	74.8#	37#	0.00
10	Hexachlorobutadiene	0.210	0.054	74.3#	37#	0.00
11 SURR	2-Methylnaphthalene-d10	0.664	0.166	75.0#	37#	0.00
12	2-Methylnaphthalene	0.786	0.197	74.9#	37#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.218	0.055	74.8#	38#	0.00
15 S	2-Fluorobiphenyl	1.499	0.380	74.6#	37#	0.01
16	Acenaphthylene	1.866	0.457	75.5#	37#	0.00
17	Acenaphthene	1.184	0.296	75.0#	36#	0.00
18	Fluorene	1.570	0.387	75.4#	36#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	139	0.00
20	4-Bromophenyl-phenylether	0.234	0.058	75.2#	37#	0.00
21	Hexachlorobenzene	0.270	0.068	74.8#	36#	0.00
22	Pentachlorophenol	0.108	0.000	100.0#	0#	-16.60#
23	Phenanthrene	1.221	0.304	75.1#	36#	0.00
24	Anthracene	1.121	0.271	75.8#	36#	0.00
25 SURR	Fluoranthene-d10	1.211	0.298	75.4#	36#	0.00
26	Fluoranthene	1.470	0.361	75.4#	36#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	140	0.00
28	Pvrene	1.396	0.351	74.9#	37#	0.00
29 S	Terphenyl-d14	0.921	0.231	74.9#	36#	0.00
30	Benzo(a)anthracene	1.446	0.367	74.6#	37#	0.00
31	Chrysene	1.403	0.346	75.3#	35#	0.00
32	Bis(2-ethylhexyl)phthalate	0.964	0.321	66.7#	53	0.00
33	Indeno(1,2,3-cd)pyrene	1.733	0.432	75.1#	36#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	136	0.00
35	Benzo(b)fluoranthene	1.329	0.338	74.6#	36#	0.00
36	Benzo(k)fluoranthene	1.305	0.334	74.4#	37#	0.00
37 C	Benzo(a)pyrene	1.253	0.325	74.1#	37#	0.00
38	Dibenzo(a,h)anthracene	1.280	0.325	74.6#	36#	0.00
39	Benzo(a,h,i)perylene	1.296	0.326	74.8#	36#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120419\
Data File : BN008836.D
Acq On : 04 Dec 2019 11:20
Operator : JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDICC0.1

Quant Time: Dec 04 14:19:56 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 14:04:21 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 1			