

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120419\
 Data File : BN008836.D
 Acq On : 04 Dec 2019 11:20
 Operator : JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDICC0.1

Quant Time: Dec 04 14:19:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 14:04:21 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	144	0.00
2	1,4-Dioxane	0.400	0.072	82.0#	30	0.00
3	n-Nitrosodimethylamine	0.400	0.000	100.0#	0	-3.51#
4 S	2-Fluorophenol	0.400	0.080	80.0#	37	0.00
5 S	Phenol-d6	0.400	0.075	81.3#	37	0.00
6	bis(2-Chloroethyl)ether	0.400	0.117	70.8#	37	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	141	0.00
8 S	Nitrobenzene-d5	0.400	0.102	74.5#	38	0.00
9	Naphthalene	0.400	0.096	76.0#	37	0.00
10	Hexachlorobutadiene	0.400	0.092	77.0#	37	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.098	75.5#	37	0.00
12	2-Methylnaphthalene	0.400	0.104	74.0#	37	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.400	0.101	74.8#	38	0.00
15 S	2-Fluorobiphenyl	0.400	0.092	77.0#	37	0.01
16	Acenaphthylene	0.400	0.086	78.5#	37	0.00
17	Acenaphthene	0.400	0.087	78.3#	36	0.00
18	Fluorene	0.400	0.088	78.0#	36	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	139	0.00
20	4-Bromophenyl-phenylether	0.400	0.096	76.0#	37	0.00
21	Hexachlorobenzene	0.400	0.102	74.5#	36	0.00
22	Pentachlorophenol	0.400	0.000	100.0#	0	-16.60#
23	Phenanthrene	0.400	0.098	75.5#	36	0.00
24	Anthracene	0.400	0.096	76.0#	36	0.00
25 SURR	Fluoranthene-d10	0.400	0.088	78.0#	36	0.00
26	Fluoranthene	0.400	0.093	76.8#	36	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	140	0.00
28	Pvrene	0.400	0.103	74.3#	37	0.00
29 S	Terphenyl-d14	0.400	0.094	76.5#	36	0.00
30	Benzo(a)anthracene	0.400	0.100	75.0#	37	0.00
31	Chrysene	0.400	0.096	76.0#	35	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.169	57.8#	53	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.096	76.0#	36	0.00
34 I	Perylene-d12	0.400	0.400	0.0	136	0.00
35	Benzo(b)fluoranthene	0.400	0.097	75.8#	36	0.00
36	Benzo(k)fluoranthene	0.400	0.097	75.8#	37	0.00
37 C	Benzo(a)pyrene	0.400	0.099	75.3#	37	0.00
38	Dibenzo(a,h)anthracene	0.400	0.097	75.8#	36	0.00
39	Benzo(a,h,i)perylene	0.400	0.099	75.3#	36	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120419\
Data File : BN008836.D
Acq On : 04 Dec 2019 11:20
Operator : JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDICC0.1

Quant Time: Dec 04 14:19:56 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 14:04:21 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 1				