

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120419\
 Data File : BN008843.D
 Acq On : 04 Dec 2019 16:11
 Operator : JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN120419

Quant Time: Dec 04 16:44:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 15:42:09 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.384	0.383	0.3	103	0.00
3	n-Nitrosodimethylamine	0.531	0.532	-0.2	116	0.00
4 S	2-Fluorophenol	1.083	1.053	2.8	101	0.00
5 S	Phenol-d6	1.475	1.398	5.2	101	0.00
6	bis(2-Chloroethyl)ether	1.327	1.276	3.8	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	-0.01
8 S	Nitrobenzene-d5	0.353	0.329	6.8	99	-0.01
9	Naphthalene	1.067	1.031	3.4	100	-0.01
10	Hexachlorobutadiene	0.210	0.204	2.9	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.664	0.632	4.8	100	0.00
12	2-Methylnaphthalene	0.786	0.750	4.6	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.218	0.198	9.2	98	-0.01
15 S	2-Fluorobiphenyl	1.499	1.440	3.9	100	0.00
16	Acenaphthylene	1.866	1.716	8.0	99	-0.01
17	Acenaphthene	1.184	1.115	5.8	99	-0.01
18	Fluorene	1.570	1.505	4.1	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4-Bromophenyl-phenylether	0.234	0.217	7.3	101	0.00
21	Hexachlorobenzene	0.270	0.249	7.8	98	-0.01
22	Pentachlorophenol	0.108	0.091	15.7	104	0.00
23	Phenanthrene	1.221	1.150	5.8	101	-0.01
24	Anthracene	1.121	1.030	8.1	101	0.00
25 SURR	Fluoranthene-d10	1.211	1.145	5.5	101	0.00
26	Fluoranthene	1.470	1.391	5.4	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
28	Pvrene	1.396	1.337	4.2	101	0.00
29 S	Terphenyl-d14	0.921	0.875	5.0	100	0.00
30	Benzo(a)anthracene	1.446	1.388	4.0	102	0.00
31	Chrysene	1.403	1.335	4.8	98	0.00
32	Bis(2-ethylhexyl)phthalate	0.964	0.816	15.4	99	0.00
33	Indeno(1,2,3-cd)pyrene	1.733	1.705	1.6	103	0.00
34 I	Perylene-d12	1.000	1.000	0.0	102	0.00
35	Benzo(b)fluoranthene	1.329	1.275	4.1	102	0.00
36	Benzo(k)fluoranthene	1.305	1.252	4.1	104	0.00
37 C	Benzo(a)pyrene	1.253	1.192	4.9	103	0.00
38	Dibenzo(a,h)anthracene	1.280	1.239	3.2	103	0.00
39	Benzo(a,h,i)perylene	1.296	1.249	3.6	103	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120419\
Data File : BN008843.D
Acq On : 04 Dec 2019 16:11
Operator : JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN120419

Quant Time: Dec 04 16:44:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 15:42:09 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			