


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120524\
Data File : BN035414.D
Acq On : 04 Dec 2024 08:28
Operator : RC/JU
Sample : DFTPP366
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP366

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN120224.MA.M
Title : SVOA CALIBRATION
Last Update : Mon Dec 02 18:25:21 2024



AutoFind: Scans 2436, 2437, 2438; Background Corrected with Scan 2430

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	22.9	70837	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	30.7	95051	PASS
70	69	0.00	2	0.6	523	PASS
127	198	10	80	40.1	124045	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	309419	PASS
199	198	5	9	6.8	20897	PASS
275	198	10	60	27.9	86336	PASS
365	198	1	100	4.9	15072	PASS
441	198	0.01	100	11.4	35245	PASS
442	198	50	100	71.8	222293	PASS
443	442	15	24	19.4	43152	PASS

m/z	Abundance
39.90	386.0
44.00	1301.0
72.90	351.0
77.10	160.0
191.00	167.0
207.00	1237.0
208.00	266.0
281.00	226.0

Instrument :

BNA_N

ClientSampleId :

DFTPP366

m/z	Abundance
37.00	488.0
38.00	1195.0
39.00	6178.0
40.00	622.0
44.00	937.0
45.10	289.0
50.00	24152.0
51.00	86848.0
52.00	4979.0
53.00	209.0
55.00	477.0
56.00	2738.0
57.00	6469.0
58.00	337.0
61.00	1240.0
62.00	1639.0
63.00	4307.0
64.10	521.0
65.00	2050.0
65.90	160.0
67.10	301.0
69.00	113880.0
70.00	732.0
73.00	2213.0
74.00	13021.0
75.00	19000.0
76.10	6609.0
77.00	125600.0
78.00	9004.0
79.00	9467.0
80.00	6260.0
81.00	9645.0
82.00	2375.0
83.00	2163.0
83.90	308.0
85.00	1426.0
86.00	2896.0
87.00	1313.0
88.00	495.0
88.90	252.0
90.90	2257.0
91.90	2537.0
93.00	18280.0
94.00	1194.0
95.10	279.0
96.00	1012.0
98.00	14232.0
99.00	10265.0
100.10	891.0
101.00	5083.0
101.90	321.0
102.90	1959.0
104.00	3628.0
105.00	3798.0
107.00	45560.0
108.00	6530.0
110.00	72312.0
111.00	11380.0
112.00	1382.0
113.00	552.0
114.90	180.0
116.00	2544.0
117.00	45288.0
118.00	3501.0
118.90	359.0
120.00	453.0
122.00	2969.0
122.90	5031.0
123.90	2208.0
125.00	2066.0
127.00	139200.0
128.00	10971.0
129.00	59272.0
130.00	5349.0
131.00	1089.0
132.00	654.0
132.90	295.0
134.00	1791.0
135.00	4978.0
135.90	1948.0
137.00	2901.0
138.00	536.0
139.00	335.0
139.90	773.0
141.00	8778.0
142.00	2927.0
143.00	1791.0
143.90	527.0
144.80	453.0
145.90	1773.0
147.00	4926.0
148.00	11613.0

Instrument :
BNA_N
ClientSampleId :
DFTPP366

149.00	2297.0
150.00	813.0
151.00	1068.0
152.10	668.0
153.00	2540.0
153.90	2116.0
155.00	4762.0
156.00	6513.0
157.00	1102.0
157.90	1493.0
159.00	1287.0
160.00	2641.0
161.00	4051.0
162.00	1190.0
163.00	323.0
163.90	449.0
164.90	3624.0
166.00	2974.0
167.00	18984.0
168.00	10653.0
169.00	1713.0
170.00	669.0
171.00	721.0
172.00	1572.0
173.00	2104.0
174.00	3931.0
175.00	6699.0
176.00	1922.0
177.00	3253.0
177.90	1176.0
179.00	13087.0
179.90	9179.0
181.00	3822.0
182.00	744.0
182.90	338.0
183.20	333.0
183.90	1160.0
185.00	6773.0
186.00	47456.0
187.00	14308.0
188.00	1546.0
188.90	3199.0
190.00	510.0
191.00	1542.0
191.90	4260.0
193.00	5304.0
194.00	965.0
196.00	8934.0
197.90	328896.0
198.90	23000.0
200.00	2067.0
201.50	1425.0
203.00	2822.0
204.00	13237.0
205.00	22024.0
206.00	89680.0
207.00	13413.0
208.00	3738.0
208.90	1368.0
210.00	1493.0
211.00	3949.0
211.80	806.0
213.00	294.0
214.90	1088.0
215.90	2277.0
216.90	25256.0
218.00	3487.0
218.90	197.0
221.00	10745.0
221.80	3138.0
223.00	5666.0
224.00	50056.0
225.00	12829.0
226.00	1458.0
227.00	24496.0
228.00	3527.0
228.90	4764.0
229.90	515.0
231.00	1811.0
232.10	213.0
233.00	352.0
233.90	1495.0
235.00	1588.0
235.90	1068.0
237.00	2251.0
237.90	253.0
238.90	1047.0
240.00	714.0
241.00	1272.0
242.00	2759.0
243.00	2855.0
244.00	36168.0
245.00	5059.0
245.90	8102.0
246.90	1656.0
248.00	411.0

Instrument :
BNA_N
ClientSampleId :
DFTPP366

248.90	1290.0
249.90	288.0
250.90	356.0
252.00	511.0
253.00	1052.0
255.00	204352.0
256.00	30480.0
257.00	2477.0
258.00	13127.0
258.90	1921.0
260.00	356.0
261.00	413.0
265.00	5347.0
266.00	677.0
267.00	268.0
270.00	290.0
271.00	866.0
271.90	673.0
273.00	6216.0
274.00	16480.0
275.00	83504.0
276.00	11848.0
276.90	9123.0
278.00	1532.0
278.90	347.0
281.00	660.0
282.00	395.0
283.00	1123.0
284.00	665.0
285.00	1503.0
285.90	309.0
287.80	164.0
288.90	396.0
289.90	319.0
290.90	311.0
291.90	249.0
293.00	1960.0
294.00	637.0
296.00	27480.0
297.00	3788.0
297.90	379.0
301.00	289.0
302.00	481.0
303.00	3283.0
304.00	781.0
307.90	385.0
309.00	291.0
309.90	328.0
312.90	230.0
314.00	1292.0
314.90	3311.0
316.00	1868.0
317.00	263.0
320.90	856.0
321.80	444.0
323.00	7650.0
324.00	1273.0
324.90	181.0
326.90	1692.0
328.00	852.0
328.90	178.0
331.90	475.0
333.00	884.0
334.00	5277.0
334.90	1394.0
338.80	184.0
341.00	998.0
342.10	271.0
346.00	1918.0
347.00	369.0
350.80	152.0
352.00	2270.0
353.00	1457.0
354.00	2699.0
355.00	648.0
358.90	202.0
364.90	13482.0
366.00	1823.0
369.90	178.0
371.00	704.0
372.00	4063.0
373.10	1180.0
382.90	1143.0
384.00	311.0
390.00	425.0
390.90	317.0
392.00	227.0
400.90	224.0
402.00	1854.0
403.00	2089.0
403.90	755.0
421.00	1698.0
422.00	1743.0
423.00	11444.0
424.00	3112.0
425.10	288.0

Instrument :
BNA_N
ClientSampleId :
DFTPP366

441.00	27384.0
442.00	179904.0
443.00	36120.0
444.00	3677.0

Instrument :
BNA_N
ClientSampleId :
DFTPP366

m/z	Abundance
37.00	500.0
38.00	1164.0
39.10	6055.0
40.00	519.0
44.00	1124.0
44.90	177.0
50.00	24320.0
51.00	85208.0
52.00	4160.0
53.00	166.0
55.00	551.0
56.00	2639.0
57.00	5889.0
57.90	296.0
61.00	1260.0
62.00	1638.0
63.00	4138.0
64.00	516.0
65.00	2064.0
67.00	234.0
69.00	112808.0
70.00	602.0
73.00	1664.0
74.00	13106.0
75.00	19584.0
76.10	7024.0
77.00	126272.0
78.00	8630.0
79.00	9856.0
80.00	7242.0
81.00	10081.0
82.00	2342.0
83.00	2000.0
83.90	260.0
85.00	1733.0
86.00	3276.0
87.00	1488.0
88.00	586.0
88.80	219.0
91.00	2880.0
92.00	2481.0
93.00	18976.0
93.90	1111.0
95.00	254.0
96.00	802.0
97.00	358.0
98.00	15453.0
99.00	10661.0
99.90	1049.0
100.90	5693.0
102.00	330.0
103.00	1818.0
104.00	3626.0
105.00	3977.0
107.00	47944.0
108.00	6753.0
110.00	77760.0
111.00	12410.0
112.00	1409.0
113.00	539.0
116.10	2771.0
117.00	47440.0
118.00	3526.0
119.00	323.0
120.00	497.0
120.90	298.0
122.00	3402.0
123.00	5075.0
123.90	2204.0
125.00	1865.0
127.00	149760.0
128.00	11686.0
129.00	62608.0
130.00	5155.0
131.00	1178.0
131.90	538.0
133.00	474.0
134.00	1971.0
135.00	5318.0
136.00	2468.0
137.00	3118.0
137.90	545.0
138.90	299.0
140.00	1044.0
141.00	9561.0
142.00	2836.0
142.90	1868.0
143.90	542.0
145.10	390.0
145.90	1792.0
147.00	5211.0
148.00	12459.0

Instrument :
BNA_N
ClientSampleId :
DFTPP366

148.90	2248.0
150.00	753.0
151.00	1111.0
151.70	647.0
152.10	598.0
153.00	3052.0
153.90	2321.0
155.00	4730.0
156.00	7571.0
157.00	1817.0
158.00	1894.0
158.90	1302.0
159.90	3244.0
161.00	4648.0
161.90	1165.0
162.90	459.0
163.90	467.0
165.00	3463.0
166.00	3029.0
167.00	20896.0
168.00	10637.0
169.00	1720.0
170.00	700.0
170.90	843.0
172.00	1859.0
173.00	2260.0
174.00	4461.0
175.00	7586.0
176.00	2007.0
176.90	3827.0
177.90	1395.0
178.90	14883.0
180.00	9889.0
180.90	4417.0
182.00	611.0
182.80	402.0
183.90	1275.0
185.00	7485.0
186.00	56224.0
187.00	15684.0
188.00	1669.0
189.00	3691.0
190.00	726.0
191.00	1856.0
192.00	4926.0
192.90	5420.0
193.90	1069.0
196.00	10125.0
197.90	372800.0
198.90	24112.0
199.90	1891.0
201.50	1560.0
202.20	344.0
202.90	3146.0
204.00	14995.0
205.00	25224.0
206.00	102432.0
207.00	15024.0
207.90	4096.0
209.00	1264.0
210.00	1699.0
210.90	4428.0
212.90	390.0
214.90	1379.0
215.90	2721.0
216.90	28376.0
217.90	3577.0
218.90	424.0
221.00	12646.0
221.80	4024.0
223.00	6472.0
224.00	58688.0
225.00	14866.0
226.00	1742.0
227.00	28128.0
227.90	3696.0
229.00	5158.0
230.00	862.0
231.00	2088.0
232.00	360.0
233.00	591.0
234.00	1890.0
234.90	2005.0
235.90	1619.0
237.00	1847.0
238.00	312.0
238.90	1183.0
240.00	835.0
240.90	1477.0
242.00	3388.0
243.10	3170.0
244.00	41784.0
245.00	5863.0
246.00	10487.0
246.90	2117.0
247.90	607.0

Instrument :
BNA_N
ClientSampleId :
DFTPP366

248.90	1802.0
249.90	407.0
250.90	379.0
251.90	584.0
252.90	937.0
255.00	237120.0
256.00	36680.0
257.00	2769.0
258.00	15868.0
259.00	2475.0
259.90	447.0
260.90	356.0
263.90	466.0
265.00	6293.0
266.00	780.0
267.00	173.0
270.00	297.0
271.00	991.0
272.00	835.0
273.00	7243.0
274.00	20240.0
275.00	106520.0
276.00	13958.0
276.90	10745.0
278.00	1866.0
279.00	418.0
281.00	641.0
281.90	329.0
283.00	1369.0
284.00	818.0
285.00	1834.0
286.00	390.0
287.90	153.0
288.90	437.0
289.90	357.0
291.10	235.0
292.10	486.0
293.00	2468.0
293.90	535.0
296.00	37560.0
297.00	4823.0
297.90	351.0
301.00	576.0
302.00	803.0
303.00	3974.0
304.00	860.0
308.00	516.0
309.00	311.0
309.90	401.0
311.00	192.0
312.00	156.0
312.90	315.0
313.90	1528.0
315.00	4042.0
316.00	2129.0
317.00	320.0
321.00	1223.0
322.00	519.0
323.00	10215.0
324.00	2008.0
325.00	198.0
327.00	2170.0
328.00	1215.0
328.90	225.0
332.00	776.0
333.00	1398.0
334.00	7678.0
335.00	1796.0
335.90	309.0
339.90	197.0
341.00	1439.0
341.90	293.0
346.00	2517.0
346.90	467.0
351.00	262.0
352.00	3391.0
353.00	2111.0
354.00	3654.0
355.10	844.0
356.00	167.0
358.90	220.0
365.00	18288.0
366.00	2386.0
370.00	371.0
371.10	998.0
372.00	5665.0
373.00	1229.0
373.90	196.0
382.90	1669.0
384.00	387.0
390.00	793.0
391.00	472.0
392.00	430.0
400.90	459.0
402.00	2473.0
403.00	3514.0

Instrument :
BNA_N
ClientSampleId :
DFTPP366

404.00	1210.0
405.00	242.0
414.70	157.0
421.00	2662.0
422.00	2584.0
423.00	17224.0
424.00	4199.0
424.90	376.0
441.00	41832.0
442.00	255104.0
443.00	54168.0
444.10	4496.0
444.90	196.0

Instrument :
BNA_N
ClientSampleId :
DFTPP366

m/z	Abundance
37.10	241.0
38.00	597.0
39.00	2843.0
40.00	383.0
44.00	1195.0
48.90	422.0
50.00	11214.0
51.00	40456.0
52.00	2056.0
54.90	344.0
55.90	1210.0
57.00	2857.0
58.00	178.0
61.00	696.0
62.10	812.0
63.00	2047.0
64.00	247.0
65.00	1217.0
69.00	58464.0
70.00	236.0
73.00	1051.0
74.00	6570.0
75.00	10728.0
76.10	3496.0
77.00	65952.0
78.00	4317.0
79.00	5110.0
80.00	3676.0
81.00	5519.0
82.00	1320.0
83.00	1115.0
84.00	229.0
85.00	837.0
85.90	1621.0
86.90	635.0
87.90	337.0
88.90	167.0
91.00	1287.0
92.00	1576.0
93.00	9233.0
93.90	562.0
95.00	200.0
96.00	565.0
98.00	7665.0
99.00	5772.0
100.00	651.0
101.00	3069.0
101.90	210.0
103.00	936.0
104.00	2188.0
105.00	1985.0
107.00	25024.0
108.00	3258.0
110.00	42776.0
111.00	6418.0
111.90	791.0
112.90	384.0
116.10	1444.0
117.00	25248.0
118.00	1720.0
119.00	186.0
120.00	244.0
120.90	167.0
122.00	1875.0
122.90	2841.0
124.00	1418.0
125.00	1326.0
127.00	83176.0
128.00	6352.0
129.00	33368.0
130.00	2792.0
130.90	622.0
131.90	268.0
133.00	238.0
134.00	1093.0
135.00	2907.0
136.00	1269.0
137.00	1594.0
138.10	234.0
138.90	265.0
140.10	469.0
141.00	4983.0
142.00	1721.0
142.90	1224.0
143.90	291.0
145.00	256.0
146.00	1078.0
147.00	2894.0
148.00	6695.0
149.00	1385.0
150.00	373.0
151.00	573.0

Instrument :
BNA_N
ClientSampleId :
DFTPP366

151.90	389.0
152.90	1868.0
153.90	1272.0
155.00	2944.0
156.00	4118.0
157.00	922.0
158.00	936.0
159.00	569.0
160.00	1634.0
161.00	2556.0
161.90	856.0
162.80	234.0
164.00	354.0
165.00	2358.0
166.00	1858.0
167.00	12234.0
168.00	6326.0
169.00	1050.0
169.80	394.0
171.10	380.0
171.90	1181.0
172.90	1345.0
174.00	2280.0
175.00	4692.0
175.90	1213.0
176.90	2118.0
178.00	796.0
178.90	8846.0
180.00	5952.0
181.00	2562.0
182.00	387.0
183.00	230.0
183.90	794.0
185.00	4281.0
186.00	33944.0
187.00	9357.0
188.00	802.0
188.90	2405.0
189.90	295.0
190.90	992.0
192.00	3082.0
193.00	3377.0
194.00	667.0
194.80	209.0
196.00	5559.0
197.90	226560.0
198.90	15579.0
199.90	1361.0
201.40	1044.0
203.00	2014.0
204.00	9055.0
205.00	16007.0
206.00	62688.0
207.00	9548.0
208.00	2458.0
209.00	946.0
210.10	995.0
210.90	2724.0
211.90	304.0
213.00	217.0
215.00	923.0
216.00	1586.0
216.90	18480.0
217.90	2126.0
219.00	246.0
220.00	320.0
221.00	8947.0
221.80	2728.0
222.90	3866.0
224.00	35688.0
225.00	9946.0
226.00	1028.0
227.00	17832.0
228.00	2050.0
228.90	3379.0
230.00	352.0
231.00	1109.0
232.00	217.0
232.80	324.0
233.90	1163.0
235.00	1299.0
235.90	877.0
237.00	1303.0
237.90	231.0
239.00	663.0
239.90	420.0
240.90	1057.0
241.90	2035.0
243.00	2431.0
244.00	28816.0
245.00	3889.0
245.90	6422.0
246.90	1215.0
248.00	289.0
248.90	904.0
250.10	178.0

Instrument :
BNA_N
ClientSampleId :
DFTPP366

250.90	360.0
251.90	249.0
252.90	776.0
255.00	157376.0
256.00	23336.0
257.00	1724.0
258.00	10129.0
258.90	1528.0
259.90	299.0
260.90	233.0
263.70	198.0
264.90	4479.0
265.90	485.0
266.80	171.0
270.00	170.0
270.90	646.0
272.00	463.0
273.00	5310.0
274.00	13390.0
275.00	68984.0
276.00	9224.0
276.90	6797.0
277.90	1150.0
278.90	290.0
281.00	345.0
282.00	456.0
282.90	858.0
283.90	655.0
285.00	1207.0
286.00	211.0
288.90	212.0
290.00	198.0
290.90	290.0
292.00	369.0
292.90	1443.0
294.00	443.0
296.00	23416.0
297.00	3174.0
298.00	281.0
301.00	344.0
302.00	340.0
303.00	2873.0
304.00	753.0
308.00	333.0
308.80	189.0
310.00	303.0
312.90	197.0
314.00	1064.0
315.00	2933.0
316.00	1441.0
317.00	244.0
320.90	658.0
321.90	452.0
323.00	8699.0
324.00	1212.0
326.00	211.0
326.90	1397.0
327.90	751.0
329.00	180.0
332.00	748.0
333.00	972.0
334.00	5537.0
335.00	1418.0
335.90	208.0
340.90	949.0
342.00	293.0
346.00	1740.0
346.80	331.0
351.10	251.0
352.10	2522.0
353.00	1760.0
354.00	2708.0
355.00	604.0
359.10	150.0
365.00	13446.0
365.90	1758.0
370.20	230.0
370.90	731.0
372.00	4647.0
373.10	1105.0
383.00	1301.0
383.90	365.0
389.90	602.0
391.00	464.0
392.10	296.0
400.90	340.0
402.00	1725.0
403.00	2683.0
404.00	802.0
421.00	2309.0
422.00	1977.0
423.00	14631.0
424.00	2921.0
425.00	390.0
441.10	36520.0
442.10	231872.0

Instrument :
BNA_N
ClientSampleId :
DFTPP366

443.00	39168.0
444.00	3781.0
445.00	193.0

Instrument : BNA_N ClientSampleId : DFTPP366
