

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120620\
 Data File : BN012909.D
 Acq On : 06 Dec 2020 09:29
 Operator : CG/JU
 Sample : L4859-05
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0BB6

Quant Time: Dec 07 02:18:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN120320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 02:16:16 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	1526	0.40	ng/ul	0.00
2) Naphthalene-d8	10.15	136	6253	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.06	164	3754	0.40	ng/ul	0.00
10) Phenanthrene-d10	0.00	188	0	0.00	ng/ul	-16.82
16) Chrysene-d12	21.04	240	6945	0.40	ng/ul	0.00
20) Perylene-d12	23.14	264	7322	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.76	152	1850	0.20	ng/ul	0.00
14) Fluoranthene-d10	18.86	212	4280	0.00	ng/ul	0.00
Target Compounds						
17) Pyrene	19.23	202	1021	0.031	ng/ul#	1
23) Benzo(a)pyrene	23.01	252	2055	0.065	ng/ul#	68


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120620\
Data File : BN012909.D
Acq On : 06 Dec 2020 09:29
Operator : CG/JU
Sample : L4859-05
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0BB6

Quant Time: Dec 07 02:18:41 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN120320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 07 02:16:16 2020
Response via : Initial Calibration





#17

Pvrene

Concen: 0.031 ng/ul

RT: 19.23 min Scan# 3619

Delta R.T. -0.03 min

Lab File: BN012909.D

Acq: 06 Dec 2020 09:29

Instrument :

BNA_N

ClientSampleId :

C0BB6


Tot Ion:202 Resp: 1021

Ion Ratio Lower Upper

202 100

101 190.5 9.1 13.7#




100 1291.1 7.5 11.3#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#23

Benzo(a)pyrene

Concen: 0.065 ng/ul

RT: 23.01 min Scan# 4750

Delta R.T. -0.04 min

Lab File: BN012909.D

Acq: 06 Dec 2020 09:29


Tgt Ion:252 Resp: 2055

Ion Ratio Lower Upper

252 100

253 46.5 26.5 39.7#

125 38.8 14.2 21.2#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E

