

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120621\
 Data File : BN017729.D
 Acq On : 06 Dec 2021 13:51
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Dec 06 14:35:01 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 06 13:46:11 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.752	152	17999	0.400	ng	0.00
7) Naphthalene-d8	10.532	136	72785	0.400	ng	# 0.00
13) Acenaphthene-d10	14.373	164	46148	0.400	ng	0.00
19) Phenanthrene-d10	17.117	188	98365	0.400	ng	0.00
29) Chrysene-d12	21.304	240	92742	0.400	ng	# 0.00
35) Perylene-d12	23.613	264	66588	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.354	112	64783	1.528	ng	0.00
5) Phenol-d6	6.931	99	71615	1.654	ng	0.00
8) Nitrobenzene-d5	8.897	82	86634	2.461	ng	0.00
11) 2-Methylnaphthalene-d10	12.125	152	223312	1.734	ng	0.00
14) 2,4,6-Tribromophenol	15.870	330	41247	1.759	ng	0.00
15) 2-Fluorobiphenyl	13.002	172	290837	1.654	ng	0.00
27) Fluoranthene-d10	19.154	212	570908	1.778	ng	0.00
31) Terphenyl-d14	19.755	244	378615	1.840	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.317	88	21586	1.616	ng	94
3) n-Nitrosodimethylamine	3.630	42	29460	1.633	ng	95
6) bis(2-Chloroethyl)ether	7.180	93	75848	1.645	ng	94
9) Naphthalene	10.583	128	297265	1.602	ng	100
10) Hexachlorobutadiene	10.887	225	94785	1.775	ng	# 100
12) 2-Methylnaphthalene	12.196	142	206767	1.706	ng	97
16) Acenaphthylene	14.094	152	315275	1.775	ng	100
17) Acenaphthene	14.436	154	201470	1.665	ng	99
18) Fluorene	15.426	166	265849	1.769	ng	100
20) 4,6-Dinitro-2-methylph...	15.502	198	5382	1.214	ng	95
21) 4-Bromophenyl-phenylether	16.314	248	103443	1.746	ng	# 72
22) Hexachlorobenzene	16.435	284	120365	1.650	ng	# 92
23) Atrazine	16.581	200	76066	2.246	ng	# 95
24) Pentachlorophenol	16.776	266	24829	1.161	ng	100
25) Phenanthrene	17.153	178	439061	1.680	ng	100
26) Anthracene	17.251	178	399874	1.775	ng	100
28) Fluoranthene	19.184	202	537275	1.728	ng	# 97
30) Pyrene	19.541	202	543181	1.839	ng	100
32) Benzo(a)anthracene	21.293	228	485052	1.789	ng	99
33) Chrysene	21.346	228	493328	1.626	ng	99
34) Bis(2-ethylhexyl)phtha...	21.240	149	236748	2.686	ng	95
36) Indeno(1,2,3-cd)pyrene	25.982	276	242520	1.498	ng	# 92
37) Benzo(b)fluoranthene	22.915	252	411177	1.877	ng	# 97
38) Benzo(k)fluoranthene	22.959	252	369771	1.722	ng	# 97
39) Benzo(a)pyrene	23.510	252	322231	1.681	ng	# 96
40) Dibenzo(a,h)anthracene	26.002	278	171320	1.418	ng	# 96
41) Benzo(g,h,i)perylene	26.704	276	233891	1.560	ng	# 94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120621\
Data File : BN017729.D
Acq On : 06 Dec 2021 13:51
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Dec 06 14:35:01 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 06 13:46:11 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.752 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:152 Resp: 17999
 Ion Ratio Lower Upper
 152 100
 150 155.5 123.0 184.4
 115 59.2 45.6 68.4



#2
 1,4-Dioxane
 Concen: 1.616 ng
 RT: 3.317 min Scan# 6
 Delta R.T. -0.000 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51

Tgt Ion: 88 Resp: 21586
 Ion Ratio Lower Upper
 88 100
 43 31.3 24.7 37.1
 58 72.2 63.0 94.6









#3
 n-Nitrosodimethylamine
 Concen: 1.633 ng
 RT: 3.630 min Scan# 40
 Delta R.T. -0.000 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion: 42 Resp: 29460
 Ion Ratio Lower Upper
 42 100
 74 117.0 97.8 146.6
 44 5.7 5.5 8.3



#4
 2-Fluorophenol
 Concen: 1.528 ng
 RT: 5.354 min Scan# 227
 Delta R.T. -0.000 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51

Tgt Ion: 112 Resp: 64783
 Ion Ratio Lower Upper
 112 100
 64 62.3 47.6 71.4
 63 34.5 24.6 37.0









#5
Phenol-d6
Concen: 1.654 ng
RT: 6.931 min Scan# 398
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 71615
Ion Ratio Lower Upper
99 100
42 21.9 17.0 25.6
71 39.6 24.8 37.2#



#6
bis(2-Chloroethyl)ether
Concen: 1.645 ng
RT: 7.180 min Scan# 425
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51

Tgt Ion: 93 Resp: 75848
Ion Ratio Lower Upper
93 100
63 75.4 65.8 98.6
95 32.4 26.6 39.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.532 min Scan# 71
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:136 Resp: 72785
Ion Ratio Lower Upper
136 100
137 10.8 8.8 13.2
54 7.7 4.0 6.0#
68 5.8 3.0 4.6#



#8
Nitrobenzene-d5
Concen: 2.461 ng
RT: 8.897 min Scan# 594
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51

Tgt Ion: 82 Resp: 86634
Ion Ratio Lower Upper
82 100
128 36.9 27.7 41.5
54 49.1 49.2 73.8#









#9
Naphthalene
Concen: 1.602 ng
RT: 10.583 min Scan# 71
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 297265
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.2 10.6 15.8



#10
Hexachlorobutadiene
Concen: 1.775 ng
RT: 10.887 min Scan# 751
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51

Tgt Ion:225 Resp: 94785
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.1 76.7





#11

2-Methylnaphthalene-d10

Concen: 1.734 ng

RT: 12.125 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN017729.D

Acq: 06 Dec 2021 13:51

Instrument :

BNA_N

ClientSampleId :





SSTDICC1.6

Tgt Ion:152 Resp: 223312

Ion Ratio Lower Upper

152 100

151 20.3 16.2 24.4



#12

2-Methylnaphthalene

Concen: 1.706 ng

RT: 12.196 min Scan# 953

Delta R.T. -0.005 min

Lab File: BN017729.D

Acq: 06 Dec 2021 13:51




Tgt Ion:142 Resp: 206767


Ion Ratio Lower Upper

142 100

141 88.5 70.6 106.0

115 38.0 25.8 38.6









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.373 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:164 Resp: 46148
 Ion Ratio Lower Upper
 164 100
 162 101.0 81.4 122.2
 160 45.1 36.8 55.2



#14
 2,4,6-Tribromophenol
 Concen: 1.759 ng
 RT: 15.870 min Scan# 1288
 Delta R.T. -0.000 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51

Tgt Ion:330 Resp: 41247
 Ion Ratio Lower Upper
 330 100
 332 0.0 0.0 0.0
 141 27.4 28.9 43.3#









#15
2-Fluorobiphenyl
Concen: 1.654 ng
RT: 13.002 min Scan# 1094
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 290837
Ion Ratio Lower Upper
172 100
171 34.2 28.7 43.1
170 22.3 19.4 29.2



#16
Acenaphthylene
Concen: 1.775 ng
RT: 14.094 min Scan# 1148
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51

Tgt Ion:152 Resp: 315275
Ion Ratio Lower Upper
152 100
151 19.6 15.5 23.3
153 13.2 10.4 15.6









#17
Acenaphthene
Concen: 1.665 ng
RT: 14.436 min Scan# 1175
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 201470
Ion Ratio Lower Upper
154 100
153 112.4 91.2 136.8
152 54.4 44.3 66.5



#18
Fluorene
Concen: 1.769 ng
RT: 15.426 min Scan# 1253
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51

Tgt Ion:166 Resp: 265849
Ion Ratio Lower Upper
166 100
165 97.1 77.9 116.9
167 13.5 10.9 16.3





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.117 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN017729.D

Acq: 06 Dec 2021 13:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:188 Resp: 98365

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 6.4 6.1 9.1



#20

4,6-Dinitro-2-methylphenol

Concen: 1.214 ng

RT: 15.502 min Scan# 1259

Delta R.T. -0.013 min

Lab File: BN017729.D

Acq: 06 Dec 2021 13:51




Tgt Ion:198 Resp: 5382


Ion Ratio Lower Upper

198 100

182 12.4 11.4 17.2

77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.746 ng
RT: 16.314 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 103443
Ion Ratio Lower Upper
248 100
250 97.6 71.8 107.8
141 44.1 70.9 106.3#



#22
Hexachlorobenzene
Concen: 1.650 ng
RT: 16.435 min Scan# 1334
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51

Tgt Ion:284 Resp: 120365
Ion Ratio Lower Upper
284 100
142 26.4 28.2 42.2#
249 30.2 24.4 36.6









#23
 Atrazine
 Concen: 2.246 ng
 RT: 16.581 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:200 Resp: 76066
 Ion Ratio Lower Upper
 200 100
 173 27.0 17.6 26.4#
 215 48.4 40.2 60.4



#24
 Pentachlorophenol
 Concen: 1.161 ng
 RT: 16.776 min Scan# 1362
 Delta R.T. -0.012 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51

Tgt Ion:266 Resp: 24829
 Ion Ratio Lower Upper
 266 100
 264 62.4 49.8 74.8
 268 63.8 51.3 76.9









#25
Phenanthrene
Concen: 1.680 ng
RT: 17.153 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 439061
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.6
179 15.2 12.3 18.5



#26
Anthracene
Concen: 1.775 ng
RT: 17.251 min Scan# 1401
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51

Tgt Ion:178 Resp: 399874
Ion Ratio Lower Upper
178 100
176 18.1 14.6 22.0
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 1.778 ng
RT: 19.154 min Scan# 1693
Delta R.T. -0.000 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 570908
Ion Ratio Lower Upper
212 100
106 9.2 10.5 15.7#
104 5.1 5.8 8.6#



#28
Fluoranthene
Concen: 1.728 ng
RT: 19.184 min Scan# 1699
Delta R.T. 0.005 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51

Tgt Ion:202 Resp: 537275
Ion Ratio Lower Upper
202 100
101 8.2 8.9 13.3#
203 17.4 13.8 20.6





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.304 min Scan# 1987

Delta R.T. -0.000 min

Lab File: BN017729.D

Acq: 06 Dec 2021 13:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:240 Resp: 92742

Ion Ratio Lower Upper

240 100

120 8.3 4.2 6.2#

236 26.6 20.0 30.0



#30

Pyrene

Concen: 1.839 ng

RT: 19.541 min Scan# 1771

Delta R.T. -0.000 min

Lab File: BN017729.D

Acq: 06 Dec 2021 13:51




Tgt Ion:202 Resp: 543181


Ion Ratio Lower Upper

202 100

200 20.7 16.5 24.7

203 17.5 14.0 21.0









#31
 Terphenyl-d14
 Concen: 1.840 ng
 RT: 19.755 min Scan# 1846
 Delta R.T. -0.000 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244 Resp: 378615
 Ion Ratio Lower Upper
 244 100
 212 7.6 5.7 8.5
 122 8.4 8.6 13.0#



#32
 Benzo(a)anthracene
 Concen: 1.789 ng
 RT: 21.293 min Scan# 1986
 Delta R.T. -0.000 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51

Tgt Ion:228 Resp: 485052
 Ion Ratio Lower Upper
 228 100
 226 25.9 21.3 31.9
 229 19.8 15.7 23.5





#33

Chrysene

Concen: 1.626 ng

RT: 21.346 min Scan# 1981

Delta R.T. -0.000 min

Lab File: BN017729.D

Acq: 06 Dec 2021 13:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 493328

Ion Ratio Lower Upper

228 100

226 28.4 23.3 34.9

229 19.9 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 2.686 ng

RT: 21.240 min Scan# 1981

Delta R.T. 0.010 min

Lab File: BN017729.D

Acq: 06 Dec 2021 13:51




Tgt Ion:149 Resp: 236748


Ion Ratio Lower Upper

149 100

167 29.9 21.8 32.6

279 5.0 3.4 5.0









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.613 min Scan# 2476
 Delta R.T. -0.000 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:264 Resp: 66588
 Ion Ratio Lower Upper
 264 100
 260 25.3 19.4 29.2
 265 26.7 18.6 28.0



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 1.498 ng
 RT: 25.982 min Scan# 3135
 Delta R.T. -0.011 min
 Lab File: BN017729.D
 Acq: 06 Dec 2021 13:51

Tgt Ion:276 Resp: 242520
 Ion Ratio Lower Upper
 276 100
 138 17.0 19.7 29.5#
 277 25.0 20.0 30.0





#37

Benzo(b)fluoranthene

Concen: 1.877 ng

RT: 22.915 min Scan# 21

Delta R.T. -0.004 min

Lab File: BN017729.D

Acq: 06 Dec 2021 13:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 411177

Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 6.7 7.8 11.8#



#38

Benzo(k)fluoranthene

Concen: 1.722 ng

RT: 22.959 min Scan# 2252

Delta R.T. -0.004 min

Lab File: BN017729.D

Acq: 06 Dec 2021 13:51




Tgt Ion:252 Resp: 369771


Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5

125 6.9 8.2 12.2#









#39
Benzo(a)pyrene
Concen: 1.681 ng
RT: 23.510 min Scan# 2446
Delta R.T. -0.004 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 322231
Ion Ratio Lower Upper
252 100
253 22.1 17.9 26.9
125 7.5 9.2 13.8#



#40
Dibenzo(a,h)anthracene
Concen: 1.418 ng
RT: 26.002 min Scan# 3141
Delta R.T. -0.011 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51


Tgt Ion:278 Resp: 171320
Ion Ratio Lower Upper
278 100
139 12.1 13.0 19.6#
279 23.8 19.2 28.8





#41
Benzo(g,h,i)perylene
Concen: 1.560 ng
RT: 26.704 min Scan# 31
Delta R.T. -0.007 min
Lab File: BN017729.D
Acq: 06 Dec 2021 13:51

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:276 Resp: 233891
Ion Ratio Lower Upper
276 100
277 23.7 19.0 28.4
138 15.1 17.1 25.7#

