


2,4,6-Tribromophenol

Response Ratio




$$\text{Response} = 2.520\text{e-}001 * \text{Amt} - 8.740\text{e-}002$$

Coef of Det (r^2) = 0.999264 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120621.M

Calibration Table Last Updated: Mon Dec 06 15:58:05 2021


4,6-Dinitro-2-methylphenol



$R = 2.471e-003 A^2 + 4.112e-003 A - 1.259e-003$
 Coef of Det (r^2) = 0.999931 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120621.M
 Calibration Table Last Updated: Mon Dec 06 15:58:05 2021

Pentachlorophenol


Response Ratio



$R = 3.766e-003 A^2 + 5.607e-002 A - 2.299e-002$
 Coef of Det (r^2) = 0.999789 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120621.M
 Calibration Table Last Updated: Mon Dec 06 15:58:05 2021

Atrazine

Response Ratio



$$\text{Response} = 2.114\text{e-}001 * \text{Amt} - 5.183\text{e-}002$$

Coef of Det (r^2) = 0.999628 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120621.M

Calibration Table Last Updated: Mon Dec 06 15:58:05 2021