

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120621\
 Data File : BN017755.D
 Acq On : 07 Dec 2021 07:21
 Operator : CG/JU
 Sample : M4918-01
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-37

Quant Time: Dec 07 16:02:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 06 15:58:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.752	152	24340	0.400	ng	0.00
7) Naphthalene-d8	10.532	136	100110	0.400	ng	# 0.00
13) Acenaphthene-d10	14.373	164	65458	0.400	ng	0.00
19) Phenanthrene-d10	17.117	188	139281	0.400	ng	0.00
29) Chrysene-d12	21.304	240	131872	0.400	ng	# 0.00
35) Perylene-d12	23.613	264	116508	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.373	112	7998	0.153	ng	0.02
5) Phenol-d6	6.950	99	4748	0.085	ng	0.02
8) Nitrobenzene-d5	8.897	82	20292	0.300	ng	0.00
11) 2-Methylnaphthalene-d10	12.126	152	42651	0.233	ng	0.00
14) 2,4,6-Tribromophenol	15.870	330	13150	0.458	ng	0.00
15) 2-Fluorobiphenyl	13.003	172	77951	0.322	ng	0.00
27) Fluoranthene-d10	19.169	212	124322	0.265	ng	0.01
31) Terphenyl-d14	19.755	244	119747	0.354	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.326	88	173	0.011	ng	# 1
6) bis(2-Chloroethyl)ether	7.208	93	290	0.005	ng	# 18
9) Naphthalene	10.583	128	11369	0.046	ng	# 93
10) Hexachlorobutadiene	10.875	225	43	0.001	ng	# 66
12) 2-Methylnaphthalene	12.201	142	11138	0.065	ng	# 96
16) Acenaphthylene	14.094	152	3164	0.013	ng	# 79
17) Acenaphthene	14.436	154	43442	0.255	ng	99
18) Fluorene	15.426	166	11407	0.052	ng	# 97
21) 4-Bromophenyl-phenylether	16.314	248	62	0.001	ng	# 1
22) Hexachlorobenzene	16.436	284	91	0.001	ng	# 1
24) Pentachlorophenol	16.776	266	185	0.169	ng	# 84
25) Phenanthrene	17.154	178	25600	0.070	ng	99
26) Anthracene	17.251	178	5586	0.017	ng	# 86
28) Fluoranthene	19.199	202	7585	0.017	ng	# 92
30) Pyrene	19.547	202	6674	0.013	ng	# 84
32) Benzo(a)anthracene	21.293	228	4028	0.010	ng	# 31
33) Chrysene	21.346	228	2180	0.005	ng	# 17
34) Bis(2-ethylhexyl)phtha...	21.240	149	287265	1.381	ng	98
36) Indeno(1,2,3-cd)pyrene	25.992	276	2028	0.008	ng	# 73
37) Benzo(b)fluoranthene	22.919	252	3276	0.008	ng	# 1
38) Benzo(k)fluoranthene	22.963	252	2452	0.006	ng	# 1
39) Benzo(a)pyrene	23.514	252	2270	0.007	ng	# 1
40) Dibenzo(a,h)anthracene	25.999	278	1379	0.008	ng	# 1
41) Benzo(g,h,i)perylene	26.711	276	1999	0.008	ng	# 22


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120621\
Data File : BN017755.D
Acq On : 07 Dec 2021 07:21
Operator : CG/JU
Sample : M4918-01
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-37

Quant Time: Dec 07 16:02:12 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 06 15:58:05 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.752 min Scan# 48
 Delta R.T. 0.000 min
 Lab File: BN017755.D
 Acq: 07 Dec 2021 07:21




Instrument :
 BNA_N
 ClientSampleId :
 MW-37


Tgt Ion:152 Resp: 24340
 Ion Ratio Lower Upper
 152 100
 150 156.9 123.0 184.4
 115 59.5 45.6 68.4



#2
 1,4-Dioxane
 Concen: 0.011 ng
 RT: 3.326 min Scan# 7
 Delta R.T. 0.009 min
 Lab File: BN017755.D
 Acq: 07 Dec 2021 07:21

Tgt Ion: 88 Resp: 173
 Ion Ratio Lower Upper
 88 100
 43 587.3 24.7 37.1#
 58 42.2 63.0 94.6#









#4
2-Fluorophenol
Concen: 0.153 ng
RT: 5.373 min Scan# 21
Delta R.T. 0.018 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion: 112 Resp: 7998
Ion Ratio Lower Upper
112 100
64 59.2 47.6 71.4
63 26.0 24.6 37.0



#5
Phenol-d6
Concen: 0.085 ng
RT: 6.950 min Scan# 400
Delta R.T. 0.018 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion: 99 Resp: 4748
Ion Ratio Lower Upper
99 100
42 28.6 17.0 25.6#
71 57.6 24.8 37.2#









#6
bis(2-Chloroethyl)ether
Concen: 0.005 ng
RT: 7.208 min Scan# 425
Delta R.T. 0.028 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion: 93 Resp: 290
Ion Ratio Lower Upper
93 100
63 0.0 65.8 98.6#
95 0.0 26.6 39.8#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.532 min Scan# 723
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion: 136 Resp: 100110
Ion Ratio Lower Upper
136 100
137 11.1 8.8 13.2
54 8.2 4.0 6.0#
68 6.9 3.0 4.6#









#8
Nitrobenzene-d5
Concen: 0.300 ng
RT: 8.897 min Scan# 594
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion: 82 Resp: 20292
Ion Ratio Lower Upper
82 100
128 36.1 27.7 41.5
54 51.3 49.2 73.8



#9
Naphthalene
Concen: 0.046 ng
RT: 10.583 min Scan# 727
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion: 128 Resp: 11369
Ion Ratio Lower Upper
128 100
129 14.3 8.8 13.2#
127 15.5 10.6 15.8









#10
Hexachlorobutadiene
Concen: 0.001 ng
RT: 10.875 min Scan# 71
Delta R.T. -0.013 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion:225 Resp: 43
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 37.2 51.1 76.7#



#11
2-Methylnaphthalene-d10
Concen: 0.233 ng
RT: 12.126 min Scan# 937
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion:152 Resp: 42651
Ion Ratio Lower Upper
152 100
151 20.7 16.2 24.4





#12

2-Methylnaphthalene

Concen: 0.065 ng

RT: 12.201 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN017755.D

Acq: 07 Dec 2021 07:21

Instrument :

BNA_N

ClientSampleId :

MW-37





Tgt Ion:142 Resp: 11138

Ion Ratio Lower Upper

142 100

141 87.9 70.6 106.0

115 39.2 25.8 38.6#



#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.373 min Scan# 1170

Delta R.T. 0.000 min

Lab File: BN017755.D

Acq: 07 Dec 2021 07:21




Tgt Ion:164 Resp: 65458


Ion Ratio Lower Upper

164 100

162 100.5 81.4 122.2

160 45.8 36.8 55.2









#14
2,4,6-Tribromophenol
Concen: 0.458 ng
RT: 15.870 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion: 330 Resp: 13150
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 26.4 28.9 43.3#



#15
2-Fluorobiphenyl
Concen: 0.322 ng
RT: 13.003 min Scan# 1062
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion: 172 Resp: 77951
Ion Ratio Lower Upper
172 100
171 34.5 28.7 43.1
170 22.7 19.4 29.2









#16
Acenaphthylene
Concen: 0.013 ng
RT: 14.094 min Scan# 1148
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion:152 Resp: 3164
Ion Ratio Lower Upper
152 100
151 22.9 15.5 23.3
153 29.2 10.4 15.6#



#17
Acenaphthene
Concen: 0.255 ng
RT: 14.436 min Scan# 1175
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion:154 Resp: 43442
Ion Ratio Lower Upper
154 100
153 112.8 91.2 136.8
152 54.2 44.3 66.5









#18
Fluorene
Concen: 0.052 ng
RT: 15.426 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion:166 Resp: 11407
Ion Ratio Lower Upper
166 100
165 96.4 77.9 116.9
167 20.0 10.9 16.3#



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.117 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion:188 Resp: 139281
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.5 6.1 9.1





#21

4-Bromophenyl-phenylether

Concen: 0.001 ng

RT: 16.314 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN017755.D

Acq: 07 Dec 2021 07:21

Instrument :

BNA_N

ClientSampleId :

MW-37





Tgt Ion:248 Resp: 62

Ion Ratio Lower Upper

248 100

250 113.1 71.8 107.8#

141 731.4 70.9 106.3#



#22

Hexachlorobenzene

Concen: 0.001 ng

RT: 16.436 min Scan# 1334

Delta R.T. -0.000 min

Lab File: BN017755.D

Acq: 07 Dec 2021 07:21




Tgt Ion:284 Resp: 91


Ion Ratio Lower Upper

284 100

142 318.7 28.2 42.2#

249 52.7 24.4 36.6#









#24
Pentachlorophenol
Concen: 0.169 ng
RT: 16.776 min Scan# 1362
Delta R.T. -0.013 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion:266 Resp: 185
Ion Ratio Lower Upper
266 100
264 44.9 49.8 74.8#
268 56.2 51.3 76.9



#25
Phenanthrene
Concen: 0.070 ng
RT: 17.154 min Scan# 1393
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion:178 Resp: 25600
Ion Ratio Lower Upper
178 100
176 18.5 15.0 22.6
179 14.9 12.3 18.5









#26
 Anthracene
 Concen: 0.017 ng
 RT: 17.251 min Scan# 1401
 Delta R.T. -0.000 min
 Lab File: BN017755.D
 Acq: 07 Dec 2021 07:21




Instrument :
 BNA_N
 ClientSampleId :
 MW-37


Tgt Ion:178 Resp: 5586
 Ion Ratio Lower Upper
 178 100
 176 19.5 14.6 22.0
 179 26.8 12.2 18.4#



#27
 Fluoranthene-d10
 Concen: 0.265 ng
 RT: 19.169 min Scan# 1696
 Delta R.T. 0.015 min
 Lab File: BN017755.D
 Acq: 07 Dec 2021 07:21

Tgt Ion:212 Resp: 124322
 Ion Ratio Lower Upper
 212 100
 106 10.1 10.5 15.7#
 104 5.9 5.8 8.6









#28
Fluoranthene
Concen: 0.017 ng
RT: 19.199 min Scan# 11
Delta R.T. 0.020 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion:202 Resp: 7585
Ion Ratio Lower Upper
202 100
101 17.4 8.9 13.3#
203 18.6 13.8 20.6



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.304 min Scan# 1987
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion:240 Resp: 131872
Ion Ratio Lower Upper
240 100
120 11.8 4.2 6.2#
236 27.7 20.0 30.0









#30
Pyrene
Concen: 0.013 ng
RT: 19.547 min Scan# 11
Delta R.T. 0.005 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion:202 Resp: 6674
Ion Ratio Lower Upper
202 100
200 23.4 16.5 24.7
203 29.6 14.0 21.0#



#31
Terphenyl-d14
Concen: 0.354 ng
RT: 19.755 min Scan# 1814
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion:244 Resp: 119747
Ion Ratio Lower Upper
244 100
212 8.1 5.7 8.5
122 10.4 8.6 13.0









#32
Benzo(a)anthracene
Concen: 0.010 ng
RT: 21.293 min Scan# 1986
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion:228 Resp: 4028
Ion Ratio Lower Upper
228 100
226 50.7 21.3 31.9#
229 65.0 15.7 23.5#



#33
Chrysene
Concen: 0.005 ng
RT: 21.346 min Scan# 1991
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion:228 Resp: 2180
Ion Ratio Lower Upper
228 100
226 60.1 23.3 34.9#
229 73.7 15.6 23.4#









#34
Bis(2-ethylhexyl)phthalate
Concen: 1.381 ng
RT: 21.240 min Scan# 1980
Delta R.T. 0.011 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion:149 Resp: 287265
Ion Ratio Lower Upper
149 100
167 28.4 21.8 32.6
279 4.8 3.4 5.0



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.613 min Scan# 2443
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion:264 Resp: 116508
Ion Ratio Lower Upper
264 100
260 25.6 19.4 29.2
265 28.1 18.6 28.0









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.008 ng
RT: 25.992 min Scan# 3138
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21




Instrument :
BNA_N
ClientSampleId :
MW-37


Tgt Ion:276 Resp: 2028
Ion Ratio Lower Upper
276 100
138 0.0 19.7 29.5#
277 27.5 20.0 30.0



#37
Benzo(b)fluoranthene
Concen: 0.008 ng
RT: 22.919 min Scan# 2240
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21


Tgt Ion:252 Resp: 3276
Ion Ratio Lower Upper
252 100
253 77.8 17.7 26.5#
125 73.7 7.8 11.8#








#38
Benzo(k)fluoranthene
Concen: 0.006 ng
RT: 22.963 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Instrument :
BNA_N
ClientSampleId :
MW-37







Tgt Ion:252 Resp: 2452
Ion Ratio Lower Upper
252 100
253 89.6 17.7 26.5#
125 88.9 8.2 12.2#



#39
Benzo(a)pyrene
Concen: 0.007 ng
RT: 23.514 min Scan# 2414
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion:252 Resp: 2270
Ion Ratio Lower Upper
252 100
253 93.2 17.9 26.9#
125 93.1 9.2 13.8#









#40
Dibenzo(a,h)anthracene
Concen: 0.008 ng
RT: 25.999 min Scan# 3144
Delta R.T. -0.014 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Instrument :
BNA_N
ClientSampleId :
MW-37

Tgt Ion:278 Resp: 1379
Ion Ratio Lower Upper
278 100
139 149.7 13.0 19.6#
279 94.8 19.2 28.8#



#41
Benzo(g,h,i)perylene
Concen: 0.008 ng
RT: 26.711 min Scan# 3348
Delta R.T. -0.000 min
Lab File: BN017755.D
Acq: 07 Dec 2021 07:21

Tgt Ion:276 Resp: 1999
Ion Ratio Lower Upper
276 100
277 56.1 19.0 28.4#
138 64.3 17.1 25.7#

