



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120621\
 Data File : BN017743.D
 Acq On : 06 Dec 2021 22:21
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120621.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Dec 06 15:58:05 2021






AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2470

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.0	250347	PASS
68	69	0.00	2	1.3	3491	PASS
69	198	0.00	100	35.6	262528	PASS
70	69	0.00	2	0.6	1685	PASS
127	198	10	80	37.5	276437	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	736597	PASS
199	198	5	9	6.7	49029	PASS
275	198	10	60	27.7	204011	PASS
365	198	1	100	3.6	26237	PASS
441	198	0.01	100	12.1	88936	PASS
442	442	50	100	100.0	538048	PASS
443	442	15	24	19.4	104176	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120621\
Data File : BN017743.D
Acq On : 06 Dec 2021 22:21
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 07 08:42:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 07 08:42:11 2021
Response via : Initial Calibration



TIC: BN017743.D\data.ms

(70) Pentachlorophenol (C)

16.751min (0.000) 22872.65 ng




response 919034

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	67.73
264.00	61.60	65.69
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120621\
Data File : BN017743.D
Acq On : 06 Dec 2021 22:21
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 07 08:42:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 07 08:42:11 2021
Response via : Initial Calibration



TIC: BN017743.D\data.ms

(77) Benzidine

19.351min (0.000) 619285.76 ng

response 6070529

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.39
183.00	13.20	11.84
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/6/2021	BNA_N	<u>BN017943.D</u>
Compound Name	Response	Retention Time
DDT	2524356	20.592
DDD	92126	20.151
DDE	6346	19.645
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
98472	2622828	3.75

Instrument :
BNA_N
ClientSampleId :
DFTPP