

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120622\
 Data File : BN023078.D
 Acq On : 06 Dec 2022 17:45
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN120622

Quant Time: Dec 07 02:12:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 07 02:02:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	71	0.00
2	1,4-Dioxane	0.502	0.465	7.4	76	0.00
3	n-Nitrosodimethylamine	0.502	0.435	13.3	73	0.00
4 S	2-Fluorophenol	0.941	0.769	18.3	67	0.00
5 S	Phenol-d6	1.190	0.949	20.3	67	0.00
6	bis(2-Chloroethyl)ether	1.282	1.174	8.4	74	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	70	0.00
8 S	Nitrobenzene-d5	0.300	0.246	18.0	70	0.00
9	Naphthalene	1.187	1.085	8.6	74	0.00
10	Hexachlorobutadiene	0.222	0.209	5.9	76	0.00
11 SURR	2-Methylnaphthalene-d10	0.803	0.621	22.7	59	0.00
12	2-Methylnaphthalene	0.211	0.160	24.2	68	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.171	0.127	25.7#	62	0.00
15 S	2-Fluorobiphenyl	1.767	1.566	11.4	69	0.00
16	Acenaphthylene	1.884	1.617	14.2	72	0.00
17	Acenaphthene	1.365	1.240	9.2	72	0.00
18	Fluorene	1.639	1.465	10.6	71	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	67	0.00
20	4,6-Dinitro-2-methylphenol	0.053	0.046	13.2	68	0.00
21	4-Bromophenyl-phenylether	0.247	0.212	14.2	68	0.01
22	Hexachlorobenzene	0.312	0.296	5.1	74	0.00
23	Atrazine	0.177	0.145	18.1	69	0.00
24	Pentachlorophenol	0.086	0.083	3.5	75	0.00
25	Phenanthrene	1.362	1.247	8.4	72	0.00
26	Anthracene	1.097	0.940	14.3	70	0.00
27 SURR	Fluoranthene-d10	1.101	0.910	17.3	67	0.00
28	Fluoranthene	1.483	1.306	11.9	71	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	67	0.00
30	Pyrene	1.718	1.576	8.3	71	0.00
31 S	Terphenyl-d14	0.827	0.714	13.7	67	0.00
32	Benzo(a)anthracene	1.495	1.304	12.8	69	0.00
33	Chrysene	1.642	1.547	5.8	72	0.00
34	Bis(2-ethylhexyl)phthalate	0.634	0.526	17.0	71	0.00
35 I	Perylene-d12	1.000	1.000	0.0	68	0.00
36	Indeno(1,2,3-cd)pyrene	2.157	1.975	8.4	75	0.00
37	Benzo(b)fluoranthene	1.843	1.676	9.1	73	0.00
38	Benzo(k)fluoranthene	1.880	1.740	7.4	75	0.00
39 C	Benzo(a)pyrene	1.495	1.295	13.4	73	0.00
40	Dibenzo(a,h)anthracene	1.707	1.547	9.4	74	0.00
41	Benzo(g,h,i)perylene	1.768	1.571	11.1	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN120622