

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120622\
 Data File : BN023078.D
 Acq On : 06 Dec 2022 17:45
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN120622

Quant Time: Dec 07 02:12:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 07 02:02:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	71	0.00
2	1,4-Dioxane	0.400	0.371	7.3	76	0.00
3	n-Nitrosodimethylamine	0.400	0.346	13.5	73	0.00
4 S	2-Fluorophenol	0.400	0.327	18.3	67	0.00
5 S	Phenol-d6	0.400	0.319	20.3	67	0.00
6	bis(2-Chloroethyl)ether	0.400	0.366	8.5	74	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	70	0.00
8 S	Nitrobenzene-d5	0.400	0.328	18.0	70	0.00
9	Naphthalene	0.400	0.366	8.5	74	0.00
10	Hexachlorobutadiene	0.400	0.376	6.0	76	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.309	22.8	59	0.00
12	2-Methylnaphthalene	0.400	0.304	24.0	68	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.400	0.297	25.8#	62	0.00
15 S	2-Fluorobiphenyl	0.400	0.354	11.5	69	0.00
16	Acenaphthylene	0.400	0.343	14.2	72	0.00
17	Acenaphthene	0.400	0.363	9.3	72	0.00
18	Fluorene	0.400	0.358	10.5	71	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	67	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.349	12.8	68	0.00
21	4-Bromophenyl-phenylether	0.400	0.344	14.0	68	0.01
22	Hexachlorobenzene	0.400	0.378	5.5	74	0.00
23	Atrazine	0.400	0.328	18.0	69	0.00
24	Pentachlorophenol	0.400	0.383	4.3	75	0.00
25	Phenanthrene	0.400	0.366	8.5	72	0.00
26	Anthracene	0.400	0.343	14.2	70	0.00
27 SURR	Fluoranthene-d10	0.400	0.331	17.3	67	0.00
28	Fluoranthene	0.400	0.352	12.0	71	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	67	0.00
30	Pyrene	0.400	0.367	8.3	71	0.00
31 S	Terphenyl-d14	0.400	0.345	13.8	67	0.00
32	Benzo(a)anthracene	0.400	0.349	12.8	69	0.00
33	Chrysene	0.400	0.377	5.8	72	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.332	17.0	71	0.00
35 I	Perylene-d12	0.400	0.400	0.0	68	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.366	8.5	75	0.00
37	Benzo(b)fluoranthene	0.400	0.364	9.0	73	0.00
38	Benzo(k)fluoranthene	0.400	0.370	7.5	75	0.00
39 C	Benzo(a)pyrene	0.400	0.347	13.3	73	0.00
40	Dibenzo(a,h)anthracene	0.400	0.363	9.3	74	0.00
41	Benzo(g,h,i)perylene	0.400	0.355	11.3	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN120622