

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120624\
 Data File : BN035441.D
 Acq On : 06 Dec 2024 00:54
 Operator : RC/JU
 Sample : PB165400BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK400

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/07/2024
 Supervised By :mohammad ahmed 12/13/2024

Quant Time: Dec 06 02:05:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN120224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 02 18:25:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.299	152	98055	20.000	ng/u1	0.00
20) Naphthalene-d8	10.046	136	413630	20.000	ng/u1	0.00
38) Acenaphthene-d10	13.957	164	325212	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.722	188	737930	20.000	ng/u1	0.00
79) Chrysene-d12	20.980	240	751529	20.000	ng/u1	0.00
88) Perylene-d12	23.651	264	880554	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.958	96	10248m	4.938	ng/uL	0.00
4) Pyridine-d5	3.334	84	129642	23.701	ng/u1	0.00
7) Phenol-d5	6.499	99	163296	23.377	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.658	67	96333	24.729	ng/u1	0.00
11) 2-Chlorophenol-d4	6.840	132	137834	23.285	ng/u1	0.00
15) 4-Methylphenol-d8	8.010	113	142713	23.443	ng/u1	0.00
21) Nitrobenzene-d5	8.434	128	68210	25.382	ng/u1	0.00
24) 2-Nitrophenol-d4	9.146	143	79457	25.545	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.681	165	155733	22.908	ng/u1	0.00
31) 4-Chloroaniline-d4	10.193	131	211255	25.166	ng/u1	0.00
46) Dimethylphthalate-d6	13.387	166	590838	25.302	ng/u1	0.00
49) Acenaphthylene-d8	13.645	160	620446	24.126	ng/u1	0.00
54) 4-Nitrophenol-d4	14.192	143	83956	21.834	ng/u1	0.00
60) Fluorene-d10	14.969	176	481053	23.855	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.098	200	80403	20.660	ng/u1	0.00
73) Anthracene-d10	16.822	188	823346	25.297	ng/u1	0.00
81) Pyrene-d10	19.145	212	969917	24.436	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.468	264	1094352	25.665	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120624\
Data File : BN035441.D
Acq On : 06 Dec 2024 00:54
Operator : RC/JU
Sample : PB165400BL
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK400

Quant Time: Dec 06 02:05:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN120224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 02 18:25:21 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/07/2024
Supervised By :mohammad ahmed 12/13/2024

